

RDE Calibration Report

Date: 2026-05-24 14:45:15
RDE IDN: bartkepl,RDE,15A07EC4,1.0
Meter IDN: HEWLETT-PACKARD,34401A,0,4-1-1
Mode: 4W (FRES), NPLC=100.0, Avg=13

Decade	d=0	d=1	d=2	d=3	d=4	d=5	d=6	d=7	d=8	d=9
1 Ohm (d1)	20	1264	2288	3272	4293	5278	6300	7284	8284	9270
10 Ohm (d2)	19	10249	20260	30229	40238	50207	60217	70185	80167	90137
100 Ohm(d3)	18	100174	200106	300002	399923	499819	599716	699610	799536	899435
1 kOhm (d4)	18	999116	1998034	2996850	3995659	4994489	5993378	6992187	7991022	8989879
10 kOhm(d5)	17	9989037	19980397	29969786	39960183	49949781	59943149	69932556	79921805	89911895
100 kOhm(d6)	17	99898523	199781508	299671577	399558658	499449693	599327475	699218828	799106822	898998239

Deviation from Nominal:

Decade	d=0 (mO)	d=1 %	d=2 %	d=3 %	d=4 %	d=5 %	d=6 %	d=7 %	d=8 %	d=9 %
1 Ohm (d1)	20	26.40	14.40	9.07	7.32	5.56	5.00	4.06	3.55	3.00
10 Ohm (d2)	19	2.49	1.30	0.76	0.60	0.41	0.36	0.26	0.21	0.15
100 Ohm(d3)	18	0.17	0.05	0.00	0.02	0.04	0.05	0.06	0.06	0.06
1 kOhm (d4)	18	0.09	0.10	0.10	0.11	0.11	0.11	0.11	0.11	0.11
10 kOhm(d5)	17	0.11	0.10	0.10	0.10	0.10	0.09	0.10	0.10	0.10
100 kOhm(d6)	17	0.10	0.11	0.11	0.11	0.11	0.11	0.11	0.11	0.11

Summary:

Decade	Short d=0 (mO)	Mean dev %	Max dev %
1 Ohm (d1)	20	8.71	26.40
10 Ohm (d2)	19	0.73	2.49
100 Ohm(d3)	18	0.06	0.17
1 kOhm (d4)	18	0.11	0.11
10 kOhm(d5)	17	0.10	0.11
100 kOhm(d6)	17	0.11	0.11
ALL DECADES	-	1.63	26.40