

Oxygen charge state distribution from injected BeO^- , O_2^- and O^- ions in a tandem accelerator

B. C. Gonsalves^a, K. Mizohata^a, K. Helariutta^a, P. O. Tikkanen^a and J. A. Räisänen^a

^aDepartment of Physics, University of Helsinki, P.O. Box 43, Helsinki, 00014, Uusimaa, Finland

ARTICLE INFO

Keywords:
AMS
Oxygen
Charge state

ABSTRACT

Charge exchange of ions in a tandem accelerator is a crucial part of Accelerator Mass Spectrometry (AMS). In AMS measurements of Be, molecular beams of beryllium oxide are chosen to achieve higher negative ion yields from the ion source. In this study, oxygen charge state distribution dependence of the injected particle, atomic or molecular particles, were examined. Charge state distribution of oxygen with several different terminal potentials from injected O , O_2 and BeO were measured and results were compared to common charge exchange theories.

1. Introduction

The Helsinki Accelerator Mass Spectrometry (HAMS) facility is based on an EGP-10-II 5-MV tandem accelerator. Currently it is dedicated to ^{14}C determination in Black carbon measurements and in Bio-fractionation-assays, and ^{14}C measurements for applications in Archeology, Environmental sciences, Biological sciences and Pharmaceuticals. Currently there is interest in expanding into the geosciences domain by measuring cosmogenic radionuclides starting with ^{10}Be followed on by ^{26}Al . This means increasing the variety of isotopes to analyze with the HAMS system.

AMS is a technique to measure the ratio of rare isotopes to the total number of that element's atoms in a sample using a particle accelerator, ion detectors and other beam line apparatus. A crucial aspect of this is molecular background elimination, based on breaking of the molecules in the stripper. This happens within the stripping media and results in charge exchange. There is a choice of stripping media which can be: gas, foil or a combination of these. The charge exchange process (or electron stripping process) is defined as the capture or loss of electrons in a target medium, resulting in change of the electrical charge state of the bombarding ions. Charge state distributions is of special importance for optimisation of AMS systems. The charge exchange process depends on several factors, including the properties of both the bombarding ions and the stripper medium [1, 2].

The charge exchange theory has its origin in the 1940s with the study of fission fragments and with heavy ions in the 1950s and 1960s. The theory has been summarised by the review paper of H-D Betz [2]. The similarities and differences between charge stripping in gases and in solids are discussed herein. In literature, the effect of molecules passing through a solid stripper has been investigated [3, 4, 5]. The basic charge exchange process—the cross sections, which indicate the probability of this process occurring—are assumed to be essentially the same in solids and gases.

*Corresponding author

 basil.gonsalves@helsinki.fi (B.C.G.)

ORCID(s):

However, in gases i) there are no free electrons like in solids, ii) there are no sharp transitions at the surface of the medium as seen in solids, and iii) the distances between collisions are orders of magnitude larger because of the low density of gases.

In a gas stripper, there is no sharp transition at the surface as in the case of solids, but in both cases, when a molecular projectile loses electrons in the first few layers of the stripper medium, it can break-up into charged fragments. This is known as the "Coulomb explosion". As a result, the created charged atoms are pushed apart by the strong electrostatic repulsion [3]. The Coulomb explosion can lead to beam losses that are more pronounced in solid stripper but the use of gas stripper in addition reduces the effect [6]. If an atomic projectile passes through a gaseous stripper medium the average charge state is lower than in a solid stripper medium due to the collision frequency in solids being several orders of magnitude larger than in gas. This is explained through the density effect [7] caused by the influence of excited projectile states that are formed when the projectile atom collides with the target medium. In a solid stripper, the excited projectile states may not have decayed into strongly bound states before experiencing subsequent collisions, leading to an enhancement of the electron loss cross section [1].

At high energies, an important technical application of electron stripping is to use solid stripper foils to produce high-intensity, high-charge-state ion beams for heavy ion accelerators. The efficient use of gas-filled heavy-ion spectrometers relies on knowing the equilibrium charge distribution of nuclear reaction recoils at similar energies. In the intermediate energy regime, gas strippers are used in accelerator mass spectrometry and the understanding of the different behaviour of atomic and molecular projectiles during the charge exchange processes are relevant.

In the context of Be AMS it is desirable to have as high as possible beam transmission as well as a high charge state yield. The main factors which contribute to beam transmission loss are 1.) Coulomb repulsion leading to beam emittance growth and 2.) beam scattering with stripper media in the charge exchange canal. Another factor possibly affecting beam transmission and yield of selected ion is the nature of the injected ions. It is known that break up with molecules in a stripper canal induces energy and angular spread in the beam [6]. However in ^{10}Be AMS the only option is to inject beryllium in a molecular form as the Source of Negative Ions by Cesium Sputtering (SNICS) ion source typically only produces nA of Be^- beams, which is too low to be useful. Oxygen was selected to study the effect of injecting atomic or molecular ions on charge state distributions after the stripper. Measuring different yields for injecting atomic or molecular oxygen ions might indicate similar behaviour for Be and can be used to indirectly compare expected yields from injecting Be as atomic ion, or in the form of BeO^- and BeF^- .

Unfortunately, all the previous measurements of O charge state distributions in gases employed atomic ions only [8, 9, 10]. The great majority of published experimental data on charge exchange cross sections are on neutral or single charged ions at relatively low energies ($E/M < 1$ keV/u) or on multiple charged ions at high energies ($E/M > 1$ MeV/u), and data from the intermediate energy regime (10 keV/u $< E/M < 1$ MeV/u) is more

scarce. There are a few interesting points of comparison in the literature regarding the molecular effect on charge state distribution. Maor et al. [3] reported the charge distribution shifting towards lower charge states for molecular ions as compared to atomic ions. This was explained in terms of enhanced electron capture probability. Brunelle et al. [4] measured the charge distributions of carbon clusters and compared them to carbon ions. One of the findings of Brunelle et al. [4] suggest that the average charge state of carbon cluster constituents is much less than that of single carbon ions of the same velocity. This was explained through the enhanced electron capture for closely spaced ions.

Kumar et al. [5] reported a difference in the stripping process for molecular beams as compared to atomic beams. Their study claims there can be some kind of interference due to the presence of the molecular ion passing through the stripping foil. By using two foils and using molecular beams they found the yield of the higher charge states could be increased. It is important to note that [5] although reported higher yields of higher charge states, used two foil strippers in order to mitigate the vicinage effect which reduces high charge state yields. By using the first stripper they get low positive charge states and by subsequently passing the beam through a second stripper, they can access higher charge states.

Following on previous study on beryllium charge states [11], oxygen charge state distributions are measured in argon gas. Measurements were performed at terminal voltages of 1.5 – 5 MV. In the results section, comparisons are made both to the literature data and predictions of the Convolution approximation for swift Particles (CasP) code. We also investigated if there is a molecular effect on the charge state distribution depending on the nature of the injected particles. If any molecular effects are present, such as a shift in the charge state distribution in one direction, one could potentially exploit this to increase the yield of higher charge state ^{10}Be . The measurement of oxygen charge state distributions is useful as of itself. These distributions are the most interesting aspect of this work. They provide new values to the previously available scarce data. Apart from their use in gas-filled heavy-ion spectrometers used in nuclear physics, they are useful for Be AMS facilities as measurement of $^{17}\text{O}^{+5}$ yield is used to determine stable ^9Be yield by injecting $^9\text{Be}^{17}\text{O}^-$ simultaneously.

2. Experimental methods

Measurements were conducted using the 5-MV tandem accelerator equipped with a gas stripper at the University of Helsinki [11]. The first part of the system is a Multi Cathode Gas Feed Source of Negative Ions by Cesium Sputtering (MCGSNICS) ion source [12], with BeO rods (from Goodfellow Cambridge Ltd) fitted in Al cathodes. The same BeO sample was used in all measurements in order to suppress possible sample-dependent effects, but the ion source can adapt up to 39 samples in the sample wheel if necessary. Extracted negative ions (BeO^- , O_2^- and O^-) are pre-accelerated to 70 keV and guided through an electrostatic analyser, focused and injected into the accelerator with a 90° injector magnet. Ar gas was used as stripper gas and was recirculated by a turbopump at half of the nominal pumping

speed of 150 L s^{-1} . In order to compensate for the escape of stripper gas to the accelerating tubes and to keep the gas pressure constant during the measurements, argon gas is continuously added via a Pulse Width Modulation (PWM)-controlled thermomechanical valve to the returning gas flow which is fed to the middle of the stripper canal. The canal has an inner diameter of 8 mm and is 546.5 mm long. After passing through the accelerator stage, the beam of positive ions is focused by an electrostatic quadrupole doublet and momentum-to-charge analysed in a 90° analysing magnet. In order to ensure charge state equilibrium at the stripper canal, the $^{16}\text{O}^{+3}$ current was monitored and maximised in terms of Ar gas pressure in the stripper canal.

Ion charge state distributions were obtained by measuring the beam intensity of each selected charge state separately after the analysing magnet. Beam profilometers and Faraday cups played a central role in measuring beam currents. A schematic of the facility showing positions of the Faraday cups and the BPMs as well as other components used in this investigation can be seen in Figure 1 of [11]. The measurement procedure consisted of the injected negative current being measured by the beam profilometer (BPM_IM) and simultaneously the desired charge state (say +1) on the high energy side was measured by a Faraday cup (FCSW). To move on to the next charge state a change in the high energy side components was done to measure the next charge state (say +2). This is then followed by measuring the injected negative current (i.e. from the low energy side) again by the beam profilometer and the next charge state (say +2) by Faraday cup (i.e. from the high energy side) simultaneously. This process was repeated until all desired charge states were measured. This method was advantageous to us as it allowed simultaneous measurement of beam current without blocking the beam with a Faraday cup at the low-energy side. BPM_IM calibration was done against FCIN, so that we could measure currents with the beam profilometers. Atomic oxygen charge states from +1 to +5 after the stripper were measured from injected BeO^- , O^- , and O_2^- . Beam Profile Monitors (BPMs) before and after the main acceleration were used to observe the beam position and focus as well as to monitor the stability of the beam current. While the low energy beam stayed stable with constant current during the experiments, in each subsequent change of selected charge state, the beam focus at the high energy side was checked. All the experimental data and run parameters we used during our measurement on our tandem are stored in our LabView database and can be retrieved at any time. Also the values used for our BPM calibration are traceable from the same database.

2.1. Data evaluation

In order to obtain the yields, currents of beams which passed through the injector magnet I_{IM} ($q = -1$) and analysing magnet I_{AM} ($q = +1, \dots, +5$) were measured by Faraday cups. These were then converted to absolute yields $Y_A = (I_{AM}/I_{IM} \cdot q) \cdot 100$. The relative yield Y_R for each charge state is then calculated as the ratio of the absolute yield to the sum of yield of charge states from +1 to +5 measured at a particular terminal voltage and normalising the

values to 100%. i.e. $Y_R = (Y_A / \sum_q Y_A) \cdot 100$. In the case of molecular oxygen O_2^- the absolute yields had to be divided in half because O_2^- consists of 2 oxygen atoms.

Various parameterised and semi-empirical models exist to calculate charge state yields for interaction with solid and gas targets [13, 14]. The model predictions are usually produced either i) by parameterising the mean equilibrium charge state and the width of the distribution as a function of the projectile energy E_p and charge number Z_p , and of the charge number Z_t of the medium, or ii) by calculating first the charge exchange cross sections—for electron loss and electron capture—that determine the charge state distributions via the rate equations. In the first method, the model parameters are determined by fitting to existing experimental data on equilibrium charge distributions. For example, the CasP code implements a widely used semi-empirical parameterisation for both gas and solid media [13]. In the second method, although the cross sections can be calculated from first principles for some simple systems or at very low or very high energies, it's most often required to use simplifying approximations to parameterise and fit the cross sections to experimental data [15].

The information on the equilibrium charge state distribution is thus contained in two parameters: the equilibrium average charge state ($\bar{q}$) and the width of the charge state distribution ($\Delta_{\bar{q}}$). The equilibrium average charge state ($\bar{q}$) is calculated as a weighted mean and the $\Delta_{\bar{q}}$ parameter is obtained as the standard deviation of the relative yields according to $\bar{q} = \sum_q q F_q^{\text{eq}}$ and $\Delta_{\bar{q}} = \left[\sum_q (q - \bar{q})^2 F_q^{\text{eq}} \right]^{1/2}$, where the equilibrium charge fractions F_q^{eq} are normalised to unity: $\sum_q F_q^{\text{eq}} = 1$. The parameter values can also be determined by fitting a univariate normal to each of the relative yield data (see Table 1).

In this investigation, we compare our results against predictions calculated with version 5.2 of the CasP code, input parameters being the incident particle Z_p (nuclear charge of the beam particles) and its energy as E/M and medium composition Z_t (nuclear charge of the stripper material). The current version of the CasP implements a slightly modified parameterisation to account for shell effects that are experimentally seen at slow velocities ($v \approx v_0$) and light projectile-target systems. The original semi-empirical parameterisation by Schiwietz and Grande [13] predicts the equilibrium average charge in gases according to

$$\begin{aligned} \bar{q} &= Z_p \frac{376x + x^6}{1428 - 1206x^{0.5} + 690x + x^6}, \\ \text{with} \\ x &= \left(v_p / v_0 Z_p^{-0.52} Z_t^{0.03 - 0.017 Z_p^{-0.52} v_p / v_0} \right)^{1 + 0.4 / Z_p} \end{aligned} \quad (1)$$

where v_p is the velocity of the projectile and $v_0 = 2.188 \times 10^6$ m/s is the Bohr velocity.

In practice in order to generate the predicted curves of CasP as a function of energy (cf. Figure 2), it is easier to write a python script using the above Equation 1 along with defining a Gaussian function as well as functions to convert units, so that so that all parameters can be extracted for a given energy range in one go. This would be easier than running CasP every time for each energy.

3. Results and discussion

Measurements were performed to study the charge distribution of oxygen ions injected from molecular BeO^- in Ar stripper gas. This can be seen in Figure 1, which are depicted in terms of absolute yields. O transmission of the accelerator was normalised to injector particle current during the measurement which was around 80% for Ar stripper gas. Transmission dropped steeply for terminal voltages below 2 MV. At 2 MV, we have a transmission of around 80%, but at 1.5 MV we have a transmission around 15%. This is a dramatic drop in transmission when compared to [11], which presents the transmission of Be ions. In [11] there is a more gradual drop in transmission from around close to 80% at 5 MV to around 45% at 1.5 MV and a drop down to 12.5% at 1 MV. Present measurements for O ions transmission holds at around 80% from around 5 MV to 2 MV and then we see the step-drop. We didn't expect to see such a drop for O ions; we expected a similar behaviour (i.e. a more gradual decline from higher terminal voltages and not just a step-drop from 2 - 1.5 MV) as is the case for Be ions as both ions come from the same precursor molecule so it should have the same behaviour. A possible explanation for this can be [16] at low energies we have inefficient stripping resulting in many negative and a few neutral ions which are not transmitted; the charge exchange process remains unbalanced and net charge stays low; or more seriously beam optics can get worse at low energies resulting more angular scattering and spread of the beam (divergence) at these energies resulting in poor focusing and beam losses before detection. We can get an estimation of the amount of neutral particles in Figure 2 (pink dashed from CasP and dashed-dotted from Schiwietz and Grande). It is around 10% at 1 MV and decays inversely and reaches almost nothing at 2.5 MV. At 1.5 - 2 MV is estimated to be less than 5%. Another possible explanation for our observation can be due to the fact that we have injected a molecular species, there is a Coulomb explosion [6]. It also creates an increased energy spread and angular spread resulting in beam losses and reduced transmission. This is more prominent at low terminal voltages where we have a larger energy spread; larger angular spread, which is dependent on ion velocity so, lower velocity will lead to larger angular deviations and less phase-space compression resulting in reduced transmission. At higher terminal voltage we have higher ion velocity, so less energy spread and angular divergence leading to higher transmission. The Coulomb explosion is also stated [17] to be the cause of significant loss of transmission of Be (or equivalently O) from BeO^- in both foil and gas strippers. Maximum yield of just under 50%, was reached at 2 MV terminal voltages when the +2 charge state was selected. As a visual guide, so one can follow how absolute yield changes as a function of terminal voltage, CasP curves have been convoluted to the transmission, using a Fermi-Dirac

Table 1

Equilibrium charge fractions F_q^{eq} (in %) of oxygen charge states from injected atomic and molecular negative ions after argon gas stripper. Entries in the last two columns show the parameters of a univariate normal distribution $\mathcal{N}(\bar{q}, \Delta_{\bar{q}})$ fitted to the data. For comparison, the last three rows include previous experimental data [9, 8, 10].

Injected Ion	Terminal voltage U_T (MV)	E/M (MeV/u)	Positive ion charge state q					Mean $\bar{q}$	std $\Delta_{\bar{q}}$
			1+	2+	3+	4+	5+		
${}^9\text{Be}{}^{16}\text{O}^-$	1.50	0.06	29.35	54.07	14.44	2.14		1.89	0.72
	2.00	0.08	24.29	58.05	17.66			1.93	0.64
	2.50	0.10	17.10	49.48	31.75	1.67		2.18	0.72
	3.00	0.12	11.76	40.72	35.23	11.51	0.79	2.49	0.87
	3.50	0.14	8.54	28.53	44.19	17.30	1.44	2.75	0.89
	4.00	0.16	4.54	26.79	45.61	23.07		2.87	0.81
	4.50	0.18	2.87	21.13	43.98	28.92	3.09	3.08	0.86
	5.00	0.20	2.59	15.36	41.43	34.52	6.09	3.26	0.88
${}^{16}\text{O}_2^-$	4.00	0.13	14.83	44.76	35.91	4.49		2.30	0.77
${}^{16}\text{O}^-$	1.92	0.12	8.65	38.35	38.70	13.75	0.55	2.59	0.85
1	2.50	0.156	0.71	35.55	47.17	16.58		2.80	0.71
2	n/a	0.203		12.0	38.0	39.0	11.0	3.49	0.84
3	n/a	0.216	5.8	23.2	43.9	24.0	3.1	2.95	0.91

¹ Injected ion ${}^{16}\text{O}^-$ on O_2 stripper gas (charge equilibrium uncertain) [9].

² Accelerated ${}^{16}\text{O}^{q+}$ ions on external Ar gas stripper (charge equilibrium) [8].

³ Accelerated ${}^{16}\text{O}^{q+}$ ions on external He gas stripper (charge equilibrium) [10].

step function, which models our transmission data points quite well. In terms of highest absolute yields we can see that from 1 MV - 3.2 MV + 2 charge state has the highest absolute yield and from around 3.2 MV - 5.5 MV + 3 has the highest absolute yield. We measure this change over at around 3.2 MV. This information would be useful for anybody needing to know the highest yielding oxygen charge states over said terminal voltage range. The main source of uncertainty in the absolute yields is the uncertainty (in the form of standard deviation) of the beam current measured in the Faraday cups. This is quite often due to the non-linearities of the measurement instrument in use. The average uncertainty for the transmission is 2.21% and for the measured charge state is 0.90%. The errors quoted here are relative uncertainties.

In Table 1, the relative yields of oxygen charge states from injected BeO^- ions are shown. Measurements are made from 1.5 MV to 5 MV in steps of 0.5 MV. In the last two columns, the mean and standard deviation of the resultant Gaussian distribution are presented, for a given terminal voltage. The Gaussian distribution is central to charge state distributions and represents the equilibrium of the many-step charge exchange process. From the measurements of 8 terminal voltages of BeO^- , we can see the mean charge state increases with increasing ion energy. The width of the Gaussian distributions for BeO^- show no trend apart from some small changes. In the 9th and 10th rows in Table 1, we have the relative yields of oxygen charge states from injected O_2^- ions and O^- ions at 4 MV, and 1.92 MV respectively. They will be discussed in more detail in Figure 3. The last 3 rows in this table are from literature values.

In Figure 2, we present the relative yields of oxygen charge states as a function of terminal voltage (i.e. energy) just from injected BeO^- ions. All measurements were carried out at a stripper gas pressure that was high enough for reaching equilibrium charge state distribution, so that our relative yields for each charge state q approximate rather well the limiting equilibrium charge fractions F_q^{eq} of theoretical models by CasP and Schiwietz and Grande [13]. A full error analysis has been done for all experimental data points. The error bars can be seen for all our measured points in Figure 2. The uncertainty in the relative yields is then obtained by error propagation calculation and is presented in percentages relative to the values in Figure 2. The average uncertainty in relative yield for O charge states measured from BeO^- over all terminal voltages was 1.55 %. All percentage uncertainties stated here refer to relative uncertainty.

Considering Figure 2, starting with the +1 measured data points (red), we can see at 1.5 MV our data point matches that of Schiwietz and Grande. All measurements above 1.5 MV tend to agree more with the predictions of CasP. For +2 measured data points (green), our 2 MV data point is above both predictions, with the rest tending to agree with CasP within error bars. However there are some exceptions such as a 3.5 and 5 MV, our measurement points seem to be closer with that of Schiwietz and Grande. For +3 (blue) at 1.5 MV we have agreement with Schiwietz and Grande, above that there is agreement within error bars with CasP, but for 4.5 and 5 MV there is more agreement with Schiwietz and Grande. For +4 (yellow) we have agreement with Schiwietz and Grande for all data points except 2.5 MV, which is well below both predictions. Also one should note that from 1.5 - 2.5 MV relative yields are close to 0 and so are not particularly significant. For +5 (light blue) relative yields are quite low until above 4.5 MV, yields are below 5%. At 5 MV our point agrees with Schiwietz and Grande, however the yield is around 6%.

In general, we can see that at low energies there is more agreement with Schiwietz and Grande, and at higher energies there's more agreement with the semi-empirical parametrisation in the CasP code. A possible explanation for this can be the low transmission at low terminal voltages. However one should bear in mind that at low energies we have higher uncertainties and there are extremely low transmissions especially below 2 MV (i.e. refer to Figure 1). Reasons for the low transmission are given with regards to the discussion of Figure 1. After 2 MV we obtain flat top transmission, which means from 2 MV and above we can be sure that the true and accurate charge state distribution is being measured. Also we see in the cases of +4 and +5 more agreement with Schiwietz and Grande. There is a different formula used by CasP and by Schiwietz and Grande [18], so both of them are calculated slightly differently.

In addition, comparisons are made against literature values [8, 9, 10]. The experimental setups of these measurements differ from each other and from the current one: Hubbard and Lauer used a separate Ar gas stripper after the accelerator and Preikschat and Cramer [9] used oxygen gas stripper in a tandem accelerator. Liu et al [10] used a separate He gas stripper after the main accelerator. Preikschat and Cramer [9] 's data points match ours for +3. For +2 and +4 they don't. For +1 they are close to 0 and hence meaningless. It's worth noting that all the points in our relative figure have been scaled to 100%, so if something is overestimated and it drops all other points will go up.

Hubbard and Lauer measured oxygen charge distributions with an external Ar gas stripper (length 508 mm, diam. 3.175 mm) at an equilibrium target thickness. In terms of their data points +2 is below ours, and +5 is above ours. We can say that the difference is small between their points and ours and considering the fact that the yields are also quite low, it is possible that our points are not too far apart. Also their +4 is above ours and our +3 is above theirs. Although again they are quite close so it might almost be the same result. To our knowledge, this is the only previous experiment at similar conditions to ours reported in the literature. These authors observed that the equilibrium charge distribution is independent of the used stripping gas for oxygen and neon beams at energies of about 7.1 MeV. While this may not be true in general and especially at lower energies, we assume it meaningful to compare our results also with the data from Ref. [9] although the latter were obtained by using oxygen gas stripper. However, their results may have suffered from beam losses caused by angular straggling because the yields of higher charge states do not saturate to a constant level when the gas pressure is increased. The charge distribution data probably do not correspond to charge equilibrium. The use of a separate, external gas stripper by Hubbard and Lauer seems more beneficial in this respect.

In a recent experiment, Liu et al [10] studied oxygen charge state distributions in a windowless He gas target of effective length of 44 mm and observed that a target thickness of about 20×10^{16} atoms/cm² was required for charge equilibrium at 7.2 MeV oxygen ion energy. Looking at their data we see, it is around 5.5 MV, which is above our measurement range. +3 agrees quite well with CasP. +2 and +4 are bad and don't even agree with CasP. +1 and +5 are also bad, the ordering and ratios are inverse, and don't agree with CasP, however the relative yields are quite low. In general we see large deviations between their measurements and the predictions however they use a different stripping gas i.e. He. Their theoretical analysis employs a plane-wave Born approximation to calculate the ionisation cross sections. This is a first order perturbation theory. To calculate the electron capture cross sections they use a continuum distorted wave approximation. This theory is used to describe heavy-particle collisions. They found out that the electron capture cross sections are reduced as density of the gas target increases. In a dilute gas, excited electron states of the projectile decay before the atom experiences the next collision, while in a dense gas a further collision may result in ionisation. The effect is even more pronounced at lower energies, which might account for larger deviations of experimental measurements from predictions in that energy range.

Figure 3 shows relative yields as a function of the oxygen charge states measured from BeO⁻ (red square), O₂⁻ (blue square), and O⁻ (green square) at terminal voltages 3 MV, 4 MV, and 1.92 MV, respectively. In Figure 3 the shape of the measured charge state distributions follow that of a Gaussian as predicted by Betz [2]. Solid lines represent Gaussian fits to experimental data points, using a Scaled Levenberg-Marquardt algorithm. The terminal voltages for BeO⁻ and O₂⁻ were chosen so that in each case the energy of O after molecular break-up at the terminal is comparable to atomic O⁻, i.e., equal to 0.12 MeV/u. Comparisons are made with predictions of the CasP code for oxygen ions impacting Ar gas (black line) at an energy of 120 keV/u which corresponds to the terminal voltage of atomic O⁻ at

Table 2

Gaussian fitting parameters of the four distributions studied

fit	mean	standard deviation
CasP	2.422	0.825
BeO ⁻	2.4884 ± 0.0007	0.8732 ± 0.0780
O ⁻	2.5920 ± 0.2127	0.8499 ± 0.0319
O ₂ ⁻	2.3006 ± 0.1020	0.7725 ± 0.0379

1.92 MV. The purpose of this figure is to see whether there is any noticeable change in the charge state distribution of oxygen ions depending on the precursor i.e. the originating species being BeO⁻, O₂⁻ and O⁻. This is known as the "molecular effect".

An important aspect in being able to detect any molecular effect is to quantify the uncertainties of the measured points. We have calculated the statistical uncertainty on all measured points in Figure 3 these can be seen in the form of error bars. In order to get the maximum extent of the uncertainty we have fitted two additional curves to our Gaussian fit of the experimental data points. Therefore for each Gaussian fit of the experimental data point set at a particular terminal voltage, we have an upper fit and a lower fit. The upper fit consists of fitting the points of the first two upper limits of the error bars with the second two lower limits of the error bars. The lower fit consists of the opposite which means fitting the first two lower limits of the error bars with the second two upper limits of the error bars. For each charge state distribution we really only have four data points to work with (+5 is close to zero). By carrying out this procedure we are able to overestimate our uncertainty on the means and standard deviations of our experimental data sets. By taking the differences of the upper and lower fits we are able to determine the errors on the means and standard deviations of our experimental data Gaussian fits. Although this method overestimates the error, it strengthens the validity of our result. The results can be seen in Table 2.

Table 2, is presented pictorially in Figure 4. Here we can see that BeO⁻ has the lowest uncertainty followed by O₂⁻ and lastly O⁻ which has the greatest uncertainty. The means are closest between CasP and BeO⁻, which is surprising because CasP calculates charge state distributions for atomic species not molecular species, CasP should be closest to O⁻. It is still possible that is still the case because the error bars for O⁻ are quite large, so that the means can fall within error range. One can also see there is a difference in means between O₂⁻; BeO⁻ and CasP; and beyond the error bars of O₂⁻. This indicates that there is a difference in charge state distribution even after overestimating our uncertainty. We would also like to point out that +2 of O₂⁻ in Figure 3 is far more pronounced and is the highest of all measured yields. Also comparing the width of these three distributions (c.f. Table 1), we can see that the width of the O₂⁻ distribution is narrower at 0.77, whereas the width for BeO⁻ is 0.87 and the width for O⁻ is 0.85, which are quite similar. We believe these are the contributing reasons why the O₂⁻ charge state distribution is different to the rest. The large error bars in O⁻ make it difficult to say whether there is a molecular effect in that case. All we can say is there

could be something. A possible explanation for what we are seeing could be that the gas pressure was set to give an equilibrium charge state distribution for BeO^- and the same pressure was used in all measurements.

Now it is worth taking a step back and looking at our results in light of earlier literature. References [3, 4, 5] which are earlier alluded to concern molecular breakup in a solid stripper foil. Their findings suggest that molecular ions result in lower mean charge states due to increased electron capture as compared to atomic ions, although [5] suggested it is possible to control this by using multiple strippers. In this sense our results agree with what is in literature as O^- has the highest mean charge state followed by BeO^- and finally O_2^- . Of course in our case as we have an argon stripper gas and not a solid carbon foil, in gaseous strippers we can assume the distance between stripper atoms is so huge and the interaction point distances are also huge that any electrons released during the interaction process (under equilibrium we have a balance between electron capture and electron loss) does not have an effect on the final charge state distribution. In stripper gasses we may also assume when injected molecules are broken the fragments are so far from each other that they don't interact with each other. However in the study of Fifield et al. [17] reports that the Coulomb explosion is less for gas stripping as compared to solid strippers, but "it is still present". Indeed they do present results where $^9\text{Be}^{3+}$ of BeO^- origin experiences broadening of energy distribution when compared to similar ion of atomic origin in N_2 stripping gas. We should also consider the uncertainties on O^- and O_2^- and possible differences between the calculation of CasP and experimental reality.

4. Conclusion

In this study, oxygen charge state distributions from a tandem accelerator using Ar stripper gas at terminal voltages of 2 – 5 MV were measured. Negative ions when oxygen was injected in the accelerator of both atomic (O^-) and molecular (BeO^- , and O_2^-) form were used. Absolute yields were measured along with total transmissions. For the measured range, transmission holds at around 80% except for between 2 - 1.5 MV where we observe a step-drop to about 15%. It was found out for oxygen ions transmission drops steeply between 1.5 and 2 MV. +2 is the most abundant oxygen charge state up to 3.2 MV after which +3 becomes the most abundant charge state. There is a difference in charge state distribution between O_2^- ; BeO^- and CasP, however the error bars on O^- are quite large thus based on the data presently available we cannot make a conclusive general statement on the molecular effect in a gaseous stripper except to say there can be some effect. The literature says there is a molecular effect for solid strippers however for gaseous strippers this effect is much less though it is still observable. This is significant as we sought to find out if the origin of the oxygen ions has an observable effect. This can be valid for all types of ions that may be measured by AMS. Though molecular ions may still affect non-equilibrium charge distributions, especially in solid strippers.

We obtained a series of symmetric Gaussian distributions as it is expected from charge state distribution theory. Relative yields of charge states $q = +1$ to $+5$ were determined and compared to semi-empirical predictions of CasP

code and the Schiwietz and Grande formula. In general, the measured values follow the same trend as the predicted values. Although we obtained a good general agreement within uncertainty, there were certain regions where agreement was better with Schiwietz and Grande and others where agreement was better with CasP. However, we still believe that CasP is a good tool for predicting and planning future experiments. The main factors affecting the charge state yield are the velocity of the ion and the interactions between the electrons of the ion and the electrons of the stripper medium.

References

- [1] Hans D Betz. Review of stripping of heavy ions. *IEEE Trans. Nucl. Sci.*, 18:1110–1114, 1971.
- [2] Hans-Dieter Betz. Charge states and charge-changing cross sections of fast heavy ions penetrating through gaseous and solid media. *Reviews of Modern Physics*, 44(3):465, 1972.
- [3] D Maor, PJ Cooney, A Faibis, EP Kanter, W Koenig, and BJ Zabransky. Charge-state distributions of nitrogen ions resulting from the foil-induced dissociation of 4.2-MeV N^{2+} ions. *Physical Review A*, 32(1):105, 1985.
- [4] A Brunelle, S Della-Negra, J Depauw, D Jacquet, Y Le Beyec, and M Pautrat. Reduced charge state of MeV carbon cluster constituents exiting thin carbon foils. *Physical Review A*, 59(6):4456, 1999.
- [5] Pankaj Kumar, Gunther Korschinek, S Chopra, T Faestermann, P Ludwig, G Rugel, D Seiler, Anton Wallner, S Ojha, S Gargari, et al. Charge state distribution studies of SrF_3 , MnF_3 and CaF_3 molecules using single and double stripping in a Tandem accelerator. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 269(18):1986–1991, 2011.
- [6] Martin Suter, L Keith Fifield, and Sascha Maxeiner. The impact of the break-up of molecules in the stripper on the AMS performance. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 533:96–103, 2022.
- [7] H Okuno, N Fukunishi, A Goto, H Hasebe, H Imao, O Kamigaito, M Kase, H Kuboki, Y Yano, S Yokouchi, et al. Low-Z gas stripper as an alternative to carbon foils for the acceleration of high-power uranium beams. *Physical Review Special Topics-Accelerators and Beams*, 14(3):033503, 2011.
- [8] Edward L. Hubbard and Eugene J. Lauer. Charge distributions of oxygen and neon ions passing through gases. *Physical Review*, 98(6):1814–1817, 1955.
- [9] E Preikschat and JG Cramer. Charge Distribution of ^{16}O Ions Produced from Tandem Accelerator. *Review of Scientific Instruments*, 37(12):1722–1726, 1966.
- [10] Shengjin Liu, Makoto Sakurai, Kenshi Sagara, Takashi Teranishi, Kunihiro Fujita, Hiroyuki Yamaguchi, Sayaka Matsuda, Tatsuki Mitsuzumi, Makoto Iwazaki, Mariya T. Rosary, Daiji Kato, and I.Yu. Tolstikhina. Charge state distribution of ^{16}O from the $^4He(^{12}C, ^{16}O)\gamma$ reaction of astrophysical interest studied both experimentally and theoretically. *Nuclear Instruments and Methods in Physics Research B*, 328:14–19, 2014.
- [11] Basil C Gonsalves, Kenichiro Mizohata, Pertti O Tikkanen, and Jyrki A Räisänen. Charge state optimisation for beryllium accelerator mass spectrometry. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 469:33–36, 2020.
- [12] P Tikkanen, V Palonen, H Jungner, and Juhani Keinonen. Ams facility at the university of helsinki. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 223:35–39, 2004.

- 356 [13] G Schiwietz and PL Grande. Improved charge-state formulas. *Nuclear Instruments and Methods in Physics Research Section B: Beam*
357 *Interactions with Materials and Atoms*, 175:125–131, 2001.
- 358 [14] G Schiwietz, K Czerski, M Roth, F Staufenberg, and PL Grande. Femtosecond dynamics—snapshots of the early ion-track evolution. *Nuclear*
359 *Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms*, 225(1-2):4–26, 2004.
- 360 [15] B Friedman and G DuCharme. Semi-empirical scaling for ion–atom single charge exchange cross sections in the intermediate velocity regime.
361 *Journal of Physics B: Atomic, Molecular and Optical Physics*, 50:115202, 2017.
- 362 [16] Meetha Lal Meena. Analysing the charge distribution and transmission of carbon and oxygen in nitrogen gas using tandem accelerators. *Indian*
363 *Journal of Pure & Applied Physics (IJPAP)*, 63(8):707–712, 2025.
- 364 [17] Leslie K Fifield, Martin Suter, Michaela B Froehlich, Dominik Koll, Stefan Pavetich, Zuzana Slavkovská, Stephen G Tims, and Anton Wallner.
365 Coulomb explosion of beo- molecular ions—revisited. *Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions*
366 *with Materials and Atoms*, 538:205–211, 2023.
- 367 [18] Private Communication.




Figure 1: Absolute yield measurements [as a % of total injected beam] for O charge states +1 – +5 measured from BeO, as a function of terminal voltage in MV. Legend – Measured: ■ – Total transmission; ● – +1; ● – +2; ● – +3; ● – +4; ● – +5. Error bars shown refer to statistical uncertainty, not machine error. CasP convoluted: – – 0; – – +1; – – +2; – – +3; – – +4; – – +5.




Figure 2: Relative yield measurements [as a % of total injected beam] for O charge states +1 – +5 measured from BeO. Legend – Measured: ● – +1; ● – +2; ● – +3; ● – +4; ● – +5. Error bars shown refer to statistical uncertainty. Experimental data for atomic O⁻ are shown for comparison purposes. CasP calculations: – – 0; – – +1; – – +2; – – +3; – – +4; – – +5. Schiwietz and Grande [13] calculations: – – 0; – – +1; – – +2; – – +3; – – +4; – – +5.




Figure 3: Comparison of oxygen charge state distributions, from oxygen of three different species (BeO^- , O_2^- , and O^-), all at the same energy. Legend – Measured: ■ – BeO^- , 3 MV; ■ – O^- , 1.92 MV; ■ – O_2^- , 4 MV; Error bars shown refer to statistical uncertainty. Calculated: — – CasP 3 MV $Z_p = 8$, $Z_i = 18$. Fit: — – BeO^- — – O^- — – O_2^- . Upper fit: ⋯ – BeO^- ; ⋯ – O^- ; ⋯ – O_2^- ; Lower fit: - - - – BeO^- ; - - - – O^- ; - - - – O_2^- .




Figure 4: Graphical depiction of the Gaussian fitting parameters for easy comparison