AUTHORIZATION			
	NAME AND TITLE	**SIGNATURE**	**DATE**
Issued by:	Candace Alva, Director RAQA	See M-Files	See M-Files
Approved by:	Adi Zmiri, Sr. Director of Project Management Office and Technical Consulting	See M-Files	See M-Files
Approved by:	Brian Bishop, Chief Operating Officer	See M-Files	See M-Files
Approved by:	Candace Alva, Director RAQA	See M-Files	See M-Files

REVISION HISTORY				
REV. #	**CO #**	**DESCRIPTION**	**PREPARED BY**	**DATE**
06.02	DCO00015	Original File	Brian Bishop	8/6/2020
06.03	DCO00047	Update PeriGen logo. Correct minor typos throughout. Update 5.1.4 to reflect appropriate responsibility. Added section 5.2 for room validation.	Candace Alva	6/25/2021

DISTRIBUTION LIST		
COPY #	**NAME**	**TITLE/ DEPARTMENT**
Original	Master File	Quality Assurance
1		
2		

©PeriGen, Inc. – Confidential and Proprietary

This document may not be copied without express written permission.

1. Purpose

The purpose of this document is to describe the procedures for product installations, product upgrades, facility expansions, room expansions, and product training.

2. Scope

2.1. The procedures described in this document apply to PeriGen products.

2.2. This procedure **does not** apply to products that do not meet the definition of a medical device at the target market or product that defined as medical device but is exempted from quality system.

3. Applicable Documents

Standards and Regulations:

ISO 13485:2016 – Medical devices-Quality Management Systems-Requirements for regulatory purposes, section 7.5.3.

FDA Quality System Regulations – 21 CFR Part 820.160, 820.170, and 820.200.

Procedures and Forms:

FQMS-7.5.1-002, DHR Documentation Log

QMS-8.5.1 - Customer Support and Complaint handling.

QMS-7.3.8 - Design Transfer and Release procedure

QMS-7.2.2 - Contract Review

QMS-8.3.0 - Control of Nonconforming Product

QMS-8.3.1 – Control of Nonconforming Material (Hardware)

QMS-6.2.1 Training and Qualification

OP-WI-001 - Maintaining Sites Pages

OP-WI-002 – Product Installation – Cary

OP-F-007 – Workstation Wallplate Config Template

OP-F-010 – PeriWatch Room Validation

4. Definitions and Abbreviations

Device Master Record (DMR) - A compilation of records containing the procedures and specifications for a finished device (production file).

Device History Record (DHR) - A compilation of records containing the production history of a finished device.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

TL – Technical Lead (includes technical leads and remote operation specialists)

CC – Clinical Consultant

PM – Project Manager

QA Manager – Manager**,** Quality Assurance

Go Live – a project phase when a system becomes operational

SSO - Supplemental Service Order

5. Procedure

5.1. Product Installation, Upgrade, and Expansion

5.1.1. Procedures for new product installations, upgrades, and expansions are included in the DMR for each version of software product.

5.1.2. Product Installations are performed for new clients. From time to time, based on analysis of Customer Complaints, Inquiries, Customer requests and/or insights from product development activities, existing clients may have their installed system upgraded with a newer version of the same product. All active clients are under a Software Maintenance Agreement which affords them access to any software product upgrades consistent with their current license agreement. Existing clients may also have an additional software product and/or an interface installed after the initial implementation as part of Facility or Room Expansions.

5.1.3. There shall be a signed contract or SSO for all product installations, upgrades, and expansions per **QMS–7.2.2**, *Contract Review* procedure prior to installation, upgrade, or expansion.

5.1.4. When a contract or SSO is signed for a project, the project shall be included in the Master Project Status Report maintained by operations. This Master Project Status Report shall indicate the status of the project throughout implementation.

5.1.5. Only products that were formally released for distribution shall be installed at the customer site. The TL is responsible to verify that the product installed is formally released for marketing in accordance with **QMS-7.3.8**, *Design Transfer and Product Release*.

5.1.6. All installations, upgrades, expansions and their associated verification steps may be performed onsite or remotely.

5.1.7. User training needed to ensure specified performance and safe use of the medical device will be provided to the customer during or shortly after

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

installation. User Training for PeriGen products is performed in accordance with **OP-WI-002**, *Product Installation - Cary*.

Training materials shall be formalized, approved and kept as part of the DMR.

Training and instruction of representatives, authorized agents, distributors, users and customers are given by employees who have been qualified to do so according to **QMS-6.2.1** *Training and Qualification***.** Each training session shall be recorded. Training could include emails of training materials or release notes, face-to-face training meetings, webinars, computer-based training (CBT), etc.

Records of training should include the following information:

- The name and details of the person being instructed
- The subjects given by the instructor and the number of hours provided
- Date and place of instruction
- Name and signature of the instructor

5.1.8. Customer and user training will typically engage a user or group of users currently employed by the customer, according to the customer's request and depending on user availability. There is no requirement to train all users: training of users who missed the PeriGen training session will be done by the customer, as will further training of new users, using peer-to-peer on-the-job training, instructor led training, or other methodologies at the customer's discretion.

5.1.9. Re-training or additional training activities shall be considered on a case-by-case basis for each project (i.e. new installation, upgrade, and/or expansion) and release (e.g. major software release that includes new functionality). The TL and/or CC will communicate with the customer to determine if additional training is necessary.

5.1.10. Installations, upgrades, and expansions may be performed for new clients and existing clients. Installation, upgrades, and expansions for PeriGen products shall be performed in accordance with **OP-WI-002**, *Product Installation – Cary*.

5.1.11. **When the project is completed by the TL and CC, the PM** reviews the project documentation and ensures all items have been completed. Once all items are confirmed to be completed, the PM submits the project documents no later than in thirty (30) days following project completion to the QA Manager for review and storage in the client's DHR. The QA Manager records the data in **FQMS-7.5.1-002**, *DHR Documentation Log*.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

5.2. **Room validations**

5.2.1. If a customer is adding new rooms to an existing PeriWatch system, it shall be documented on **OP-F-010**, *PeriWatch Room Validation* form.

5.2.2. The TL acquires the customer's signature on the document.

5.2.3. PeriGen support completes/updates **OP-F-007**, *Workstation Wallplate Config Template* to complete the documentation.

5.2.4. The TL submits the signed remote room validation form and the workstation configuration to the PM for inclusion in the DHR.

5.3. **Non-conformities identified**

5.3.1. Non-conformities identified in released products during the installation, upgrade, or expansion process are submitted by the TL or CC or PM with a description of the non-conformity issue in accordance with **QMS-8.5.1,** *Customer Support and Complaint Handling* procedure.

5.3.2. Problems identified after the software has been validated at the customer site are evaluated and documented as described in **QMS-8.5.1**, *Customer Support and Complaint Handling*, to determine the appropriate response.

5.3.3. For PeriGen supplied equipment only:

In the event that nonconforming product (software or cables) is identified during installation or after commercial distribution (product release), they shall be handled according to **QMS-8.3.0**, *Control of Nonconforming Product*.

In the event that nonconforming product (e.g. DAS unit) is identified during installation or after commercial distribution (product release), they shall be handled according to **QMS-8.3.1**, *Control of Non-Conforming Material (Hardware)*.

5.4. **Location of DHRs and Customers configurations and access related information**

5.4.1. DHRs for all customers are kept in a secure network location that is backed up regularly or in the approved document management system by Quality.

6. Responsibilities

6.1. The Sr. Director of Project Management Office and Technical Consulting has the overall responsibility for the implementation of this procedure.

6.2. Quality is responsible for performing final documentation review and approval and filing of the submitted documents in the appropriate client's DHR. Quality records the data in **FQMS-7.5.1-002**, *DHR Documentation Log*.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

Signatures:

Rejected document:	I reject this document		
Name:	**Candace Alva** **perigen.com\candace.alva**	Title:	
	Candace Alva		2021-09-09 14:25:41 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Rejected document:	I reject this document		
Name:	**Candace Alva** **perigen.com\candace.alva**	Title:	
	Candace Alva		2021-09-17 14:30:51 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Document approval:	I approve this document		
Name:	**Brian Bishop** **perigen.com\brian.bishop**	Title:	
	Brian Bishop		2021-10-01 14:51:03 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Document approval:	I approve this document		
Name:	**Adi Zmiri** **perigen.com\adi.zmiri**	Title:	
	Adi Zmiri		2021-10-10 21:38:25 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Document approval:	I approve this document		
Name:	**Andrew Krieman** **perigen.com\andrew.krieman**	Title:	
	Andrew Krieman		2021-10-18 13:11:50 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Document approval:	I approve this document		
Name:	**Candace Alva** **perigen.com\candace.alva**	Title:	
	Candace Alva		2021-10-18 20:15:42 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	