AUTHORIZATION			
	NAME AND TITLE	**SIGNATURE**	**DATE**
Issued by:	Brian Bishop, Chief Product Officer	See M-Files	See M-Files
Approved by:	Andrew Krieman, Director Technical Client Services	See M-Files	See M-Files

REVISION HISTORY				
REV. #	**CO #**	**DESCRIPTION**	**PREPARED BY**	**DATE**
NA	NA	No description of changes to previous revisions was maintained.	NA	NA
02.03	NA	Updated to match the new support call data base; modified complaint review	Pamela Haswell	Sept 17th 2012
03.00	006	QA-SOP-0017 - Customer Support was changed to QMS-8.5.1-QMS-7.3.7. Updates following periodic review and QMS transfer to the PeriGen Israel office have been included.	Tsvia Erlich, Regulatory Affairs Consultant	April 5th, 2013
3.01	040	1. Paragraph 5.5.4 – changed according to internal audit recommendation to mark complaint as resolved after its validation. 2. Minor grammar changes.	Idan Navias, Help Desk Team Leader	May 8th, 2014
04.00	078	Removed reference to EU requirements; updated link to current work instructions; emphasized personnel who are authorized to close enquiries, complaints and investigations; Support Engineer role was replaced with Technical Services Director; other typo fixes Updated to comply with ISO 13485:2016	Sharon Resnick, QA Manager	September 18th, 2018
04.01	081	Addition of CS-WI-002 Support call case creation and management	Natan Blanks	Jan 9, 2019

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

REVISION HISTORY				
REV. #	**CO #**	**DESCRIPTION**	**PREPARED BY**	**DATE**
4.02	4	Addition of CS-WI-003 Root Cause Analysis	Brian Bishop	27 Sept 2019

DISTRIBUTION LIST		
COPY #	**NAME**	**TITLE/ DEPARTMENT**
Original	Master File	Quality Assurance
1		
2		

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

1. Purpose

The intent of this procedure is to:

1.1. Ensure that Customer Inputs (enquiries, complaints, etc.) are handled in a timely manner and that they are properly handled, classified, documented and, when appropriate, investigated.

1.2. Ensure that:

1.2.1. Customer Complaints are identified and handled in a timely manner.

1.2.2. Analysis and resolution of the root cause is effectively performed.

1.2.3. Corrective and preventive action(s) are determined, when applicable.

1.2.4. Any decision not to open an investigation to determine the root cause of a complaint and/ or not to take corrective action in response to a complaint shall be justified recorded and authorized by the Director of Technical Services.

1.2.5. Notification to Regulatory Authorities and/ or recall (field action) is conducted in compliance with the requirements in relevant markets, when applicable.

1.2.6. Other actions as may be appropriate or warranted, are determined and implemented.

2. Scope

This procedure applies to any Customer Input, related to a regulated medical device released for marketing, received via the telephone support helpline and/ or any input received by Customer Support or any PeriGen representative through other channels (e-mails, during installation, visit to the customer etc.).

This procedure does not apply to PeriGen Products that do not meet the definition of a medical device in the target market or are exempt from QSR, including Customer Complaint requirements.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

3. Definitions and Abbreviations

Enquiry – A request for information about PeriGen software, assistance with a feature or function, or a notification of events such as a power failure or a planned down time.

Complaint – Any written, electronic, or oral communication that alleges deficiencies related to the identity, quality, durability, reliability, safety, effectiveness, or performance of PeriGen's marketed medical devices.

Complaint Investigation – An analysis conducted by PeriGen to determine the underlying process or product failure that resulted in a complaint.

Cleared Product – a product that was released for marketing.

Support Call Data Base (SCDB) – A software system or environment (such as project, data base, etc.), used to manage quality system issues including:

- Enquiries
- Complaints
- Internal investigations

Issue (Customer Input) – The generic name for a record in **SCDB**. In the context of this work instruction, an issue is an *Enquiry* or a *Complaint*. Each issue in the SCDB is uniquely identified by its number.

DHF – Design History File.

DHR – Design History Record.

4. Applicable Documents

Standards and Regulations:

ISO 13485:2003 – Medical Devices-Quality Management Systems-Requirements for regulatory purposes section 7.2.3, 8.2.1 and 8.5.1.
Health Canada Medical Devices Regulations, sections 57 and 58.
FDA Quality System Regulations – 21CFR Part 820.198(a) – 820.198(g)

Procedures and Forms:

QMS-5.6.1 - Management Review

QMS-7.5.1 - Product Installations and Upgrades

QMS-8.5.2 – Corrective and Preventive Actions.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

QMS-8.5.5 – Post-Marketing Surveillance.

Work Instructions:

RA-WI-005 – FDA Recall and Reporting Requirements.
RA-WI-006 – Canadian Recall and Reporting Requirements.
CS-WI-001 - Ticket priority classification.
CS-WI-002 – Support call case creation and management.
CS-WI-003 - Root Cause Analysis

5. Customer Support - General

5.1. Customer Access

5.1.1. All PeriGen customers have access to a support Help Desk 24 hours a day, 7 days a week. The Customer Input can be obtained via phone call, e-mail, fax, oral or written correspondence or any input from sales or service representatives.

5.1.2. The Help Desk representative receives basic customer information, product information, and the nature of the question or complaint (problem) and records the data in the SCDB, using the designated form.

5.1.3. The Help Desk representative evaluates the received information and addresses the Customer Input or requests additional information, as needed.

5.1.4. Other PeriGen representatives who become aware of any Customer Input refer that information to Help Desk via email at support@perigen.com, so that it can be handled through the complaint system.

5.1.5. Any request by the Customer should be referred by email to the Help Desk. Any feedback from installation or upgrades that had impact on the go live should also be referred to Help desk by email.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

5.1.6. All company employees shall obtain training for initial handling of Customer Input and on the importance of Customer Inputs.

5.2. **Support Call Data Base**

5.2.1. Each Customer Input (whether considered a complaint or enquiry) is recorded in the SCDB as soon as practical, as described in *CS-WI-002 Support call case creation and management*. Each case is identified with a unique SCDB number that is assigned automatically by the SCDB tool. A separate Issue is opened for each new incident at each customer site. Subsequent calls from the same customer site about the same incident should be noted on the original SCDB record. The Help Desk representative who receives the complaint determines when a new case is required for repeated occurrences of the same incident at the same site

5.2.2. Terminology used in the SCDB may differ for technical reasons, from the terminology used in this procedure. Specifically, some SCDB implementations might have reserved terms for specific operations such as attaching a different meaning to “resolve” to that used in this procedure. In such cases, the work instruction WI-CS-002 for the specific SCDB will clarify the terminology usage as appropriate.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

5.3. **Customer Inputs initial handling**

5.3.1. Once the Customer input is recorded in the SCDB, the Help Desk representative notifies the customer via phone or e-mail about the assigned Customer Input number and the Input summary.

5.3.2. The Help Desk representative evaluates the communication, classifies the issue as an enquiry or a complaint, and sets its priority as described in *CS-WI-001, Ticket priority classification*. The issue type is recorded in the SCDB.

5.3.3. The initial evaluation and prioritization shall be done as soon as possible but, in any case, no later than 24 hours from issue creation time.

5.4. **Enquiry Handling**

5.4.1. Data collected and recorded includes:

- Basic information about the caller and his/ her question.
- Product information.
- The actions taken to answer/ resolve the enquiry.
- The type of the enquiry.

5.4.2. An enquiry can be closed when the customer acknowledges that the question has been answered or the notification has been documented and the Help Desk representative has documented the action taken in the SCDB.

5.4.3. When needed, the Help Desk representative escalates the enquiry to the Technical Services Director and/ or Operations Department member who will address and resolve the issue.

Note that a Help Desk representative or the Technical Services Director is authorized to close enquiries.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

5.5. Customer Complaint Handling

5.5.1. The Help Desk representative collects the additional information required for recording the complaint, including details about the product, the software version and the problem reported. Data collected and recorded includes:

- Basic information about the caller and his/ her question
- Product information
- The complaint problem type
- The actions taken to resolve the complaint

5.5.2. The Help Desk representative reviews each complaint within 24 hours to determine if the problem could pose a risk to patient safety and IMMEDIATELY escalates any potential safety issue to the, Technical Services Director and the QA/ RA Director for further handling.

5.5.3. When needed, the Help Desk representative forwards the complaint to the Operations Department member who will address and resolve the issue.

5.5.4. A Complaint can be marked as resolved when the root cause of the reported problem has been resolved and the client notified of the resolution, and either the client or a Help Desk representative has verified that the system is working properly.

5.5.5. The Technical Services Director shall determine up to one working week (5 business days) after a Complaint is received, whether to conduct an investigation. The Help Desk team leader will oversee the implementation of all activities until completion and closeout.

5.5.6. The Complaint can be closed only when:

- The customer reported problem was resolved, and Recorded in the SCDB with Customer acknowledgement, or if no action is taken, the reason for inaction should be recorded.
- An Investigation has been completed (or a decision not to conduct an investigation was recorded and authorized) and the root cause is determined per **CS-WI-003**, *Root Cause Analysis*.
- In the case where an investigation was initiated, a CAPA was initiated or a decision not to open a CAPA was recorded and authorized)

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

Note that only the Technical Services Director or RA/QA Director is authorized to close complaints.

5.5.7. If the complaint involves a death or serious injury, or if death or serious injury appears likely if the problem occurs again, the QA/ RA Director evaluates the complaint to determine if the incident must be reported to relevant regulatory authorities.

5.5.8. If a Recall/ Field Safety Corrective Action/ Advisory Notice is required, it will be handled as per the Recall and Vigilance Procedures.

6. Complaint Investigations

6.1. Any complaint involving the possible failure of the device, or its labeling or packaging to meet device performance specifications shall be investigated unless such investigation has already been performed for a similar complaint and another investigation is not necessary.

6.2. An investigation shall be opened within 5 business days from obtaining the complaint. The Company aims to complete investigations within 30 days of the decision to initiate investigation, as applicable.

6.3. The Technical Services Director or his/her delegate shall complete the investigation record in the SCDB, including, when applicable, to whom the complaint is assigned for investigation or any specific instructions to the assignee.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

6.4. If no investigation is carried out, the reason must be recorded on the complaint record in the SCDB along with the name of the approver.

6.5. If investigation is required, the following applies:

6.5.1. The Technical Services Director will determine the nature and extent of the investigation, and to whom it is assigned to lead the investigation.

6.5.2. The person assigned to lead the investigation shall assign additional members from the various groups in PeriGen, such as clinical delegate, SW development, R&D, installation, QA/ RA, etc. to assist with the investigation. An investigation includes, as applicable:

- Customer Input details.
- Review of design and development records (DHF).
- Review of production records (such as DHR, installation record, risk analysis, labeling, etc.).
- Determination of other software versions or products that could be affected by the problem.
- Actions taken to resolve the problem.
- Investigation activities and testing.
- An evaluation of the impact of the problem on user and patient safety and on the effectiveness of the product.
- The need to conduct hazard analysis to evaluate potential impact of the problem and the outcome of the hazard analysis.

6.6. The team involved and all activities, including testing, evaluations, analysis and decisions taken shall be documented in the investigation record (SCDB).

6.7. The investigation results, including the analysis that has been conducted, test results and conclusions, will be recorded in the pertinent part of the investigation record.

6.8. If the investigation reveals that an external organization, such as subcontractor or distributor, is involved, the QA/ RA Director shall notify them in writing and assure that pertinent corrective actions are performed, according to the **QMS-8.5.2,** *Corrective and Preventive Actions* procedure. In addition, activities resulting from and/or related to the complaint (e.g., replacement of the defective product, contacts with the complainant,

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

etc.) should be properly recorded on the pertinent parts of the complaint record. All related correspondence, with the customer shall be attached to the investigation record.

6.9. The Investigation can be closed only when:

- All investigation activities have been documented in the SCDB.
- The root cause of the problem has been identified, per **CS-WI-003**, *Root Cause Analysis*.
- CAPA was initiated (or a decision not to open a CAPA was recorded and authorized).
- Results of the investigation have been reported to the customer, if requested.

6.10. The Technical Services Director is responsible to oversee Investigation handling and closing.

7. Corrective Action

7.1. Each complaint record, for which an investigation was conducted, includes a reference to any corrective/ preventive action taken or an authorized justification for not taking such action.

7.2. Corrective and Preventive action, as a result of the investigation shall be initiated by the QA/RA Director and handled according to **QMS-8.5.2**, *Corrective and Preventive Actions* procedure.

7.3. If no Corrective or Preventive action is taken, the reason shall be documented in the investigation record.

8. Trending and Reviews and Data analysis

8.1. Information from the complaint handling process provides input for analysis of data in other areas of the quality management system, including but not limited to product risk management, product monitoring, post marketing surveillance and management review.

8.2. Other potential sources of Customer Input include post marketing surveillance activities such as user groups, satisfaction surveys, journal articles, and information concerning similar products, as detailed in the **QMS-8.5.5**, *Post-Marketing Surveillance*) procedure.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

8.3. As part of preventive actions, the Technical Services Director will review the complaint records periodically to detect recurring failures and problems with specific products. Such cases will be further investigated. The analysis, analysis results and action taken (including CAPA) shall be recorded and kept in the complaint file.

8.4. The Technical Services Director shall prepare, at least once a year, a trend analysis of all complaints. This analysis shall be recorded, reported to and discussed during Management Review meeting. (See **QMS-5.6.1,** *Management Review*).

8.5. To monitor and analyze Support call frequencies and types, periodic trending reports are prepared by the Technical Services Director. These trend reports are presented during Management Reviews.

8.6. Enquiries do not require routine review, but are included in the support call trending reports presented for Management Review.

9. Responsibilities

9.1. The Technical Services Director is responsible for the implementation of this procedure.

9.2. The QA/RA Director is responsible for reviewing Customer Complaints that are safety related within 24 hours and initiate recall, field action or reporting to regulatory authority when appropriate. In addition, the QA/RA Director is responsible for reviewing and verifying that investigations were properly conducted and documented.

9.3. The QA/RA Director is responsible for initiating a suitable CAPA process when applicable.

9.4. All company personnel are responsible for informing the Support department of any complaint received.

©PeriGen, Inc. – Confidential and Proprietary
This document may not be copied without express written permission.

Signatures:

Document approval:	I approve this document		
Name:	**Brian Bishop** **perigen.com\brian.bishop**	Title:	
	Brian Bishop		2019-10-27 11:51:02 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	

Document approval:	I approve this document		
Name:	**Andrew Krieman** **perigen.com\andrew.krieman**	Title:	
	Andrew Krieman		2019-10-28 00:07:43 (UTC+00:00)
Electronically Signed in	PeriGen Advanced Perinatal Systems	Timestamp	