

Table of contents

\\Research

Investigators

Belleau

R01StressDyn

localizer AAHScout T1_MEMPRAGE_64ch slice_positioning 22 degree angle rfMRI_REST1_ME_AP rfMRI_REST_ME_PA_distortion SpinEchoFieldMap_AP SpinEchoFieldMap_PA EMIST1_ME_AP EMIST2_ME_AP rfMRI_REST2_ME_AP HMIST1_ME_AP HMIST2_ME_AP rfMRI_REST3_ME_AP t2_tse_dark-fluid_tra pd+t2_tse_tra resolve_4scan_trace_tra_p2_192 rfMRI_REST4_ME_AP rfMRI_REST_ME_PA_distortion SpinEchoFieldMap_AP SpinEchoFieldMap_PA

\\Research\Investigators\Belleau\R01StressDyn\localizer

TA: 0:19 PM: REF Voxel size: 0.5×0.5×10.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
TR	20.0 ms
TE	5.00 ms
Averages	1
Concatenations	7
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	20.0 ms
TE	5.00 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	40 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
Base resolution	256
Phase resolution	50 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
TR	20.0 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	7

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	Isocenter

Geometry - AutoAlign

Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	20.0 ms
Concatenations	7
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	50 %

Physio - PACE

Resp. control	Off
Concatenations	7

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	180 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\Research\Investigators\Belleau\R01StressDyn\AAHScout

TA: 0:17 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 2 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip angle	8 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off

System - Adjustments

Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

Sequence - Assistant

Mode	Off
------	-----

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	8 deg
Measurements	1
Time to center	7.5 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

\\Research\\Investigators\\Belleau\\R01StressDyn\\T1_MEMPRAGE_64ch

TA: 6:02 PM: REF Voxel size: 1.0×1.0×1.0 mmPAT: 2 Rel. SNR: 1.00 : tfl_me

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	176
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.00 mm
TR	2530.0 ms
TE 1	1.69 ms
TE 2	3.55 ms
TE 3	5.41 ms
TE 4	7.27 ms
Averages	1
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	2530.0 ms
TE 1	1.69 ms
TE 2	3.55 ms
TE 3	5.41 ms
TE 4	7.27 ms
Magn. preparation	Non-sel. IR
TI	1100 ms
Flip angle	7.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %

Resolution - Common

Slice thickness	1.00 mm
Base resolution	256
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	176
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.00 mm
TR	2530.0 ms
Multi-slice mode	Single shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	R2.2 P1.7 F31.8
R	2.2 mm
P	1.7 mm
F	31.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	176 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2530.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
T1	1100 ms
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	100.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
----------	-----

Inline - Common

Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	4
Flow comp. 1	No
Multi-slice mode	Single shot
Echo spacing	9.8 ms
Bandwidth 1	650 Hz/Px
Bandwidth 2	650 Hz/Px
Bandwidth 3	650 Hz/Px
Bandwidth 4	650 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	176

Sequence - Special

Readout polarity	Positive
Readout trajectory	Bipolar
Gradient spoiling	Siemens
Gradient moment factor	1
Averaging	RMS

Sequence - Assistant

Mode	Off
------	-----

\\Research\\Investigators\\Belleau\\R01StressDyn\\slice_positioning 22 degree angle

TA: 0:21 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	1300 ms
TE	22.80 ms
Averages	1
Multi-band accel. factor	3
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1300 ms
TE	22.80 ms
MTC	Off
Magn. preparation	None
Flip angle	64 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	50

Resolution - iPAT

Reference scan mode	Single-shot
---------------------	-------------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	72
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	208 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	1300 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	3

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
--------------	----------

System - Adjustments

B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1300 ms
Multi-band accel. factor	3

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.53 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
Gradient mode	Performance

Sequence - Part 2

Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST1_ME_AP

TA: 8:24 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	360
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	360
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST_ME_PA_distortion

TA: 0:39 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : eptid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	10
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	10
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\Investigators\Belleau\R01StressDyn\SpinEchoFieldMap_AP

TA: 8.0 s PM: FIX Voxel size: 2.1×2.1×2.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8036 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8036 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Physio recording	Off
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\SpinEchoFieldMap_PA

TA: 8.0 s PM: FIX Voxel size: 2.1×2.1×2.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8036 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8036 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Physio recording	Off
Triggering scheme	Standard

\\Research\Investigators\Belleau\R01StressDyn\EMIST1_ME_AP

TA: 4:25 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	180
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	180
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\Investigators\Belleau\R01StressDyn\EMIST2_ME_AP

TA: 4:25 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	180
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	180
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST2_ME_AP

TA: 8:24 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	360
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	360
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\HMIST1_ME_AP

TA: 4:25 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	180
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	180
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\Investigators\Belleau\R01StressDyn\HMIST2_ME_AP

TA: 4:25 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	180
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	180
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST3_ME_AP

TA: 8:24 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	360
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	360
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\Investigators\Belleau\R01StressDyn\t2_tse_dark-fluid_tra

TA: 4:14 PM: REF Voxel size: 0.7×0.7×4.0 mmPAT: Off Rel. SNR: 1.00 : tir

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	96.9 %
Slice thickness	4.0 mm
TR	9000.0 ms
TE	83.0 ms
Averages	1
Concatenations	2
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	9000.0 ms
TE	83.0 ms
TD	0.0 ms
MTC	Off
Magn. preparation	Slice-sel. IR
T1	2500 ms
Flip angle	150 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Freeze suppressed tissue	On

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	96.9 %
Slice thickness	4.0 mm
Base resolution	320
Phase resolution	70 %

Resolution - Common

Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	96.9 %
Slice thickness	4.0 mm
TR	9000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**System - Miscellaneous**

Positioning mode	REF
Table position	H

System - Miscellaneous

Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	90.00 deg
R >> L	214 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	9000.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	Slice-sel. IR
TI	2500 ms
Fat suppr.	Fat sat.
Dark blood	Off
FoV read	220 mm
FoV phase	96.9 %
Phase resolution	70 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off

Inline - Common

Save original images	On
----------------------	----

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	8.28 ms
Bandwidth	289 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	13
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Performance
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	17

Sequence - Assistant

Mode	Off
Allowed delay	60 s

\\Research\Investigators\Belleau\R01StressDyn\pd+t2_tse_tra

TA: 2:11 PM: FIX Voxel size: 0.3×0.3×4.0 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3600.0 ms
TE 1	9.4 ms
TE 2	94 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	3600.0 ms
TE 1	9.4 ms
TE 2	94 ms
MTC	Off
Magn. preparation	None
Flip angle	160 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	320
Phase resolution	100 %
Phase partial Fourier	Off
Trajectory	Cartesian

Resolution - Common

Interpolation	On
---------------	----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	30
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	3600.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**System - Miscellaneous**

Positioning mode	FIX
Table position	H

System - Miscellaneous

Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3600.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	100.0 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	2
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	9.36 ms
Bandwidth	252 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	35
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Normal
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	5

Sequence - Assistant

Mode	Min flip angle
Min flip angle	130 deg
Allowed delay	60 s

\\Research\\Investigators\\Belleau\\R01StressDyn\\resolve_4scan_trace_tra_p2_192

TA: 1:51 PM: REF Voxel size: 1.1×1.1×4.0 mmPAT: 2 Rel. SNR: 1.00 : resolve

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	4040 ms
TE 1	55 ms
TE 2	93 ms
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HC1-7;NC1,2

Contrast - Common

TR	4040 ms
TE 1	55 ms
TE 2	93 ms
MTC	Off
Magn. preparation	None
Flip angle	180 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off
Readout partial Fourier	5/8
Readout segments	7
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	96
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	On
------------	----

Geometry - Common

Slice group	1
Slices	30
Dist. factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	4.0 mm
TR	4040 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	4040 ms
Concatenations	1

Diff - Neuro

Diffusion mode	4-Scan Trace
Diff. directions	4
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	Off
Trace weighted images	On
ADC maps	On
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	100

Diff - Body

Diffusion mode	4-Scan Trace
Diff. directions	4
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	Off
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	Off

Diff - Body

Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	100

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Contrasts	2
Optimization	Min. TE
Multi-slice mode	Interleaved
Echo spacing	0.32 ms
Bandwidth	766 Hz/Px

Sequence - Part 2

EPI factor	96
RF pulse type	Normal
Gradient mode	Fast
Reacquisition mode	On

Sequence - Assistant

Mode	Off
------	-----

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST4_ME_AP

TA: 8:24 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : eptid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	360
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	360
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\rfMRI_REST_ME_PA_distortion

TA: 0:39 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : eptid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	On
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
Averages	1
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	1330 ms
TE 1	12.60 ms
TE 2	29.16 ms
TE 3	45.72 ms
TE 4	62.28 ms
MTC	Off
Magn. preparation	None
Flip angle	67 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	10
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	72 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	1330 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares

System - Miscellaneous

Matrix Optimization	Performance
AutoAlign	Head > Brain
Coil Select Mode	On - AutoCoilSelect

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1330 ms
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	10
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	4
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off

Sequence - Part 1

Echo spacing	0.5 ms
Bandwidth	3230 Hz/Px

Sequence - Part 2

EPI factor	62
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	3820 us
Min. prep scans	0
Min. prep scans SB	0
Inter-TE delay	0 us
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\SpinEchoFieldMap_AP

TA: 8.0 s PM: FIX Voxel size: 2.1×2.1×2.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8036 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8036 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Physio recording	Off
Triggering scheme	Standard

\\Research\\Investigators\\Belleau\\R01StressDyn\\SpinEchoFieldMap_PA

TA: 8.0 s PM: FIX Voxel size: 2.1×2.1×2.5 mmPAT: Off Rel. SNR: 1.00 : epse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	8036 ms
TE	66.00 ms
MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
Base resolution	104
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	60
Dist. factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
FoV read	216 mm
FoV phase	100.0 %
Slice thickness	2.50 mm
TR	8036 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	T > C-22.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-22.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	Off

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	150 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249510 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	8036 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	Off
Threshold	4.00
Paradigm size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion correction	Off
Spatial filter	Off
Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Interleaved
Free echo spacing	On
Echo spacing	0.58 ms
Bandwidth	2290 Hz/Px

Sequence - Part 2

EPI factor	104
RF pulse type	Normal
Gradient mode	Performance
Excitation	Standard

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Physio recording	Off
Triggering scheme	Standard