

Table of contents

\\Research

Investigators - validated on FIT

Frederick

ADMS_CCF

localizer AAHScout T1_MEMPRAGE_64ch t2_tse_dark-fluid_tra pd+t2_tse_tra resolve_4scan_trace_tra_p2_192 slice_positioning 22 degree angle_CCF aAAC_preMAST_AP_CCF_1 aAAC_preMAST_AP_CCF_2 aAAC_preMAST_AP_CCF_3 aAAC_preMAST_PA_CCF_distortion SpinEchoFieldMap_AP_CCF slice_positioning 22 degree angle_CCF aAAC_postMAST_AP_CCF_1 aAAC_postMAST_AP_CCF_2 aAAC_postMAST_AP_CCF_3 aAAC_postMAST_PA_CCF_distortion SpinEchoFieldMap_AP_CCF
--

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\localizer

TA: 19 sec Coil Selection: Auto Voxel Size: 0.5×0.5×10.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	5
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	5
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	5
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	10.0 mm
TR	20.0 ms
TE	5.00 ms
Averages	1
Concatenations	7
AutoAlign	---

Contrast - Common

TR	20.0 ms
TE	5.00 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	40 deg

Contrast - Common

Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
SWI	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	10.0 mm
Base Resolution	256
Phase Resolution	50 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	None
Deep Resolve	Off
Phase Partial Fourier	Off
Asymmetric Echo	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
POCS	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	5
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	5
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	5
Distance Factor	20 %
Position	Isocenter

Geometry - Common

Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	280 mm
FOV Phase	100.0 %
Slice Thickness	10.0 mm
TR	20.0 ms
Multi-Slice Mode	Sequential
Series	Interleaved
Concatenations	7

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	20.0 ms
Segments	1
Concatenations	7

Physio - Cardiac

Tagging	None
Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	280 mm
FOV Phase	100.0 %
Phase Resolution	50 %

Physio - PACE

Resp. Control	Off
Concatenations	7

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
Measurements	1

Inline - Subtraction

StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	40 deg
Measurements	1
Contrasts	1
TE	5.00 ms
TR	20.0 ms
Save Original Images	On

Sequence - Part 1

Sequence Name	fl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Fast
Gradient Mode	Normal
Flow Compensation	None
Bandwidth	180 Hz/Px
Asymmetric Echo	Off
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\Research\Investigators - validated on FIT\Frederick\ADMS_CCF\AAHScout

TA: 17 sec Coil Selection: Auto Voxel Size: 1.6×1.6×1.6 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	All Segments
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
AutoAlign	Head

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip Angle	8 deg
Fat-Water Contrast	Standard
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Time to Center	7.5 s

Resolution - Common

FOV Read	260 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	1.6 mm
Base Resolution	160
Phase Resolution	100 %
Slice Resolution	69 %
Trajectory	Cartesian

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	24
Acceleration Factor 3D	1
Phase Partial Fourier	6/8
Slice Partial Fourier	6/8
Asymmetric Echo	Weak

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	3D
Normalize	Off
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	128
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	260 mm
FOV Phase	100.0 %
Slice Thickness	1.6 mm
TR	3.15 ms
Multi-Slice Mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	Isocenter
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm

Geometry - AutoAlign

H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	8 deg
Measurements	1
Contrasts	1
TE	1.37 ms
TR	3.15 ms
Save Original Images	On

Sequence - Part 1

Sequence Name	fl
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Normal
Bandwidth	540 Hz/Px
Asymmetric Echo	Weak

Sequence - Part 2

Introduction	On
RF Spoiling	On
Breast Application	Off
Phase Enc. Order	Automatic

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\T1_MEMPRAGE_64ch

TA: 6:02 min Coil Selection: Auto Voxel Size: 1.0x1.0x1.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
Base Resolution	256
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	2530.0 ms
TE 1	1.69 ms
TE 2	3.55 ms
TE 3	5.41 ms
TE 4	7.27 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	32
Acceleration Factor 3D	1
Phase Partial Fourier	Off
Slice Partial Fourier	Off
Asymmetric Echo	Off
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	On
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	2530.0 ms
Multi-Slice Mode	Single Shot
Series	Interleaved
Concatenations	1

Contrast - Common

TR	2530.0 ms
TE 1	1.69 ms
TE 2	3.55 ms
TE 3	5.41 ms
TE 4	7.27 ms
Magn. Preparation	Non-sel. IR
TI	1100 ms
Flip Angle	7 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	4
Reconstruction	Magnitude

Geometry - AutoAlign

Slab Group	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	R2.2 P1.7 F31.8
R	2.2 mm
P	1.7 mm
F	31.8 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	176 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2530.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	Non-sel. IR
TI	1100 ms
Dark Blood	Off
FOV Read	256 mm
FOV Phase	100.0 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Inline - MapIt

MapIt	None
Flip Angle	7 deg
Measurements	1
Contrasts	4
TE 1	1.69 ms
TE 2	3.55 ms
TE 3	5.41 ms
TE 4	7.27 ms
TR	2530.0 ms
Save Original Images	On

Sequence - Part 1

Sequence Name	tfl_me
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Fast
Flow Compensation 1	None
Flow Compensation 2	None
Flow Compensation 3	None
Flow Compensation 4	None
Reordering	Linear
Bandwidth 1	650 Hz/Px
Bandwidth 2	650 Hz/Px
Bandwidth 3	650 Hz/Px
Bandwidth 4	650 Hz/Px
Echo Spacing	9.84 ms
Asymmetric Echo	Off
Turbo Factor	176

Sequence - Part 2

Introduction	Off
RF Spoiling	On
Incr. Gradient Spoiling	Off

Sequence - Special

Readout polarity	Positive
Readout trajectory	Bipolar
Gradient spoiling	Siemens
Gradient moment factor	1.00
Averaging	RMS

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\t2_tse_dark-fluid_tra

TA: 4:14 min Coil Selection: Auto Voxel Size: 0.7×0.7×4.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
TR	9000.0 ms
TE	83.00 ms
Averages	1
Concatenations	2
AutoAlign	Head > Brain

Contrast - Common

TR	9000.0 ms
TE	83.00 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	Slice-sel. IR
TI	2500 ms
Freeze Suppr. Tissue	On
Flip Angle Mode	Constant
Flip Angle	150 deg
Fat-Water Contrast	Fat Saturation
Fat Saturation	Strong
Dark Blood	Off
Contrasts	1
Wrap-up Magn.	None
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
--------------	----------

Contrast - Dynamic

Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
Base Resolution	320
Phase Resolution	70 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Deep Resolve	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
TR	9000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice Group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Orientation	Transversal

Geometry - AutoAlign

Initial Rotation	90.00 deg
------------------	-----------

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	37 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	90.00 deg
R >> L	214 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
-----------------	------

Physio - Signal

TR	9000.0 ms
Concatenations	2

Physio - Cardiac

Fat-Water Contrast	Fat Saturation
Magn. Preparation	Slice-sel. IR
TI	2500 ms
Dark Blood	Off
FOV Read	220 mm
FOV Phase	96.9 %
Phase Resolution	70 %
Motion Correction	None

Physio - PACE

Resp. Control	Off
Concatenations	2

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Motion Correction	None
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tir
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	289 Hz/Px
Echo Spacing	8.28 ms
Free Echo Spacing	Off
Define	Turbo Factor
Turbo Factor	17
Echo Trains per Slice	13

Sequence - Part 2

Introduction	On
Phase Correction	Automatic

Sequence - Part 2

Compensate T2 Decay	Off
Fast Mode	Off
WARP	Off
Red. EC Sensitivity	Off
Acoustic noise reduction	Off
Reduce Motion Sens.	On
Motion Correction	None

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	60 s

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\pd+t2_tse_tra

TA: 2:11 min Coil Selection: Auto Voxel Size: 0.3x0.3x4.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Resolution - Common

FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
Base Resolution	320
Phase Resolution	100 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	30
Deep Resolve	Off
Phase Partial Fourier	Off

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	3600.0 ms
TE 1	9.40 ms
TE 2	94.00 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	3600.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	3600.0 ms
TE 1	9.40 ms
TE 2	94.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	Constant
Flip Angle	160 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	2
Wrap-up Magn.	None
Reconstruction	Magnitude

Geometry - AutoAlign

Slice Group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Orientation	Transversal
Initial Rotation	90.00 deg

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	37 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3600.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	220 mm
FOV Phase	100.0 %
Phase Resolution	100 %
Motion Correction	None

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Motion Correction	None
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tse
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Normal
Flow Compensation	None
Bandwidth	252 Hz/Px
Echo Spacing	9.36 ms
Free Echo Spacing	Off
Define	Turbo Factor
Turbo Factor	5
Echo Trains per Slice	35

Sequence - Part 2

Introduction	On
Phase Correction	Automatic
Compensate T2 Decay	Off
Fast Mode	Off
WARP	Off
Red. EC Sensitivity	Off
Acoustic noise reduction	Off

Sequence - Part 2

Reduce Motion Sens.	On
Motion Correction	None

Sequence - Assistant

SAR Assistant	Flip Angle
Min Flip Angle	130 deg
Allowed Delay	60 s

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\resolve_4scan_trace_tra_p2_192

TA: 1:55 min Coil Selection: Auto Voxel Size: 1.1×1.1×4.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Resolution - Common

FOV Phase	100.0 %
Slice Thickness	4.0 mm
Base Resolution	192
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Accel. Mode	GRAPPA
Reference Scans	EPI/Separate
Acceleration Factor PE	2
Reference Lines PE	96
Phase Partial Fourier	Off
Readout Partial Fourier	5/8
Readout Segments	7

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	4190.0 ms
TE 1	54 ms
TE 2	92 ms
Concatenations	1
AutoAlign	Head > Brain

Resolution - Filter

Raw Filter	On
Distortion Correction	Off
Normalize	Prescan
Noise Masking	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	4190.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	4190.0 ms
TE 1	54 ms
TE 2	92 ms
MTC	Off
Magn. Preparation	None
Flip Angle	180 deg
Fat-Water Contrast	Fat Saturation
Fat Saturation	Strong
Contrasts	2
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1

Resolution - Common

FOV Read	220 mm
----------	--------

Geometry - AutoAlign

Slice Group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	37 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
---------	----------

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4190.0 ms
Concatenations	1

Diff

Diffusion Mode	4-Scan Trace
Diff. Directions	4

Diff

Diffusion Scheme	Monopolar
Diff. Weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
Averages 1	1
Averages 2	1
Invert Gray Scale	Off
Diff. Weighted Images	Off
Trace Weighted Images	On
Tensor	Off
FA Maps	Off
ADC Maps	On
Exponential ADC Maps	Off
b-value >=	0 s/mm ²
ADC Noise Threshold	100
Noise Masking	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	resolve
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	766 Hz/Px
Echo Spacing	0.32 ms
Optimization	Min. TE
EPI Factor	96

Sequence - Part 2

Introduction	On
Reacquisition Mode	On

Sequence - Assistant

SAR Assistant	Off
Optimization	Min. TE

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\slice_positioning 22 degree angle_CC F
TA: 10 sec Coil Selection: Auto Voxel Size: 2.0x2.0x2.0 mm ³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
----------	--------

Resolution - Common

FOV Phase	100.0 %
Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms

Physio - Signal

Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\Frederick\ADMS_CCF\AAC_preMAST_AP_CCF_1

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\Frederick\ADMS_CCF\AAC_preMAST_AP_CCF_2

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\Frederick\ADMS_CCF\AAC_preMAST_AP_CCF_3

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\Investigators - validated on FIT\Frederick\ADMS_CCF\AAC_preMAST_PA_CCF_distortion

TA: 18 sec Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	10
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	10
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\SpinEchoFieldMap_AP_CCF

TA: 8 sec Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	8036.0 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---

Contrast - Common

TR	8036.0 ms
TE	66.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	8036.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	8036.0 ms
Log Signals	Off
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	Off
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	On
EPI Factor	104

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Force GPA balance	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\slice_positioning 22 degree angle_CC F
TA: 10 sec Coil Selection: Auto Voxel Size: 2.0x2.0x2.0 mm ³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
----------	--------

Resolution - Common

FOV Phase	100.0 %
Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms

Physio - Signal

Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\aAAC_postMAST_AP_CCF_1

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0x2.0x2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\aAAC_postMAST_AP_CCF_2

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\aAAC_postMAST_AP_CCF_3

TA: 7:34 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	On
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	555
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	555
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\aAAC_postMAST_PA_CCF_distortion

TA: 18 sec Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc.: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
TE	37.00 ms
Averages	1
Multi-band accel. factor	8
AutoAlign	---

Contrast - Common

TR	800.0 ms
TE	37.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	10
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	8

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	Ext./Trigger
Average Cycle	No Signal ms
Acquisition Window	800 ms
Trigger Pulse	1
Trigger Delay	0 ms
TR	800.0 ms
Log Signals	Off
Multi-band accel. factor	8
Phases	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	10
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	Off
EPI Factor	104

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	On
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\ADMS_CCF\\SpinEchoFieldMap_AP_CCF

TA: 8 sec Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	8036.0 ms
TE	66.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---

Contrast - Common

TR	8036.0 ms
TE	66.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
Base Resolution	104
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	72
Distance Factor	0 %
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	208 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	8036.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	T > C-22.0
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	T > C
T > C	-22.00
> S	0.00
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	T > C-22.0
Rotation	0.00 deg
A >> P	208 mm
R >> L	208 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	8036.0 ms
Log Signals	Off
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	Off
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2290 Hz/Px
Echo Spacing	0.58 ms
Free Echo Spacing	On
EPI Factor	104

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Force GPA balance	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----