

Table of contents

\\Research

Investigators - validated on FIT

Frederick

cmrr base

cmrr_mbep2d_bold	*
cmrr_mbep2d_diff	*
cmrr_mbep2d_se	*

\\Research\\Investigators - validated on FIT\\Frederick\\cmrr base\\cmrr_mbep2d_bold *

TA: 8 sec Coil Selection: Auto Voxel Size: 3.9×3.9×5.0 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	250.0 ms
TE	60.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---

Contrast - Common

TR	250.0 ms
TE	60.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	20
Delay in TR	0.00 ms

Resolution - Common

FOV Read	500 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	5.0 mm
Base Resolution	128
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	2D
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	250.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	500 mm
R >> L	500 mm
F >> H	5 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	250.0 ms
Log Signals	Off
Multi-band accel. factor	1

BOLD

GLM Statistics	On
Ignore Meas. at Start	0
Ignore After Transition	0

BOLD

Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion Correction	On
Spatial Filter	Off
Measurements	20
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1502 Hz/Px
Echo Spacing	0.71 ms
Free Echo Spacing	Off
EPI Factor	128

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	2560 us
Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
FFT scale factor	1.00

Sequence - Special

Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\cmrr base\\cmrr_mbep2d_diff *

TA: 1 sec Coil Selection: Auto Voxel Size: 3.9×3.9×5.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	500.0 ms
TE	200.00 ms
Multi-band accel. factor	1
AutoAlign	---

Contrast - Common

TR	500.0 ms
TE	200.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	500 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	5.0 mm
Base Resolution	128
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	500.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	500 mm
R >> L	500 mm
F >> H	5 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	500.0 ms
Multi-band accel. factor	1

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	1

Diff

Diffusion Mode	MDDW
Diff. Directions	6
Diffusion Scheme	Bipolar
Diff. Weightings	1
b-value	0 s/mm ²
Averages	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	Off
ADC Maps	Off
Exponential ADC Maps	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	1502 Hz/Px
Echo Spacing	0.71 ms
Free Echo Spacing	Off
EPI Factor	128

Sequence - Part 2

Introduction	Off
RF Spoiling	Off
Phase Correction	Internal

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Force GPA balance	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators - validated on FIT\\Frederick\\cmrr base\\cmrr_mbep2d_se *

TA: 7 sec Coil Selection: Auto Voxel Size: 3.9×3.9×5.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	177.0 ms
TE	95.60 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---

Contrast - Common

TR	177.0 ms
TE	95.60 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	20
Delay in TR	0.00 ms

Resolution - Common

FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
Base Resolution	128
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	2D
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	50 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	500 mm
FOV Phase	100.0 %
Slice Thickness	5.0 mm
TR	177.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	500 mm
R >> L	500 mm
F >> H	5 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.253118 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	177.0 ms
Log Signals	Off
Multi-band accel. factor	1

BOLD

GLM Statistics	On
Ignore Meas. at Start	0
Ignore After Transition	0

BOLD

Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion Correction	On
Spatial Filter	Off
Measurements	20
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	1502 Hz/Px
Echo Spacing	0.71 ms
Free Echo Spacing	Off
EPI Factor	128

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Force GPA balance	Off
Disable freq. update	Off
Suppress 16-bit DICOM	On
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off

Sequence - Special

Triggering scheme	Standard
-------------------	----------

Sequence - Assistant

SAR Assistant	Off
---------------	-----