

Table of contents

\\Research

<table border="1"><tr><td>Investigators</td></tr></table>	Investigators
Investigators	

<table border="1"><tr><td>Dickstein</td></tr></table>	Dickstein
Dickstein	

<table border="1"><tr><td>Irritability_PRR 1st</td></tr></table>	Irritability_PRR 1st
Irritability_PRR 1st	

<table border="1"><tr><td><table border="1"><tr><td>localizer</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>t1_mprage</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_REST</td></tr><tr><td>cmrr_DSI_1.5_HARDI</td></tr><tr><td>pd+t2_tse_tra</td></tr><tr><td>t2_tse_dark-fluid_tra</td></tr><tr><td>resolve_4scan_trace_tra_p2_192</td></tr><tr><td>ep2D_mb_PRR</td></tr></table></td></tr></table>	<table border="1"><tr><td>localizer</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>t1_mprage</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_REST</td></tr><tr><td>cmrr_DSI_1.5_HARDI</td></tr><tr><td>pd+t2_tse_tra</td></tr><tr><td>t2_tse_dark-fluid_tra</td></tr><tr><td>resolve_4scan_trace_tra_p2_192</td></tr><tr><td>ep2D_mb_PRR</td></tr></table>	localizer	ep2D_mb_PRR	ep2D_mb_PRR	ep2D_mb_PRR	ep2D_mb_PRR	t1_mprage	ep2D_mb_POSNER	ep2D_mb_POSNER	ep2D_mb_REST	cmrr_DSI_1.5_HARDI	pd+t2_tse_tra	t2_tse_dark-fluid_tra	resolve_4scan_trace_tra_p2_192	ep2D_mb_PRR
<table border="1"><tr><td>localizer</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>ep2D_mb_PRR</td></tr><tr><td>t1_mprage</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_POSNER</td></tr><tr><td>ep2D_mb_REST</td></tr><tr><td>cmrr_DSI_1.5_HARDI</td></tr><tr><td>pd+t2_tse_tra</td></tr><tr><td>t2_tse_dark-fluid_tra</td></tr><tr><td>resolve_4scan_trace_tra_p2_192</td></tr><tr><td>ep2D_mb_PRR</td></tr></table>	localizer	ep2D_mb_PRR	ep2D_mb_PRR	ep2D_mb_PRR	ep2D_mb_PRR	t1_mprage	ep2D_mb_POSNER	ep2D_mb_POSNER	ep2D_mb_REST	cmrr_DSI_1.5_HARDI	pd+t2_tse_tra	t2_tse_dark-fluid_tra	resolve_4scan_trace_tra_p2_192	ep2D_mb_PRR	
localizer															
ep2D_mb_PRR															
ep2D_mb_PRR															
ep2D_mb_PRR															
ep2D_mb_PRR															
t1_mprage															
ep2D_mb_POSNER															
ep2D_mb_POSNER															
ep2D_mb_REST															
cmrr_DSI_1.5_HARDI															
pd+t2_tse_tra															
t2_tse_dark-fluid_tra															
resolve_4scan_trace_tra_p2_192															
ep2D_mb_PRR															

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\localizer

TA: 30 sec Coil Selection: Auto Voxel Size: 0.5×0.5×7.0 mm³ Acc.: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	Off
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	All Segments
Inline Movie	Off

Routine

Slice Group	1
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	8.6 ms
TE	3.69 ms
Averages	2
Concatenations	7
AutoAlign	---

Contrast - Common

TR	8.6 ms
TE	3.69 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	20 deg

Contrast - Common

Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
SWI	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
Base Resolution	256
Phase Resolution	91 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	None
Deep Resolve	Off
Phase Partial Fourier	Off
Asymmetric Echo	Allowed

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	3
Distance Factor	20 %
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal

Geometry - Common

Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	250 mm
FOV Phase	100.0 %
Slice Thickness	7.0 mm
TR	8.6 ms
Multi-Slice Mode	Sequential
Series	Interleaved
Concatenations	7

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	2
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slice Group	3
Position	L0.0 A10.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H0.0
L	0.0 mm
A	10.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
---------------------	----------

System - Adjustments

B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	8.6 ms
Segments	1
Concatenations	7

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	250 mm
FOV Phase	100.0 %
Phase Resolution	91 %

Physio - PACE

Resp. Control	Off
Concatenations	7

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	fl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Normal
Gradient Mode	Normal
Flow Compensation	None
Bandwidth	320 Hz/Px
Asymmetric Echo	Allowed
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_PRR

TA: 6:22 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	2
AutoAlign	---

Contrast - Common

TR	2300.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	156
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2300.0 ms
Log Signals	Off
Multi-band accel. factor	2

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	156
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1796 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	3800 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online

Sequence - Special

FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_PRR

TA: 6:22 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	2
AutoAlign	---

Contrast - Common

TR	2300.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	156
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2300.0 ms
Log Signals	Off
Multi-band accel. factor	2

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	156
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1796 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	3800 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online

Sequence - Special

FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_PRR

TA: 6:22 min Coil Selection: Auto Voxel Size: 2.0x2.0x2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	2
AutoAlign	---

Contrast - Common

TR	2300.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	156
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2300.0 ms
Log Signals	Off
Multi-band accel. factor	2

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	156
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1796 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	3800 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online

Sequence - Special

FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_PRR

TA: 6:22 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	2
AutoAlign	---

Contrast - Common

TR	2300.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	156
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2300.0 ms
Log Signals	Off
Multi-band accel. factor	2

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	156
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1796 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	3800 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online

Sequence - Special

FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\Investigators\Dickstein\Irritability_PRR 1st\t1_mprage

TA: 4:01 min Coil Selection: Auto Voxel Size: 1.0x1.0x1.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	R0.7 A16.9 H35.1 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	160
Phase Oversampling	0 %
Slice Oversampling	30.0 %
FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	1900.0 ms
TE	3.01 ms
Averages	1
Concatenations	1
AutoAlign	---

Contrast - Common

TR	1900.0 ms
TE	3.01 ms
Magn. Preparation	Non-sel. IR
TI	900 ms
Flip Angle	9 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
Base Resolution	256
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	30
Acceleration Factor 3D	1
Phase Partial Fourier	7/8
Slice Partial Fourier	Off
Asymmetric Echo	Off
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	R0.7 A16.9 H35.1 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	160
Phase Oversampling	0 %
Slice Oversampling	30.0 %
FOV Read	256 mm
FOV Phase	100.0 %
Slice Thickness	1.00 mm
TR	1900.0 ms
Multi-Slice Mode	Single Shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	R0.7 A16.9 H35.1 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---

Geometry - AutoAlign

Initial Position	R0.7 A16.9 H35.1
R	0.7 mm
A	16.9 mm
H	35.1 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Tim Planning Suite**

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R0.7 A16.9 H35.1 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	160 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	1900.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	Non-sel. IR
TI	900 ms
Dark Blood	Off
FOV Read	256 mm
FOV Phase	100.0 %
Phase Resolution	100 %

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tfl
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Fast*
Flow Compensation	None
Reordering	Linear
Bandwidth	230 Hz/Px
Echo Spacing	8.10 ms
Asymmetric Echo	Off
Turbo Factor	208

Sequence - Part 2

Introduction	On
RF Spoiling	On
Incr. Gradient Spoiling	On

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_POSNER

TA: 7:06 min Coil Selection: Auto Voxel Size: 2.0x2.0x2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	1000.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	4
AutoAlign	---

Contrast - Common

TR	1000.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	406
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	1000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	1000.0 ms
Log Signals	Off
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	406
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1860 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	4500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off

Sequence - Special

Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_POSNER

TA: 7:06 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	1000.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	4
AutoAlign	---

Contrast - Common

TR	1000.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	406
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	1000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	1000.0 ms
Log Signals	Off
Multi-band accel. factor	4

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	406
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1860 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	4500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off

Sequence - Special

Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_REST

TA: 9:00 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	66
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2000.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	3
AutoAlign	---

Contrast - Common

TR	2000.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	256
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	66
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	3

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	132 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2000.0 ms
Log Signals	Off
Multi-band accel. factor	3

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	256
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1336 Hz/Px
Echo Spacing	0.93 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	4500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off

Sequence - Special

Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\Investigators\Dickstein\Irritability_PRR 1st\cmrr_DSI_1.5_HARDI

TA: 9:02 min Coil Selection: Auto Voxel Size: 1.5×1.5×1.5 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	84
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	228 mm
FOV Phase	100.0 %
Slice Thickness	1.5 mm
TR	3205.0 ms
TE	94.00 ms
Multi-band accel. factor	4
AutoAlign	---

Contrast - Common

TR	3205.0 ms
TE	94.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FOV Read	228 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	1.5 mm
Base Resolution	152
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	84
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	228 mm
FOV Phase	100.0 %
Slice Thickness	1.5 mm
TR	3205.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	4

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Navigator

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	228 mm
R >> L	228 mm
F >> H	126 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3205.0 ms
Multi-band accel. factor	4

Physio - PACE

Resp. Control	Off
---------------	-----

Physio - PACE

Multi-band accel. factor	4
--------------------------	---

Diff

Diffusion Mode	Free
Diff. Directions	75
Diffusion Scheme	Monopolar
Diff. Weightings	3
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 3	2500 s/mm ²
Averages 1	1
Averages 2	1
Averages 3	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	On
Tensor	On
FA Maps	On
ADC Maps	On
Exponential ADC Maps	Off
ADC Noise Threshold	40
Calculated Image	Off

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
Gradient Mode	Performance*
Bandwidth	1566 Hz/Px
Echo Spacing	0.74 ms
Free Echo Spacing	Off
EPI Factor	152

Sequence - Part 2

Introduction	Off
RF Spoiling	Off
Phase Correction	Internal

Sequence - Special

Excite pulse duration	4480 us
Refocus pulse duration	8320 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
Time-shifted MB RF	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Force GPA balance	Off

Sequence - Special

Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\pd+t2_tse_tra

TA: 2:11 min Coil Selection: Auto Voxel Size: 0.3×0.3×4.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Resolution - Common

FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
Base Resolution	320
Phase Resolution	100 %
Interpolation	On

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	30
Deep Resolve	Off
Phase Partial Fourier	Off

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	3600.0 ms
TE 1	9.40 ms
TE 2	94.00 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	3600.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	3600.0 ms
TE 1	9.40 ms
TE 2	94.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	Constant
Flip Angle	160 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	2
Wrap-up Magn.	None
Reconstruction	Magnitude

Geometry - AutoAlign

Slice Group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Orientation	Transversal
Initial Rotation	90.00 deg

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	37 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	3600.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FOV Read	220 mm
FOV Phase	100.0 %
Phase Resolution	100 %
Motion Correction	None

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Motion Correction	None
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tse
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Normal
Flow Compensation	None
Bandwidth	252 Hz/Px
Echo Spacing	9.36 ms
Free Echo Spacing	Off
Define	Turbo Factor
Turbo Factor	5
Echo Trains per Slice	35

Sequence - Part 2

Introduction	On
Phase Correction	Automatic
Compensate T2 Decay	Off
Fast Mode	Off
WARP	Off
Red. EC Sensitivity	Off
Acoustic noise reduction	Off

Sequence - Part 2

Reduce Motion Sens.	On
Motion Correction	None

Sequence - Assistant

SAR Assistant	Flip Angle
Min Flip Angle	130 deg
Allowed Delay	60 s

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\t2_tse_dark-fluid_tra

TA: 4:14 min Coil Selection: Auto Voxel Size: 0.7×0.7×4.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	On
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
TR	9000.0 ms
TE	83.00 ms
Averages	1
Concatenations	2
AutoAlign	Head > Brain

Contrast - Common

TR	9000.0 ms
TE	83.00 ms
TD	0.00 ms
MTC	Off
Magn. Preparation	Slice-sel. IR
TI	2500 ms
Freeze Suppr. Tissue	On
Flip Angle Mode	Constant
Flip Angle	150 deg
Fat-Water Contrast	Fat Saturation
Fat Saturation	Strong
Dark Blood	Off
Contrasts	1
Wrap-up Magn.	None
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
--------------	----------

Contrast - Dynamic

Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
Base Resolution	320
Phase Resolution	70 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	None
Deep Resolve	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	96.9 %
Slice Thickness	4.0 mm
TR	9000.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice Group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Orientation	Transversal

Geometry - AutoAlign

Initial Rotation	90.00 deg
------------------	-----------

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	37 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	90.00 deg
R >> L	214 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
-----------------	------

Physio - Signal

TR	9000.0 ms
Concatenations	2

Physio - Cardiac

Fat-Water Contrast	Fat Saturation
Magn. Preparation	Slice-sel. IR
TI	2500 ms
Dark Blood	Off
FOV Read	220 mm
FOV Phase	96.9 %
Phase Resolution	70 %
Motion Correction	None

Physio - PACE

Resp. Control	Off
Concatenations	2

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Motion Correction	None
Save Original Images	On

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tir
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	289 Hz/Px
Echo Spacing	8.28 ms
Free Echo Spacing	Off
Define	Turbo Factor
Turbo Factor	17
Echo Trains per Slice	13

Sequence - Part 2

Introduction	On
Phase Correction	Automatic

Sequence - Part 2

Compensate T2 Decay	Off
Fast Mode	Off
WARP	Off
Red. EC Sensitivity	Off
Acoustic noise reduction	Off
Reduce Motion Sens.	On
Motion Correction	None

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	60 s

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\resolve_4scan_trace_tra_p2_192

TA: 1:55 min Coil Selection: Auto Voxel Size: 1.1×1.1×4.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Resolution - Common

FOV Phase	100.0 %
Slice Thickness	4.0 mm
Base Resolution	192
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Accel. Mode	GRAPPA
Reference Scans	EPI/Separate
Acceleration Factor PE	2
Reference Lines PE	96
Phase Partial Fourier	Off
Readout Partial Fourier	5/8
Readout Segments	7

Routine

Slice Group	1
Slices	30
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	4190.0 ms
TE 1	54 ms
TE 2	92 ms
Concatenations	1
AutoAlign	Head > Brain

Resolution - Filter

Raw Filter	On
Distortion Correction	Off
Normalize	Prescan
Noise Masking	Off

Geometry - Common

Slice Group	1
Slices	30
Distance Factor	25 %
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	220 mm
FOV Phase	100.0 %
Slice Thickness	4.0 mm
TR	4190.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	4190.0 ms
TE 1	54 ms
TE 2	92 ms
MTC	Off
Magn. Preparation	None
Flip Angle	180 deg
Fat-Water Contrast	Fat Saturation
Fat Saturation	Strong
Contrasts	2
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1

Resolution - Common

FOV Read	220 mm
----------	--------

Geometry - AutoAlign

Slice Group	1
Position	Isocenter
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
F	37.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	ACS All but spine
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Adaptive Combine
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
---------	----------

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4190.0 ms
Concatenations	1

Diff

Diffusion Mode	4-Scan Trace
Diff. Directions	4

Diff

Diffusion Scheme	Monopolar
Diff. Weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
Averages 1	1
Averages 2	1
Invert Gray Scale	Off
Diff. Weighted Images	Off
Trace Weighted Images	On
Tensor	Off
FA Maps	Off
ADC Maps	On
Exponential ADC Maps	Off
b-value >=	0 s/mm ²
ADC Noise Threshold	100
Noise Masking	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	resolve
Dimension	2D
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	766 Hz/Px
Echo Spacing	0.32 ms
Optimization	Min. TE
EPI Factor	96

Sequence - Part 2

Introduction	On
Reacquisition Mode	On

Sequence - Assistant

SAR Assistant	Off
Optimization	Min. TE

\\Research\\Investigators\\Dickstein\\Irritability_PRR 1st\\ep2D_mb_PRR

TA: 6:22 min Coil Selection: Auto Voxel Size: 2.0×2.0×2.0 mm³ Acc:: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Disable auto transfer to PACS	Off
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	2
AutoAlign	---

Contrast - Common

TR	2300.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	72 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	156
Delay in TR	0.00 ms

Resolution - Common

FOV Read	192 mm
FOV Phase	100.0 %

Resolution - Common

Slice Thickness	2.0 mm
Base Resolution	96
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration Mode	GRAPPA
Reference scan mode	Single-shot
Acceleration Factor PE	2
Reference Lines PE	24
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Static Field Correction	Off
Normalize	Prescan

Geometry - Common

Slice Group	1
Slices	64
Distance Factor	0 %
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FOV Read	192 mm
FOV Phase	100.0 %
Slice Thickness	2.0 mm
TR	2300.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	2

Geometry - AutoAlign

Slice Group	1
Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	R3.0 A9.6 F7.8
R	3.0 mm
A	9.6 mm
F	7.8 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	8 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Auto Coil Select
Radial Sorting	Off
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off
Coil Focus	Flat

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	On
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	R3.0 A9.6 F7.8 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	192 mm
R >> L	192 mm
F >> H	128 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249566 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	2300.0 ms
Log Signals	Off
Multi-band accel. factor	2

BOLD

GLM Statistics	Off
Ignore Meas. at Start	6
Ignore After Transition	0
Model Transition States	Off
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	12
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Active
Meas[8]	Active
Meas[9]	Active
Meas[10]	Active
Meas[11]	Active
Meas[12]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	156
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	1796 Hz/Px
Echo Spacing	0.64 ms
Free Echo Spacing	Off
EPI Factor	96

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	3800 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online

Sequence - Special

FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----