

Table of contents

\\Research

Investigators

Webb

R01_Mindfulness

localizer all slices ac_pc_64ch AAHScout_64ch sliceposition_mb6_gr2_ACPC T1_MEMPRAGE_64ch_gr2 T2_SPACE_gr2 Connectome_resting_mb6_gr2_ACPCC Connectome_resting_mb6_gr2_ACPCC Connectome_resting_mb6_gr2_ACPCC Connectome_resting_mb6_gr2_ACPCC Connectome_BCT_mb6_gr2 Connectome_gre_field_mapping pd+t2_tse_tra t2_tse_dark-fluid_tra resolve_4scan_trace_tra_p2_192
--

\\Research\\Investigators\\Webb\\R01_Mindfulness\\localizer all slices ac_pc_64ch

TA: 0:19 PM: REF Voxel size: 0.5×0.5×10.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
TR	20.0 ms
TE	5.00 ms
Averages	1
Concatenations	7
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	20.0 ms
TE	5.00 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	40 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
Base resolution	256
Phase resolution	50 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
TR	20.0 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	7

Geometry - AutoAlign

Slice group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	Isocenter

Geometry - AutoAlign

Orientation	Transversal
Phase enc. dir.	A >> P
Slice group	3
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Slice-sel.

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	20.0 ms
Concatenations	7
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	50 %

Physio - PACE

Resp. control	Off
Concatenations	7

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	180 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Fast
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Assistant

Mode	Off
Allowed delay	0 s

\\Research\\Investigators\\Webb\\R01_Mindfulness\\AAHScout_64ch

TA: 0:17 PM: REF Voxel size: 1.6×1.6×1.6 mmPAT: 2 Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	3.15 ms
TE	1.37 ms
Flip angle	8 deg

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
Trajectory	Cartesian

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1

Resolution - iPAT

Reference scan mode	Integrated
---------------------	------------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Initial Position	Isocenter
L	0.0 mm
P	0.0 mm
H	0.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

System - Miscellaneous

Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

Sequence - Part 2

RF spoiling	On
-------------	----

Sequence - Assistant

Mode	Off
------	-----

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Flip angle	8 deg
Measurements	1
Time to center	7.5 s

Inline - Inline

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	540 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.

\\Research\\Investigators\\Webb\\R01_Mindfulness\\slice\\position_mb6_gr2_ACPC

TA: 0:22 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	720 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	66 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms

Resolution - Common

FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	100 %
Phase partial Fourier	7/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32

Resolution - iPAT

Reference scan mode	Single-shot
---------------------	-------------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
--------------	----------

System - Adjustments

B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	217 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	720 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

BOLD

Measurements	1
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2528 Hz/Px

Sequence - Part 2

EPI factor	88
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	10500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	Off
Triggering scheme	Standard

\\Research\\Investigators\\Webb\\R01_Mindfulness\\T1_MEMPRAGE_64ch_gr2

TA: 4:32 PM: REF Voxel size: 1.0×1.0×1.3 mmPAT: 2 Rel. SNR: 1.00 : tfl_me

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
TR	2530.0 ms
TE 1	3.3 ms
TE 2	6.93 ms
TE 3	8.79 ms
TE 4	10.65 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	2530.0 ms
TE 1	3.3 ms
TE 2	6.93 ms
TE 3	8.79 ms
TE 4	10.65 ms
Magn. preparation	Non-sel. IR
TI	1100 ms
Flip angle	7.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm

Resolution - Common

Base resolution	256
Phase resolution	75 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
TR	2530.0 ms
Multi-slice mode	Single shot
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	R2.2 P1.7 F31.8
R	2.2 mm
P	1.7 mm
F	31.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Navigator**System - Miscellaneous**

Positioning mode	REF
------------------	-----

System - Miscellaneous

Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	171 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	2530.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI	1100 ms
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1

Inline - Common

StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	4
Flow comp. 1	No
Multi-slice mode	Single shot
Echo spacing	14.1 ms
Bandwidth 1	200 Hz/Px
Bandwidth 2	650 Hz/Px
Bandwidth 3	650 Hz/Px
Bandwidth 4	650 Hz/Px

Sequence - Part 2

RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	128

Sequence - Special

Readout polarity	Positive
Readout trajectory	Bipolar
Gradient spoiling	Integral
Gradient moment factor	3.0
Averaging	RMS

Sequence - Assistant

Mode	Off
------	-----

\\Research\\Investigators\\Webb\\R01_Mindfulness\\T2_SPACE_gr2

TA: 1:32 PM: FIX Voxel size: 1.0×1.0×1.3 mmPAT: 2 Rel. SNR: 1.00 : spc

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
TR	2800 ms
TE	307 ms
Averages	1.0
Concatenations	1
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	2800 ms
TE	307 ms
MTC	Off
Magn. preparation	None
Fat suppr.	None
Blood suppr.	Off
Restore magn.	Off

Contrast - Dynamic

Averages	1.0
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
Base resolution	256
Phase resolution	75 %
Slice resolution	100 %
Phase partial Fourier	Allowed
Slice partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice oversampling	0.0 %
Slices per slab	128
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
TR	2800 ms
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R2.2 P1.7 F31.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	R2.2 P1.7 F31.8
R	2.2 mm
P	1.7 mm
F	31.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**System - Miscellaneous**

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P

System - Miscellaneous

Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
Trigger delay	0 ms
TR	2800 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	256 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
---------	-----

Inline - MIP

MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	Off
Dimension	3D
Elliptical scanning	On
Reordering	Linear
Flow comp.	No
Echo spacing	3.04 ms
Adiabatic-mode	Off
Bandwidth	849 Hz/Px

Sequence - Part 2

Echo train duration	733 ms
RF pulse type	Normal
Gradient mode	Performance
Excitation	Non-sel.
Flip angle mode	T2 var
Turbo factor	282

Sequence - Assistant

Allowed delay	0 s
---------------	-----

\\Research\Investigators\Webb\R01_Mindfulness\Connectome_resting_mb6_gr2_ACPConnectome_resting_mb6_gr2_ACP

TA: 6:21 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	32
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
Filter	Prescan Normalize
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Contrast - Common

TR	720 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	66 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	500
Delay in TR	0 ms

Resolution - Common

FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	100 %
Phase partial Fourier	7/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	217 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	720 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off

BOLD

Spatial filter	Off
Measurements	500
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2528 Hz/Px

Sequence - Part 2

EPI factor	88
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	10500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	Off
Triggering scheme	Standard

\\Research\Investigators\Webb\R01_Mindfulness\Connectome_resting_mb6_gr2_ACPConnectome_resting_mb6_gr2_ACP

TA: 6:21 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	32
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
Filter	Prescan Normalize
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Contrast - Common

TR	720 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	66 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	500
Delay in TR	0 ms

Resolution - Common

FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	100 %
Phase partial Fourier	7/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	217 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	720 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off

BOLD

Spatial filter	Off
Measurements	500
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2528 Hz/Px

Sequence - Part 2

EPI factor	88
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	10500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	Off
Triggering scheme	Standard

\\Research\Investigators\Webb\R01_Mindfulness\Connectome_resting_mb6_gr2_ACPConnectome_resting_mb6_gr2_ACP

TA: 6:21 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Resolution - iPAT

Ref. lines PE	32
Reference scan mode	Single-shot

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
Filter	Prescan Normalize
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Contrast - Common

TR	720 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	66 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	500
Delay in TR	0 ms

Resolution - Common

FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	100 %
Phase partial Fourier	7/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	217 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	720 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off

BOLD

Spatial filter	Off
Measurements	500
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2528 Hz/Px

Sequence - Part 2

EPI factor	88
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	10500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	Off
Triggering scheme	Standard

\\Research\\Investigators\\Webb\\R01_Mindfulness\\Connectome_BCT_mb6_gr2

TA: 20:21 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: 2 Rel. SNR: 1.00 : efpid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	On
Start measurements	Single measurement

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
Filter	Prescan Normalize
Coil elements	HC1-7

Contrast - Common

TR	720 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	66 deg
Fat suppr.	Fat sat.

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1667
Delay in TR	0 ms

Resolution - Common

FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
Base resolution	86
Phase resolution	100 %
Phase partial Fourier	7/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	32

Resolution - iPAT

Reference scan mode	Single-shot
---------------------	-------------

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
FoV read	212 mm
FoV phase	102.3 %
Slice thickness	2.50 mm
TR	720 ms
Multi-slice mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Geometry - Saturation

Fat suppr.	Fat sat.
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
--------------	----------

System - Adjustments

B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	217 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	720 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

BOLD

Measurements	1667
Delay in TR	0 ms

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	0.5 ms
Bandwidth	2528 Hz/Px

Sequence - Part 2

EPI factor	88
Gradient mode	Performance
Excitation	Standard
RF spoiling	Off

Sequence - Special

Excite pulse duration	10500 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable B1 control loop	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.0 deg
Fat sat. offset	0.0 Hz
Sinc exc. pulse BWTP	5.2
Physio recording	DICOM
Triggering scheme	Standard

\\Research\\Investigators\\Webb\\R01_Mindfulness\\Connectome_gre_field_mapping

TA: 2:44 PM: FIX Voxel size: 3.3×3.3×2.5 mmRel. SNR: 1.00 : fm_r

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	1250.0 ms
TE 1	10.00 ms
TE 2	12.46 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HC1-7

Contrast - Common

TR	1250.0 ms
TE 1	10.00 ms
TE 2	12.46 ms
MTC	Off
Flip angle	60 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off

Resolution - Filter Image

Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	66
Dist. factor	0 %
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	R >> L
FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	1250.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 A6.6 F11.0
L	0.0 mm
A	6.6 mm
F	11.0 mm
Initial Rotation	90.00 deg
Initial Orientation	T > C
T > C	-25.0
> S	0.0

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 A6.6 F11.0 mm
Orientation	T > C-25.0
Rotation	90.00 deg
R >> L	212 mm
A >> P	212 mm
F >> H	165 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Flow comp.	Yes
Multi-slice mode	Interleaved
Bandwidth	290 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\Research\Investigators\Webb\R01_Mindfulness\pd+t2_tse_tra

TA: 2:11 PM: REF Voxel size: 0.3×0.3×5.0 mmPAT: 2 Rel. SNR: 1.00 : tse

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3600.0 ms
TE 1	9.4 ms
TE 2	94 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	3600.0 ms
TE 1	9.4 ms
TE 2	94 ms
MTC	Off
Magn. preparation	None
Flip angle	160 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	320
Phase resolution	100 %
Phase partial Fourier	Off
Trajectory	Cartesian

Resolution - Common

Interpolation	On
---------------	----

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	30
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3600.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**System - Miscellaneous**

Positioning mode	REF
Table position	H

System - Miscellaneous

Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3600.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	220 mm
FoV phase	100.0 %
Phase resolution	100 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	2
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	9.36 ms
Bandwidth	252 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	35
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Normal
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	5

Sequence - Assistant

Mode	Min flip angle
Min flip angle	130 deg
Allowed delay	60 s

\\Research\\Investigators\\Webb\\R01_Mindfulness\\t2_tse_dark-fluid_tra

TA: 4:14 PM: FIX Voxel size: 0.7×0.7×5.0 mmPAT: Off Rel. SNR: 1.00 : tir

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	96.9 %
Slice thickness	5.0 mm
TR	9000.0 ms
TE	83.0 ms
Averages	1
Concatenations	2
Filter	Prescan Normalize, Elliptical filter
Coil elements	HC1-7

Contrast - Common

TR	9000.0 ms
TE	83.0 ms
TD	0.0 ms
MTC	Off
Magn. preparation	Slice-sel. IR
TI	2500 ms
Flip angle	150 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Freeze suppressed tissue	On

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	220 mm
FoV phase	96.9 %
Slice thickness	5.0 mm
Base resolution	320
Phase resolution	70 %

Resolution - Common

Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	220 mm
FoV phase	96.9 %
Slice thickness	5.0 mm
TR	9000.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	2

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Water suppr.	None
Restore magn.	Off
Special sat.	None

Geometry - Navigator**System - Miscellaneous**

Positioning mode	FIX
Table position	H

System - Miscellaneous

Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	90.00 deg
R >> L	214 mm
A >> P	220 mm
F >> H	149 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	9000.0 ms
Concatenations	2

Physio - Cardiac

Magn. preparation	Slice-sel. IR
TI	2500 ms
Fat suppr.	Fat sat.
Dark blood	Off
FoV read	220 mm
FoV phase	96.9 %
Phase resolution	70 %
Trajectory	Cartesian

Physio - PACE

Resp. control	Off
Concatenations	2

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off

Inline - Common

Save original images	On
----------------------	----

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	On
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	8.28 ms
Bandwidth	289 Hz/Px

Sequence - Part 2

Define	Turbo factor
Echo trains per slice	13
Phase correction	Automatic
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Performance
Hyperecho	Off
WARP	Off
Red. EC sensitivity	Off
Turbo factor	17

Sequence - Assistant

Mode	Off
Allowed delay	60 s

\\Research\\Investigators\\Webb\\R01_Mindfulness\\resolve_4scan_trace_tra_p2_192

TA: 1:41 PM: FIX Voxel size: 1.1×1.1×5.0 mmPAT: 2 Rel. SNR: 1.00 : resolve

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	96
Reference scan mode	EPI/separate

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	On
Unfiltered images	Off

Resolution - Filter Rawdata

Raw filter	On
------------	----

Routine

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3700 ms
TE 1	55 ms
TE 2	93 ms
Concatenations	1
Filter	Raw filter, Prescan Normalize
Coil elements	HC1-7

Geometry - Common

Slice group	1
Slices	25
Dist. factor	20 %
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	3700 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 P0.0 F37.0
L	0.0 mm
P	0.0 mm
H	37.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	Fat sat.
Fat sat. mode	Strong
Special sat.	None

Contrast - Common

TR	3700 ms
TE 1	55 ms
TE 2	93 ms
MTC	Off
Magn. preparation	None
Flip angle	180 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Contrast - Dynamic

Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	220 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off
Readout partial Fourier	5/8
Readout segments	7
Interpolation	Off

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	37 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	Head > Brain
Coil Select Mode	Off - All

System - Adjustments

B0 Shim mode	Standard
B1 Shim mode	TrueForm
Adjust with body coil	On
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L0.0 P0.0 H0.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	220 mm
R >> L	220 mm
F >> H	149 mm
Reset	Off

System - pTx Volumes

B1 Shim mode	TrueForm
--------------	----------

System - Tx/Rx

Frequency 1H	123.249501 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Physio - Signal1

1st Signal/Mode	None
TR	3700 ms
Concatenations	1

Diff - Neuro

Diffusion mode	4-Scan Trace
Diff. directions	4
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	Off
Trace weighted images	On
ADC maps	On
FA maps	Off
Mosaic	Off
Tensor	Off
Noise level	100

Diff - Body

Diffusion mode	4-Scan Trace
Diff. directions	4
Diffusion Scheme	Monopolar
Diff. weightings	2
b-value 1	0 s/mm ²
b-value 2	1000 s/mm ²
b-value 1	1
b-value 2	1
Diff. weighted images	Off
Trace weighted images	On
ADC maps	On
Exponential ADC Maps	Off
FA maps	Off

Diff - Body

Invert Gray Scale	Off
Calculated Image	Off
b-Value >=	0 s/mm ²
Noise level	100

Diff - Composing

Distortion Corr.	Off
------------------	-----

Sequence - Part 1

Introduction	On
Dimension	2D
Contrasts	2
Optimization	Min. TE
Multi-slice mode	Interleaved
Echo spacing	0.32 ms
Bandwidth	766 Hz/Px

Sequence - Part 2

EPI factor	96
RF pulse type	Normal
Gradient mode	Fast
Reacquisition mode	On

Sequence - Assistant

Mode	Off
------	-----