

Table of contents

\\User

KRUSEGROUP

BRAIN

BRAIN

Localizer

ABCD_T1w_MPR_vNav_setter

ABCD_T1w_MPR_vNav

ABCD-SEFieldMap_fMRI_PA

ABCD-SEFieldMap_fMRI_AP

ABCD-fMRI-PA_rest

ABCD-fMRI-AP_rest

ABCD_dMRI_DistortionMap_PA

ABCD_dMRI_DistortionMap_AP

ABCD_dMRI_PA

ABCD_T2w_SPC_vNav_setter

ABCD_T2w_SPC_vNav

ABCD-SEFieldMap_fMRI_PA

ABCD-SEFieldMap_fMRI_AP

ABCD-fMRI-PA_rest

ABCD-fMRI-AP_rest

\\User\KRUSEGROUP\BRAIN\BRAIN\Localizer

TA: 9 sec Coil Selection: Manual Voxel Size: 1.2×1.2×5.0 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	On
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	Off
Auto Store Images	On
Load Images to Stamp Segments	On
Load Images to Graphic Segments	On
Graphic segment	All Segments
Inline Movie	Off

Resolution - Common

FoV Read	300 mm
FoV Phase	100.0 %
Slice Thickness	5.0 mm
Base Resolution	256
Phase Resolution	75 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	Off
Asymmetric Echo	Allowed

Routine

Slice Group	1
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	300 mm
FoV Phase	100.0 %
Slice Thickness	5.0 mm
TR	40.0 ms
TE	3.00 ms
Averages	1
Concatenations	1
AutoAlign	---
Coil Elements	HEA;HEP

Resolution - Filter

Raw Filter	Off
Elliptical Filter	On
Distortion Correction	2D
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slice Group	1
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	2
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	3
Slices	1
Distance Factor	20 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	300 mm
FoV Phase	100.0 %
Slice Thickness	5.0 mm
TR	40.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Contrast - Common

TR	40.0 ms
TE	3.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	15 deg
Fat-Water Contrast	Standard
Dark Blood	Off
Contrasts	1
SWI	Off
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Off

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	2
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Slice Group	3
Position	L0.0 A20.0 H0.0 mm
Orientation	Transversal

Geometry - AutoAlign

Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Saturation Mode	Standard
Special Saturation	None

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Slice-sel.
LR Balancing	Off

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	40.0 ms
Segments	1
Concatenations	1

Physio - Cardiac

Tagging	None
Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FoV Read	300 mm
FoV Phase	100.0 %
Phase Resolution	75 %
Dynamic Mode	Standard

Physio - PACE

Resp. Control	Off
Concatenations	1

Inline - Liver

Liver Registration	Off
Save Original Images	On

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - Cardiac

Magn. Preparation	None
Save Original Images	On
Contrasts	1
TE	3.00 ms
TR	40.0 ms

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Soft Tissue

Wash-in	Off
Wash-out	Off
TTP	Off
PEI	Off
MIP Time	Off
Measurements	1

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	fl
Dimension	2D
Excitation	Slice-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Flow Compensation	None
Bandwidth	260 Hz/Px
Asymmetric Echo	Allowed
Segments	1

Sequence - Part 2

Introduction	On
RF Spoiling	On
Acoustic noise reduction	Off

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_T1w_MPR_vNav_setter

TA: 1 sec Coil Selection: Manual Voxel Size: 8.0×8.0×8.0 mm³ Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	32
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
TR	11.0 ms
TE	4.80 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain
Coil Elements	BC

Contrast - Common

TR	11.0 ms
TE	4.80 ms
MTC	Off
Flip Angle	2 deg
Fat-Water Contrast	Standard
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
Base Resolution	32
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Phase Partial Fourier	Off
Slice Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	32
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
TR	11.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Tune up
B1 Shim	TrueForm
Adjustment Tolerance	Auto

System - Adjustments

Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
A >> P	263 mm
R >> L	350 mm
F >> H	350 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	8.000

Physio - Signal

1st Signal/Mode	None
TR	11.0 ms
Concatenations	1

Sequence - Part 1

Sequence Name	ABCD
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	4882 Hz/Px
Echo Spacing	0.27 ms
Free Echo Spacing	Off
EPI Factor	32

Sequence - Part 2

Introduction	Off
RF Spoiling	On

Sequence - Special

Protocol filename	MPRAGE
-------------------	--------

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_T1w_MPR_vNav

TA: 7:12 min Coil Selection: Manual Voxel Size: 1.0×1.0×1.0 mm³ Acc.: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	On
Graphic segment	1st Segment
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.0 mm
TR	2500.0 ms
TE	2.90 ms
Averages	1
Concatenations	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	2500.0 ms
TE	2.90 ms
Magn. Preparation	Non-sel. IR
T1	1070 ms
Flip Angle	8 deg
Fat-Water Contrast	Fast Water Excitation
Dark Blood	Off
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.0 mm
Base Resolution	256
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	GRAPPA
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	32
Acceleration Factor 3D	1
Phase Partial Fourier	Off
Slice Partial Fourier	Off
Asymmetric Echo	Off
Elliptical Scanning	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.0 mm
TR	2500.0 ms
Multi-Slice Mode	Single Shot
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A20.0 F30.0
L	0.0 mm
A	20.0 mm
F	30.0 mm
Initial Orientation	Sagittal
Initial Rotation	16.00 deg

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	30 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H

System - Miscellaneous

Coil Combination	Adaptive Combine
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Rotation	16.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	176 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	3.000

Physio - Signal

1st Signal/Mode	None
TR	2500.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Fast Water Excitation
Magn. Preparation	Non-sel. IR
TI	1070 ms
Dark Blood	Off
FoV Read	256 mm
FoV Phase	100.0 %
Phase Resolution	100 %
Dynamic Mode	Standard

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - Cardiac

Magn. Preparation	Non-sel. IR
Save Original Images	On
Contrasts	1
TE	2.90 ms
TR	2500.0 ms

Inline - MIP

MIP Sag	Off
MIP Cor	Off

Inline - MIP

MIP Tra	Off
MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	tfl_me
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Fast
Flow Compensation	None
Reordering	Linear
Bandwidth	240 Hz/Px
Echo Spacing	8.10 ms
Asymmetric Echo	Off
Turbo Factor	176

Sequence - Part 2

Introduction	On
RF Spoiling	On
Incr. Gradient Spoiling	Off

Sequence - Special

Readout polarity	Positive
Nav. location	Before
Apply moco to	parent and nav
Remeasure	24 TRs
Reacq. threshold	0.50
Feedback Delay	60 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	On
Add. grad time	0.00 ms
Apply freq to	parent and nav
Averaging	None

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-SEfieldMap_fmRI_PA

TA: 7 sec Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
TE	80.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	7033.0 ms
TE	80.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off

Resolution - Acceleration

Phase Partial Fourier	Off
-----------------------	-----

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm

System - Adjustments

Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	7033.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-SEfieldMap_fmRI_AP

TA: 7 sec Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
TE	80.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	7033.0 ms
TE	80.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off

Resolution - Acceleration

Phase Partial Fourier	Off
-----------------------	-----

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm

System - Adjustments

Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	7033.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-fMRI-PA_rest

TA: 5:14 min Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	800.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	383
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never

System - Adjustments

Assume Silicone	Off
-----------------	-----

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	383
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us

Sequence - Special

Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-fMRI-AP_rest

TA: 5:14 min Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	800.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	383
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never

System - Adjustments

Assume Silicone	Off
-----------------	-----

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	383
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us

Sequence - Special

Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_dMRI_DistortionMap_PA

TA: 25 sec Coil Selection: Manual Voxel Size: 1.7×1.7×1.7 mm³ Acc.: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	12400.0 ms
TE	89.20 ms
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	12400.0 ms
TE	89.20 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
Base Resolution	140
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	12400.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Advanced
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off

System - Adjustments

Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	138 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	12400.0 ms
Multi-band accel. factor	1

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	1

Diff

Diffusion Mode	Free
Diff. Directions	102
Diffusion Scheme	Monopolar
Diff. Weightings	1
b-value	0 s/mm ²
Averages	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	Off
ADC Maps	Off
Exponential ADC Maps	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	1700 Hz/Px
Echo Spacing	0.69 ms
Free Echo Spacing	On
EPI Factor	140

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Part 2

Phase Correction	Internal
------------------	----------

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	On
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_dMRI_DistortionMap_AP

TA: 25 sec Coil Selection: Manual Voxel Size: 1.7×1.7×1.7 mm³ Acc.: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	12400.0 ms
TE	89.20 ms
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	12400.0 ms
TE	89.20 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
Base Resolution	140
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	12400.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Advanced
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off

System - Adjustments

Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	138 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	12400.0 ms
Multi-band accel. factor	1

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	1

Diff

Diffusion Mode	Free
Diff. Directions	102
Diffusion Scheme	Monopolar
Diff. Weightings	1
b-value	0 s/mm ²
Averages	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	Off
ADC Maps	Off
Exponential ADC Maps	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	1700 Hz/Px
Echo Spacing	0.69 ms
Free Echo Spacing	On
EPI Factor	140

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Part 2

Phase Correction	Internal
------------------	----------

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us
SENSE1 coil combine	On
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_dMRI_PA

TA: 7:29 min Coil Selection: Manual Voxel Size: 1.7×1.7×1.7 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	4200.0 ms
TE	89.20 ms
Multi-band accel. factor	3
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	4200.0 ms
TE	89.20 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Enabled
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Multiple Series	Off
Delay in TR	0.00 ms

Resolution - Common

FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
Base Resolution	140
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	81
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	240 mm
FoV Phase	100.0 %
Slice Thickness	1.7 mm
TR	4200.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	3

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Navigator**Geometry - Saturation**

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Advanced
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off

System - Adjustments

Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	240 mm
R >> L	240 mm
F >> H	138 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	4200.0 ms
Multi-band accel. factor	3

Physio - PACE

Resp. Control	Off
Multi-band accel. factor	3

Diff

Diffusion Mode	Free
Diff. Directions	102
Diffusion Scheme	Monopolar
Diff. Weightings	2
b-value 1	0 s/mm ²
b-value 2	3000 s/mm ²
Averages 1	1
Averages 2	1
Dynamic Field Correction	Off
Invert Gray Scale	Off
Diff. Weighted Images	On
Trace Weighted Images	Off
Tensor	Off
FA Maps	On
ADC Maps	Off
Exponential ADC Maps	Off
Calculated Image	Off

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Bandwidth	1700 Hz/Px
Echo Spacing	0.69 ms
Free Echo Spacing	On
EPI Factor	140

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Part 2

RF Spoiling	Off
Phase Correction	Internal

Sequence - Special

Excite pulse duration	3840 us
Refocus pulse duration	7680 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us
Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On
PF omits higher k-space	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_T2w_SPC_vNav_setter

TA: 1 sec Coil Selection: Manual Voxel Size: 8.0×8.0×8.0 mm³ Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	32
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
TR	13.0 ms
TE	6.10 ms
Averages	1
Concatenations	1
AutoAlign	Head > Brain
Coil Elements	BC

Contrast - Common

TR	13.0 ms
TE	6.10 ms
MTC	Off
Flip Angle	2 deg
Fat-Water Contrast	Water Excitation
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement

Resolution - Common

FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
Base Resolution	32
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Phase Partial Fourier	Off
Slice Partial Fourier	6/8

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slab Group	1
Slabs	1
Distance Factor	50 %
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	32
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	8.0 mm
TR	13.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	Head > Brain
Initial Position	L0.0 A20.0 H0.0
L	0.0 mm
A	20.0 mm
H	0.0 mm
Initial Orientation	Sagittal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	0 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto

System - Adjustments

Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A20.0 H0.0 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	256 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	8.000

Physio - Signal

1st Signal/Mode	None
TR	13.0 ms
Concatenations	1

Sequence - Part 1

Sequence Name	ABCD
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Normal
Gradient Mode	Fast
Bandwidth	4882 Hz/Px
Echo Spacing	0.27 ms
Free Echo Spacing	Off
EPI Factor	32

Sequence - Part 2

Introduction	Off
RF Spoiling	On

Sequence - Special

Protocol filename	T2-SPACE
-------------------	----------

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD_T2w_SPC_vNav

TA: 6:35 min Coil Selection: Manual Voxel Size: 1.0×1.0×1.0 mm³ Acc.: 2 Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slab Group	1
Slabs	1
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.00 mm
TR	3200.0 ms
TE	565.00 ms
Averages	1.0
Concatenations	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	3200.0 ms
TE	565.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle Mode	T2 Var
Fat-Water Contrast	Standard
Dark Blood	Off
Blood Suppression	Off
Wrap-up Magn.	None
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Multiple Series	Each Measurement
Reordering	Linear

Resolution - Common

FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.00 mm
Base Resolution	256
Phase Resolution	100 %
Slice Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	GRAPPA
Total Factor	2
Reference Scans	Integrated
Acceleration Factor PE	2
Reference Lines PE	32
Acceleration Factor 3D	1
Phase Partial Fourier	Allowed
Slice Partial Fourier	Off
Elliptical Scanning	On

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Distortion Correction	Off
Normalize	Prescan
Image Filter	Off

Geometry - Common

Slab Group	1
Slabs	1
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
Slices per Slab	176
Phase Oversampling	0 %
Slice Oversampling	0.0 %
FoV Read	256 mm
FoV Phase	100.0 %
Slice Thickness	1.00 mm
TR	3200.0 ms
Concatenations	1

Geometry - AutoAlign

Slab Group	1
Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A20.0 F30.0
L	0.0 mm
A	20.0 mm
F	30.0 mm
Initial Orientation	Sagittal
Initial Rotation	16.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	30 mm
Table Position	F
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H

System - Miscellaneous

Coil Combination	Adaptive Combine
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A20.0 F30.0 mm
Orientation	Sagittal
Rotation	16.00 deg
A >> P	256 mm
F >> H	256 mm
R >> L	176 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Non-sel.

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	3.000

Physio - Signal

1st Signal/Mode	None
Trigger Delay	0 ms
TR	3200.0 ms
Concatenations	1

Physio - Cardiac

Fat-Water Contrast	Standard
Magn. Preparation	None
Dark Blood	Off
FoV Read	256 mm
FoV Phase	100.0 %
Phase Resolution	100 %
Dynamic Mode	Standard

Inline - Subtraction

Subtract	Off
Measurements	1
StdDev	Off
Save Original Images	On

Inline - Cardiac

Magn. Preparation	None
Save Original Images	On
TE	565.00 ms
TR	3200.0 ms

Inline - MIP

MIP Sag	Off
MIP Cor	Off
MIP Tra	Off

Inline - MIP

MIP Time	Off
Radial MIP	Off
Save Original Images	On
MPR Sag	Off
MPR Cor	Off
MPR Tra	Off

Inline - Composing

Inline Composing	Off
------------------	-----

Sequence - Part 1

Sequence Name	spc
Dimension	3D
Excitation	Non-sel.
RF Pulse Type	Fast
Gradient Mode	Fast
Flow Compensation	None
Reordering	Linear
Bandwidth	241 Hz/Px
Echo Spacing	6.28 ms
Turbo Factor	200
Echo Train Duration	1187 ms

Sequence - Part 2

Introduction	On
--------------	----

Sequence - Special

Include Nav.	On
Apply moco to	parent and nav
Remeasure	18 TRs
Reacq. threshold	0.50
Feedback Delay	80 ms
Moco ref. image	Use Temp Ref
K-space streaming	None
ABCD navigator	On
Apply freq to	parent and nav

Sequence - Assistant

SAR Assistant	Off
Allowed Delay	0 s

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-SEfieldMap_fmRI_PA

TA: 7 sec Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
TE	80.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	7033.0 ms
TE	80.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off

Resolution - Acceleration

Phase Partial Fourier	Off
-----------------------	-----

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm

System - Adjustments

Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	7033.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-SEfieldMap_fmRI_AP

TA: 7 sec Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00

Properties

Start measurement without further preparation	On
Wait for User to Start	Off
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
TE	80.00 ms
Averages	1
Multi-band accel. factor	1
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	7033.0 ms
TE	80.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	90 deg
Refocus flip angle	180 deg
Fat-Water Contrast	Fat Saturation
Grad. rev. fat suppr.	Disabled
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	1
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off

Resolution - Acceleration

Phase Partial Fourier	Off
-----------------------	-----

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	7033.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	1

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm

System - Adjustments

Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never
Assume Silicone	Off

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	7033.0 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	1
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epse
Dimension	2D
Excitation	Standard
RF Pulse Type	Normal
Gradient Mode	Performance
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
--------------	-----

Sequence - Special

Min. prep scans	0
Delay before PC scans	0 us

Sequence - Special

SENSE1 coil combine	On
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-fMRI-PA_rest

TA: 5:14 min Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	800.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	383
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never

System - Adjustments

Assume Silicone	Off
-----------------	-----

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	383
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us

Sequence - Special

Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----

\\User\KRUSEGROUP\BRAIN\BRAIN\ABCD-fMRI-AP_rest

TA: 5:14 min Coil Selection: Manual Voxel Size: 2.4×2.4×2.4 mm³ Acc:: None Rel. SNR: 1.00**Properties**

Start measurement without further preparation	On
Wait for User to Start	On
Start measurements	Single Measurement
Prio Recon	Off
Auto Open Inline Display	Off
Auto Close Inline Display	Off
Load Images to MR View&GO	On
Auto Store Images	On
Load Images to Stamp Segments	Off
Load Images to Graphic Segments	Off
Graphic segment	Default
Inline Movie	Off

Routine

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
TE	30.00 ms
Averages	1
Multi-band accel. factor	6
AutoAlign	---
Coil Elements	HEA;HEP

Contrast - Common

TR	800.0 ms
TE	30.00 ms
MTC	Off
Magn. Preparation	None
Flip Angle	52 deg
Fat-Water Contrast	Fat Saturation
Contrasts	1
Reconstruction	Magnitude

Contrast - Dynamic

Dynamic Mode	Standard
Measurements	383
Delay in TR	0.00 ms

Resolution - Common

FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
Base Resolution	90
Phase Resolution	100 %
Interpolation	Off

Resolution - Acceleration

Acceleration mode	None
Advanced Reconstruction	Off
Phase Partial Fourier	Off

Resolution - Filter

Raw Filter	Off
Elliptical Filter	Off
Hamming	Off
Distortion Correction	Off
Normalize	Off

Geometry - Common

Slice Group	1
Slices	60
Distance Factor	0 %
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
Phase Oversampling	0 %
FoV Read	216 mm
FoV Phase	100.0 %
Slice Thickness	2.4 mm
TR	800.0 ms
Multi-Slice Mode	Interleaved
Series	Interleaved
Multi-band accel. factor	6

Geometry - AutoAlign

Slice Group	1
Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Phase Encoding Dir.	A >> P
AutoAlign	---
Initial Position	L0.0 A10.0 H10.0
L	0.0 mm
A	10.0 mm
H	10.0 mm
Initial Orientation	Transversal
Initial Rotation	0.00 deg

Geometry - Saturation

Special Saturation	None
--------------------	------

Geometry - Tim Planning Suite

Set-n-Go Protocol	Off
Table Position	10 mm
Table Position	H
Inline Composing	Off

System - Miscellaneous

Coil Selection	Manual
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combination	Sum of Squares
Matrix Optimization	Off

System - Adjustments

Adjustment Strategy	Standard
B0 Shim	Standard
B1 Shim	TrueForm
Adjustment Tolerance	Auto
Adjust with Body Coil	Off
Confirm Frequency	Never

System - Adjustments

Assume Silicone	Off
-----------------	-----

System - Adjust Volume

Position	L0.0 A10.0 H10.0 mm
Orientation	Transversal
Rotation	0.00 deg
A >> P	216 mm
R >> L	216 mm
F >> H	144 mm
Reset	Off

System - pTx

B1 Shim	TrueForm
Excitation	Standard

System - Tx/Rx

Frequency 1H	123.262540 MHz
? Ref. Amplitude 1H	0.000 V
Reset	Off
Correction Factor	1.00
Image Scaling	1.000

Physio - Signal

1st Signal/Mode	None
TR	800.0 ms
Multi-band accel. factor	6

BOLD

GLM Statistics	Off
Ignore Meas. at Start	0
Ignore After Transition	0
Model Transition States	On
Temp. Highpass Filter	On
Threshold	4.00
Paradigm Size	3
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Active
Motion Correction	Off
Spatial Filter	Off
Measurements	383
Delay in TR	0.00 ms

Sequence - Part 1

Sequence Name	epfid
Dimension	2D
Excitation	Standard
Gradient Mode	Performance
Flow Compensation	None
Bandwidth	2778 Hz/Px
Echo Spacing	0.51 ms
Free Echo Spacing	On
EPI Factor	90

Sequence - Part 2

Introduction	Off
RF Spoiling	Off

Sequence - Special

Excite pulse duration	6600 us
Min. prep scans	0
Min. prep scans SB	0
Delay before PC scans	0 us

Sequence - Special

Single-band images	On
MB LeakBlock kernel	On
MB dual kernel	Off
MB RF phase scramble	Off
Opt. MB RF pulse BW	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
Disable freq. update	Off
Suppress 16-bit DICOM	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Fat saturation FA	110.00 deg
Fat sat. offset	0.00 Hz
Sinc exc. pulse BWTP	5.20
Physio recording	Off
Triggering scheme	Standard

Sequence - Assistant

SAR Assistant	Off
---------------	-----