

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\localizer

TA: 0:19 PAT: Off Voxel size: 2.2×1.1×10.0 mm Rel. SNR: 1.00 SIEMENS: gre

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Phase resolution	50 %
Phase partial Fourier	Off
Interpolation	On
PAT mode	None
Matrix Coil Mode	Auto (CP)
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	On
Mode	Inplane

Routine

Slice group 1	
Slices	5
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 2	
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Slice group 3	
Slices	1
Dist. factor	20 %
Position	Isocenter
Orientation	Coronal
Phase enc. dir.	R >> L
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	10.0 mm
TR	20.0 ms
TE	5.00 ms
Averages	1
Concatenations	7
Filter	Elliptical filter
Coil elements	BC

Contrast

TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	40 deg
Fat suppr.	None
Water suppr.	None
SWI	Off
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
-----------------	-----

Geometry

Multi-slice mode	Sequential
Series	Interleaved
Saturation mode	Standard
Special sat.	None
Tim CT mode	Off

System

Body	On
HEP	Off
HEA	Off
L4	Off
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	---
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Segments	1
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Liver registration	Off
Std-Dev-Sag	Off

SIEMENS MAGNETOM TrioTim syngo MR B17

Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On
Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off

Sequence

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Bandwidth	180 Hz/Px
Flow comp.	No
Allowed delay	0 s
RF pulse type	Fast
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\AAHScout

TA: 0:17

PAT: 2

Voxel size: 1.6×1.6×1.6 mm

Rel. SNR: 1.00

SIEMENS: AALScout

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	20 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0 deg
AutoAlign	Head
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	260 mm
FoV phase	100.0 %
Slice thickness	1.6 mm
TR	3.15 ms
TE	1.37 ms
Averages	1
Concatenations	1
Filter	Normalize
Coil elements	HEA;HEP

Contrast

Flip angle	8 deg
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Resolution

Base resolution	160
Phase resolution	100 %
Slice resolution	69 %
Phase partial Fourier	6/8
Slice partial Fourier	6/8
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	On
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode
Series

Sequential
Ascending

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
Auto Coil Select	Off

Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Inline

Time to center	7.5 s
----------------	-------

Sequence

Introduction	On
Dimension	3D
Asymmetric echo	Weak
Contrasts	1
Bandwidth	550 Hz/Px
RF pulse type	Fast
Gradient mode	Normal
Excitation	Non-sel.
RF spoiling	On

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\NIRS_12ch\T1_mprage_mgh

TA: 4:53 PAT: 2 Voxel size: 1.0×1.0×1.3 mm Rel. SNR: 1.00 USER: tfl_mgh_multiecho

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	Isocenter
Orientation	Sagittal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	128
FoV read	256 mm
FoV phase	100.0 %
Slice thickness	1.33 mm
TR	2100 ms
TE 1	3.31 ms
TE 2	6.99 ms
TE 3	8.85 ms
TE 4	10.71 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

Magn. preparation	Non-sel. IR
TI	1100 ms
Flip angle	7.0 deg
Fat suppr.	None
Water suppr.	None
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	24
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Integrated

Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Single shot
Series	Interleaved

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Sagittal
Rotation	0.00 deg
F >> H	256 mm
A >> P	256 mm
R >> L	171 mm

Physio

1st Signal/Mode	None
Dark blood	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	3D
Elliptical scanning	Off
Asymmetric echo	Off
Contrasts	4
Bandwidth 1	195 Hz/Px

SIEMENS MAGNETOM TrioTim syngo MR B17

Bandwidth 2	651 Hz/Px
Bandwidth 3	651 Hz/Px
Bandwidth 4	651 Hz/Px
Flow comp. 1	No
Flow comp. 2	No
Flow comp. 3	No
Flow comp. 4	No
Echo spacing	13.8 ms
RF pulse type	Fast
Gradient mode	Fast
Excitation	Non-sel.
RF spoiling	On
Readout polarity	Positive
Readout trajectory	Bipolar
Add. scale factor	8.0
Gradient spoiling	Integral
Gradient moment factor	3.0
Siemens reconstruction	On
Save raw k-space data	Off
Averaging	RMS

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\NIRS_12ch\episetup_mb4

TA: 5.7 s PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\resting_state_mb

TA: 6:20 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	721
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\BART

TA: 6:20 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	721
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\colornamingstroop_fast

TA: 5:12 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	590
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\msit_fast

TA: 7:00 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	798
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\serialsub

TA: 4:11 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	473
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\checkerflash_mls

TA: 5:11 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	588
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\RSN_pasl-center on probe

TA: 6:22 PAT: Off Voxel size: 3.5×3.5×5.0 mm Rel. SNR: 1.00 SIEMENS: ep2d_pasl

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	15
Dist. factor	0 %
Position	Isocenter
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	180.00 deg
Phase oversampling	0 %
FoV read	224 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	2500.0 ms
TE	21 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

Inversion time 2	1601.7 ms
Inversion time 1	700 ms
Saturation stop time	1600 ms
Flip angle	90 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	151
Delay in TR	0 ms
Multiple series	Off
Perfusion mode	PICORE Q2T
Inversion time 1	700 ms
Saturation stop time	1600 ms
Inversion time 2	1601.7 ms
Flow limit	100.0 cm/s

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Auto (CP)
Prescan Normalize	On
Raw filter	Off
Elliptical filter	Off

Hamming

Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	50.0 mm
Thickness	100 mm

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	180.00 deg
R >> L	224 mm
A >> P	224 mm
F >> H	75 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	1
Ignore after transition	0
Model transition states	Off
Temp. highpass filter	On
Threshold	4.00
Paradigm size	4
Meas[1]	Baseline
Meas[2]	Active
Meas[3]	Baseline
Meas[4]	Active
Motion correction	On
Interpolation	3D-K-space
Spatial filter	On
Filter setting	2.0

Sequence

Introduction	On
Bandwidth	2298 Hz/Px
Free echo spacing	Off
Echo spacing	0.5 ms

SIEMENS MAGNETOM TrioTim syngo MR B17

EPI factor	64
RF pulse type	Normal
Gradient mode	Fast

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\gre_field_mapping

TA: 2:12

Voxel size: 3.4×1.7×3.5 mm

Rel. SNR: 1.00

SIEMENS: gre_field_mapping

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.5 mm
TR	1000 ms
TE 1	10.00 ms
TE 2	12.46 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	HEA;HEP

Contrast

MTC	Off
Flip angle	90 deg
Fat suppr.	None
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution

Base resolution	128
Phase resolution	50 %
Phase partial Fourier	Off
Interpolation	Off
Matrix Coil Mode	Auto (CP)
Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Sequence

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Bandwidth	260 Hz/Px
Flow comp.	Yes
RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\colornamingstroop_conv

TA: 4:15 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	480
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\msit_bushtiming

TA: 7:00 PAT: Off Voxel size: 3.4×3.4×3.5 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	32
Dist. factor	14 %
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Phase enc. dir.	P >> A
Rotation	-180.00 deg
Phase oversampling	0 %
FoV read	220 mm
FoV phase	100.0 %
Slice thickness	3.50 mm
TR	520 ms
TE	30.0 ms
Multi-band accel. factor	4
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	43 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	798
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	64
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Triple
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	L0.0 P10.9 H20.5
Orientation	Transversal
Rotation	-180.00 deg
R >> L	220 mm
A >> P	220 mm
F >> H	128 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

SIEMENS MAGNETOM TrioTim syngo MR B17

Introduction	Off
Bandwidth	3126 Hz/Px
Echo spacing	0.45 ms
EPI factor	64
Gradient mode	Fast
Excite pulse duration	10500 us
Single-band images	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Log physiology to file	Off
Invert RO/PE polarity	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Triggering scheme	Standard
Starting ignore meas	0
Paradigm size	2
Multiplier	1
Step [1]	1
Step [2]	0

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\Flair axial Low SAR

TA: 2:08 PAT: Off Voxel size: 1.3×0.9×5.5 mm Rel. SNR: 1.00 ! USER: tse_crusher

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	On
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	20
Dist. factor	40 %
Position	R6.9 A9.7 H8.2
Orientation	Transversal
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	240 mm
FoV phase	75.0 %
Slice thickness	5.5 mm
TR	9820 ms
TE	93 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	Slice-sel. IR
TI	2500 ms
Freeze suppressed tissue	Off
Flip angle	180 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Auto (CP)
Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On
Normalize	Off
B1 filter	Off

Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None
Tim CT mode	Off

System

Body	Off
HEP	On
HEA	On
Positioning mode	REF
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	Off
Dimension	2D
Compensate T2 decay	Off
Reduce Motion Sens.	Off
Contrasts	1
Bandwidth	199 Hz/Px

SIEMENS MAGNETOM TrioTim syngo MR B17

Flow comp.	Read
Allowed delay	30 s
Echo spacing	10.3 ms
Define	Turbo factor
Turbo factor	13
Echo trains per slice	12
RF pulse type	Low SAR
Gradient mode	Fast

SIEMENS MAGNETOM TrioTim syngo MR B17

\\USER\McLean Research\Frederick\rtNIRS_12ch\double echo low sar

TA: 1:14 PAT: 2 Voxel size: 1.3×0.9×5.5 mm Rel. SNR: 1.00 SIEMENS: tse

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	20
Dist. factor	40 %
Position	R6.9 A9.7 H8.2
Orientation	Transversal
Phase enc. dir.	R >> L
Rotation	90.00 deg
Phase oversampling	0 %
FoV read	240 mm
FoV phase	75.0 %
Slice thickness	5.5 mm
TR	4000 ms
TE 1	33 ms
TE 2	98 ms
Averages	1
Concatenations	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	150 deg
Fat suppr.	None
Water suppr.	None
Restore magn.	Off
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution

Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Trajectory	Cartesian
Interpolation	Off
PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	26
Matrix Coil Mode	Auto (Triple)
Reference scan mode	Integrated
Image Filter	Off
Distortion Corr.	Off
Unfiltered images	Off
Prescan Normalize	On

Normalize	Off
B1 filter	Off
Raw filter	Off
Elliptical filter	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	Parallel F
Gap	10 mm
Thickness	50 mm
Tim CT mode	Off

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	L >> R
Coronal	P >> A
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Adaptive Combine
AutoAlign	Head > Brain
Auto Coil Select	Default
Shim mode	Tune up
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	Isocenter
Orientation	Transversal
Rotation	0.00 deg
R >> L	350 mm
A >> P	263 mm
F >> H	350 mm

Physio

1st Signal/Mode	None
Dark blood	Off
Resp. control	Off

Inline

Subtract	Off
Std-Dev-Sag	Off
Std-Dev-Cor	Off
Std-Dev-Tra	Off
Std-Dev-Time	Off
MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Sequence

Introduction	On
Dimension	2D

SIEMENS MAGNETOM TrioTim syngo MR B17

Compensate T2 decay	Off
Reduce Motion Sens.	Off
Contrasts	2
Bandwidth	179 Hz/Px
Flow comp.	Read
Allowed delay	0 s
Echo spacing	10.9 ms
Define	Turbo factor
Turbo factor	5
Echo trains per slice	17
RF pulse type	Low SAR
Gradient mode	Fast

Table of contents

\\USER

McLean Research
Frederick

rtNIRS_12ch

Optimized for 12 channel!w/physio monitoring

Head First Supine

Give a 3,2,1,0 countdown before all EPI runs

In instructions 'X' means the session number for this subject

localizer

AAHScout

T1_mprage_mgh

episetup_mb4

resting_state_mb

Start fmri_bart.py

BART

Start fmri_stroop.py

Load colornamingX_fast.sstim

colornamingstroop_fast

Start fmri_stroop.py

Load msitX.sstim

msit_fast

Start fmri_stroop.py

Load serialsubX.sstim

serialsub

Start fmri_checkerflash_testversion

Load flash_6bit_3pertr.fstim

checkerflash_mls

RSN_pasl-center on probe

gre_field_mapping

Skip to Flair if running short on time

Start fmri_stroop.py, load colornamingX_conv.sstim

colornamingstroop_conv

Start fmri_stroop.py, load msit1_bushtimingX.sstim

msit_bushtiming

Flair axial Low SAR

double echo low sar