







Assembled – compartments
placed in device




Non-removable parts of the device




11 removable, individual compartments




Inner vertical wall, side vertical wall, horizontal base, and magnets are part of the rotating component of the device



Vertical wall and the camera are part of the stationary component of the device




Step 1: take inspection image of compartment no. 1




Step 2: if passed,
position compartment
no. 1 at disburse slot


Step 3: after
prescribed time, take
inspection image to
see if compartment
no. 1 has been
removed




Step 4: take inspection image of compartment no. 2




Step 5: if passed,
position compartment
no. 2 at disburse slot




Step 6: after
prescribed time, take
inspection image to
see if compartment
no. 2 has been
removed




Step 7: take inspection image of compartment no. 3




Step 8: if passed,
position compartment
no. 3 at disburse slot




Step 9: after
prescribed time, take
inspection image to
see if compartment
no. 3 has been
removed



Step 10: take inspection image of compartment no. 4



Step 11: if passed,
position compartment
no. 4 at disburse slot



Step 12: after
prescribed time, take
inspection image to
see if compartment
no. 4 has been
removed