In-gel protein staining with Coomassie Brilliant Blue

Coomassie Brilliant Blue is a pigment used to stain proteins in acrylamide or agarose gels with a detection limit of about 40 ng per band. The staining is permanent and incompatible with Western blotting. However, Coomassie R-250 can also stain proteins on PVDF membranes. While silver staining SOP 0011 offers higher sensitivity, Coomassie-stained gels are compatible with downstream mass spectrometry.

Risk assessment

- **Glacial acetic acid causes severe skin burns and eye damage**
- Work with flammable liquids
- ▷ Wear gloves, safety glasses, lab coat
- ☐ Collect as HAZARDOUS WASTE



Reviewed: Oct 26, 2026

Procedures

›› *Coomassie Brilliant Blue staining*

- ☐ Incubation tray
- ☐ Lint-free wipe
- ☐ Coomassie destaining solution (20 mL / 1 rxn) R0123
- ☐ Coomassie staining solution (20 mL / 1 rxn) R0124

(1.) *Optional:* Place gel in a container with water. Microwave at 1 000 W for 1 min.

This is why: This helps remove residual SDS and speeds up destaining. Don't boil; gels can melt or crack if microwaved dry.

(2.) Transfer gel into a tray with Coomassie staining solution. Rock gently at room temperature for 15 min. 15 min

(3.) Discard staining solution. Rinse briefly with water to remove excess dye.

(4.) Add Coomassie destaining solution. Rock at room temperature for 45 min. 45 min

(5.) Pour off the destaining solution into a collection bottle for reuse. Replace with water and add a lint-free wipe in the tray to soak up free dye. Destain overnight until background disappears.

Troubleshooting

Coomassie Brilliant Blue staining

In Step 2:

- Faint or invisible bands
  - Increase staining time or protein loading. Ensure sufficient protein was loaded. The detection limit is approximately 40 ng per band.
  - Verify staining solution has not been exhausted from repeated reuse.
- Uneven staining across the gel
  - Ensure the gel is fully submerged and rocking evenly. Use about 20 mL staining solution per mini gel.

In Step 4:

- High background that does not clear with extended destaining
  - Replace destaining solution more frequently. Add a crumpled lint-free wipe or activated charcoal to absorb free dye.
  - Use a colloidal G-250 stain, which gives a clear background [NAT+88].

Materials

Coomassie Brilliant Blue staining

- ☐ **Incubation tray**
- ☐ **Lint-free wipe**
- ☐ **Coomassie destaining solution** ◇ R0123 (20 mL / 1 rxn)
- ☐ **Coomassie staining solution** ◇ R0124 (20 mL / 1 rxn)

Recipes

Coomassie destaining solution

Amount	Ingredient		Stock	Final
1.2 L	Isopropyl alcohol	[67-63-0]	100%	30%
	Flammable liquid (H225)			
2.4 L	Water, reagent-grade	[7732-18-5]		
0.4 L	Acetic acid, glacial	[64-19-7]	99.6%	10%
	Skin corrosion (H314)			

Add concentrated acid slowly to the stirred water/alcohol mixture. DO NOT autoclave. ***Safety:*** Work inside a certified chemical fume hood. ***Note:*** The destaining solution can be reused several times and refreshed using activated charcoal. ***This is why:*** In this adaptation, 30% isopropyl alcohol or ethanol replaces the more toxic 45% methanol.

Coomassie destaining solution

30% Isopropyl alcohol, 10% Acetic acid



DANGER

Flammable liquid; Skin and serious eye irritation; Drowsiness or dizziness

- ☐ Collect as ORGANIC SOLVENT WASTE
- ☐ Keep away from oxidizers

Date: Sign: R0123

Coomassie staining solution

Amount	Ingredient		Stock	Final
1.0 g	Brilliant Blue G	[6104-58-1]		0.5 g/L
2.0 L	Coomassie destaining solution	◇ R0123		
	Flammable liquid (H225); Skin irritation (H315); Serious eye irritation (H319)			

Note: The staining solution can be used multiple times. ***Hint:*** Coomassie R-250 stains faster, while G-250 is available in more sensitive, colloidal formulations.

Coomassie staining solution



DANGER

Flammable liquid; Skin and serious eye irritation; Drowsiness or dizziness

- ☐ Collect as ORGANIC SOLVENT WASTE
- ☐ Keep away from oxidizers

Date: Sign: R0124

List of references

G. Fairbanks, T.L. Steck, and D.F. Wallach, *Biochemistry* **10**(13), 2606-2617 (1971).

P. Matsudaira, *J. Biol. Chem.* **262**(21), 10035-10038 (1987).

V. Neuhoff, N. Arold, D. Taube, and W. Ehrhardt, *Electrophoresis* **9**(6), 255-262 (1988).

Change log

2020-06-12 **Hai T. Dao** Initial protocol; Tom Muir lab "Coomassie stain and destain solution".

2023-01-16 **Benjamin C. Buchmuller** Adaptation as SOP. Replaced methanol with less toxic isopropyl alcohol, from the Daniel Summerer lab "Coomassie staining" protocol.

2026-10-26 **Benjamin C. Buchmuller** Credited Fairbanks et al. (1971), the Tom Muir and Daniel Summerer labs, and Bio-Rad Bulletin 6040; corrected the detection limit to about 40 ng and the membrane stain to R-250 on PVDF.

From the *Lab Protocols* collection by B. C. Buchmuller and contributors (2026). Licensed under Creative Commons Attribution Share Alike 4.0 International; see https://creativecommons.org/licenses/by-sa/4.0/ for the terms.

For research use. Provided in good faith, without warranty or liability for any use or results. Users are responsible for compliance with local regulations and institutional policies.

Current when printed. Visit https://benjbuch.github.io/check/ or scan the QR code to check for updates.

50f45eb