Counting cells

Risk assessment

– Work with human-derived material or transgenic cell lines (BSL-2)

▷ Wear gloves, safety glasses, lab coat

◻ Dispose waste after inactivation as REGULATED MEDICAL WASTE

Reviewed: Mar 14, 2023

This is a bench card. Full protocol available online.

›› *Cell counting with a Neubauer cell chamber*

(1.) *Optional:* Gently mix the cell suspension with an equal volume of 0.4% Trypan blue.

(2.) Clean the hemocytometer and coverslip with a lint-free tissue to remove dust and debris.

(3.) Moisten the raised glass rails of the hemocytometer. Place the coverslip over the counting surface.

(4.) Carefully load about 10 µL of the sample into the chamber. The void will fill by capillary action.

(5.) Place the hemocytometer on the microscope stage and focus. Cells should be evenly distributed and not overlapping.

- Using the hand tally, count the cells in the four corner grids of the chamber.
- Exclude cells on the left or bottom border, count cells on the right and top border (or vice versa).
- Tally unstained (live) and stained (dead) cells separately.

Quality assurance: Repeat with a more concentrated sample if the overall tally is 200 cells or less.

(6.) Average the number of cells across the counted grids and multiply by $1 \times 10^4\,\mathrm{mL}^{-1}$ to determine the cell concentration. Be sure to account for any dilution factor introduced by Trypan blue.

Recipe (available online) Resources (available online) Troubleshooting (available online) Notes (available online)

Lab Protocols — © B. C. Buchmuller. Licensed CC-BY-SA-4.0. Research use; provided as-is, without warranty. Current when printed. Visit https://benjbuch.github.io/check/?q=de5f1a2 or scan the QR code to read the full protocol or to check for updates.