NRC FORM 591M
(09-11-2024)

U.S. NUCLEAR REGULATORY COMMISSION

Materials Inspection Report

1. Licensee/Location Inspected:	2. NRC/Regional Office
Dukes Health System, LLC d/b/a Dukes Memorial Hospital 275 West 12th Street Peru, Indiana 46970-1698 Report Number(s) 2025001	Region III U. S. Nuclear Regulatory Commission 2056 Westings Avenue, Suite 400 Naperville, IL 60563-2657

3. Docket Number(s)	4. License Number(s)	5. Date(s) of Inspection
030-14065	13-18703-01	Aug. 29 2025

LICENSEE:
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

☒ 1. Based on the inspection findings, no violations were identified.

☐ 2. Previous violation(s) closed.

☐ 3. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements, and were assessed at Severity Level IV, in accordance with the NRC Enforcement Policy.

A. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy were satisfied.
(Non-cited violation(s) was/were discussed involving the following requirement(s)

B. The following violation(s) is/are being cited in accordance with NRC Enforcement Policy. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE AND DATE
LICENSEE'S REPRESENTATIVE		
NRC INSPECTOR	Geoffrey Warren, Sr. HP	8/29/25
BRANCH CHIEF	Rhex Edwards	Digitally signed by RHEX EDWARDS Date: 2025.12.12 16:12:09 -06'00'

Add Continuation Page

NRC FORM 592M
(10-04-2022)

U.S. NUCLEAR REGULATORY COMMISSION

Materials Inspection Record

1. Licensee Name:	2. Docket Number(s):	3. License Number(s)
Dukes Health System, LLC	030-14065	13-18703-01

4. Report Number(s):	5. Date(s) of Inspection:
2025001	August 29, 2025

6. Inspector(s):	7. Program Code(s):	8. Priority:	9. Inspection Guidance Used:
Geoffrey Warren, Sr. HP	02120 7C	3	IP 87130

10. Licensee Contact Name(s):	11. Licensee E-mail Address:	12. Licensee Telephone Number(s):
Tanya Oaks - NMT Rik Stephens, MD - RSO	toaks@dukesmemorialhosp.com	765-472-8026

13. Inspection Type: ☐ Initial ☑ Routine ☐ Announced ☐ Non-Routine ☑ Unannounced

14. Locations Inspected: ☐ Hybrid ☑ Main Office ☐ Field Office ☐ Temporary Job Site ☐ Remote

15. Next Inspection Date (MM/DD/YYYY): 08/29/2028 ☑ Normal ☐ Extended ☐ Reduced ☐ No change

16. Location(s) Inspected List:
275 West 12 St Peru, IN 46970-1698

17. Scope and Observations:

This was a routine, unannounced inspection of the Dukes Health System hospital in Peru, Indiana. The licensee was authorized for diagnostic and therapeutic uses of radioactive materials under 10 CFR Parts 35.100, 35.200, and 35.300. The hospital employed one full-time technologist and one PRN who typically worked Monday through Friday, performing 3 - 5 diagnostic administrations per week, typically gastric emptying, bone, and cardiac imaging studies. Nuclear medicine staff had conducted one therapy procedure, an I-131 administration, since the previous inspection. The hospital received unit doses from a licensed nuclear pharmacy; drivers were escorted by the technologist or security. The licensee utilized an outside consultant for audits.

Performance Observations: This inspection consisted of a tour of the hot lab, package receipt observation of dose administration and preparation, interviews with licensee representatives and a review of records. The inspector observed one diagnostic administration of Tc-99m. Records of dose, area wipes, and surveys were reviewed from the date of previous inspection to the date of the inspection. There were no elevated doses or elevated levels for area wipes and surveys. The staff had up to date radiation safety and hazmat training. The inspector reviewed records, including those for dosimetry, area wipes, surveys, package receipt, waste disposal, source inventory and leak tests. The written directive for the I-131 therapy was reviewed and no violations were identified. Dosimetry was reviewed, and no overexposures were reported. The inspector noted that the authorizations for depleted uranium shielding and prepackaged kits had not been used since the previous inspection, and discussed with the licensee that they could remove the authorizations during the upcoming license renewal if they did not need to maintain them.

No violations were identified as a result of this inspection.

Signature and Date - Branch Chief

Digitally signed by RHEX EDWARDS
Date: 2025.12.12 16:11:46 -06'00'