

Inspection Report:

Order number: 43652952

Job Number: 200

Sold To: EYHURST COURT LTD

SITE INFORMATION

Customer Name: BIRTLEY HOUSE
 Street: BRAMLEY
 City: GUILDFORD
 Post Code: GU5 0LB

PRODUCT INFORMATION

Material: OTHER OVERHEADS
 Serial number: VOYA-11319
 Asset Tag:
 Product Classification:
 Latest Location: 1ST FLR ROOM 21

WORK ORDER INFORMATION

Order activity type: M Maintenance
 Work Order Description: SERVICE VISIT
 Customer Order Number: TW180376

CONTACT INFORMATION

Site Contact: BIRTLEY HOUSE
 Site Contact Email:

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments
Activity	Specification	Engineers Comments
Create Quotation	As remarks	

GENERAL INFORMATION

External notes
Annual service completed. Found left hand raise/lower push buttons not working, quotation raised to replace casing. Part: VOY700081706 casing plus one hours labour.

EQUIPMENT CHECKS

Task	Specification	Engineers Comments
Product Located		
Risk Assessment Performed		
Safe to operate		

PARTS USED

Product Code	Product Description	Qty	Accounting Indicator
--------------	---------------------	-----	----------------------

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Accounting Indicator
WORK	31.01.2019 11:00:00	31.01.2019 11:30:00	CONTRACT

Inspection Report:

Order number: 43652952

Job Number: 200

Sold To: EYHURST COURT LTD

THOROUGH EXAMINATION LIFT

SITE INFORMATION

Account Name: EYHURST COURT LTD
Customer Name: BIRTLEY HOUSE
Street: BRAMLEY
Post Code: GU5 0LB

Product and job information

Material: OTHER OVERHEADS
Serial number: VOYA-11319
Asset Tag:
Product Classification:
Date of Manufacture (if known): /
Latest Location: 1ST FLR ROOM 21
Order number: 43652952
Inspection Plan: 00028487/01/00000001
Date of Test: 20190131
Customer Order Number: 20180322

Examination Details

EXAMINATION DETAILS		
Safe working Load(s) (kg)	200	Accepted
	(No remarks)	
	Test Equipment	n/a
Date of Last Thorough Examination	25.01.2018	Accepted
	(No remarks)	
	Test Equipment	n/a
Date of Thorough Examination	30.01.2019	Accepted
	(No remarks)	
	Test Equipment	n/a
Date of Next Thorough Examination	31.01.2020	Accepted
	(No remarks)	
	Test Equipment	n/a
Is this the first examination?	No	
	(No remarks)	
If yes, assembled & installed correctly?	Yes	
	(No remarks)	
Examination interval?	Within 6 months	
	(No remarks)	
Any defects identified?	No	
	(No remarks)	

Inspection Report:

Order number: 43652952

Job Number: 200

Sold To: EYHURST COURT LTD

Is the defect of immediate danger?	No	
	(No remarks)	
Date when defect could become a danger?		n/a
	(No remarks)	
	Test Equipment	n/a
Details of remedy required ?	No	
	(No remarks)	
Details of tests carried out?	Yes	
	(No remarks)	
Is the equipment safe to operate?	Yes	
	(No remarks)	

Technician's signature

N Prosser




Date: 31.01.2019

Email:

Customer's signature

See summary report



Date: 31.01.2019

Email:
Tim@birtleyhouse.co.uk