

Inspection Report:

Order number: 43828793

Job Number: 20

Sold To: EYHURST COURT LTD

SITE INFORMATION

Customer Name: BIRTLEY HOUSE
 Street: BRAMLEY
 City: GUILDFORD
 Post Code: GU5 0LB

PRODUCT INFORMATION

Material: MAXITWIN HOIST
 Serial number: 907001386
 Asset Tag:
 Product Classification:
 Latest Location: Terrace Apt main house gnd flr

WORK ORDER INFORMATION

Order activity type: R Contract Repair
 Work Order Description: RAISING /LOWERING FAULT.
 LOCATED TERRACE
 Customer Order Number: TW180376

CONTACT INFORMATION

Site Contact: BIRTLEY HOUSE
 Site Contact Email:

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments
Device user not following the IFU		
Activity	Specification	Engineers Comments

GENERAL INFORMATION

External notes
Found emergency lower handle pulled out and turned, causing actuator to always be in emergency lower mode. Reset handle and tested okay.

EQUIPMENT CHECKS

Task	Specification	Engineers Comments
Product Located		
Risk Assessment Performed		
Safe to operate		

PARTS USED

Product Code	Product Description	Qty	Accounting Indicator
--------------	---------------------	-----	----------------------

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Accounting Indicator
WORK	13.02.2019 15:15:00	13.02.2019 15:30:00	SERVICE GOODWILL

Inspection Report:


Order number: 43828793

Job Number: 20

Sold To: EYHURST COURT LTD

Technician's signature

N Prosser

A rectangular box containing a handwritten signature in black ink. The signature appears to be 'N Prosser'.

Date: 13.02.2019

Email:

Customer's signature

S Whalley

A rectangular box containing a handwritten signature in black ink. The signature appears to be 'S Whalley'.

Date: 13.02.2019

Email:

Tim@birtleyhouse.co.uk