

LOLER: Load Test Examination Report
Lifting Operations and Equipment Regulations 1998 Schedule 1

Client name & Address:

OpeMed Ltd
Birtley House Group Ltd
Birtley House
Bramley
Guildford
GU5 0LB

Tel:

Location of Examination: Room 16

Make: Opemed **Model:** OT200 **Serial No:** 17090724 **SWL:** 200Kgs

Local Number: **Start-up Date:** 2012 **MA Number:** Item 1 OH

Commissioning Examination: No Safe to Operate: Yes
--

Advisory: Hoist will require lift tape on next visit
--

Schedule: Six Months

Safe to Operate: Yes

Defective Parts (Immediate Attention)

<u>Part No</u>	<u>Description</u>	<u>Defect</u>	<u>Action Taken</u>
----------------	--------------------	---------------	---------------------

Defects requiring rectification at a later date:

<u>Part No</u>	<u>Description</u>	<u>Defect</u>	<u>Action Taken</u>
----------------	--------------------	---------------	---------------------

Conducted according to **BS EN ISO 10535**

Carried out on: Jan 2019 **Next due:** July 2019

Name of Examiner: A.Howes **: Service Engineer:** 08