

Inspection Report:

Order number: 42148172

Job Number: 50

Sold-To Name: EYHURST COURT LTD

SITE INFORMATION

Customer Name: BIRTLEY HOUSE
Street: BRAMLEY
City: GUILDFORD
Post Code: GU5 0LB

PRODUCT INFORMATION

Material: BHM MAXI SKY 600
Serial number: LD451289471
Asset Tag:
Product Classification: Good
Latest Location: GND FLR ANNEX

WORK ORDER INFORMATION

Order activity type: M Maintenance
Work Order Description: Service Maintenance Visit
Customer PO: KA/16012

CONTACT INFORMATION

Site Contact: BIRTLEY HOUSE
Site Contact Email:

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments
Activity	Specification	Engineers Comments

GENERAL INFORMATION

External notes
Annual service completed

PARTS USED

Product Code	Product Description	Qty	Account Indicator
--------------	---------------------	-----	-------------------

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Account Indicator
WORK	25.01.2018 09:30 AM	26.01.2018 03:00 PM	Bronze Contract

Inspection Report:

Order number: 42148172

Job Number: 50

Sold-To Name: EYHURST COURT LTD

THOROUGH EXAMINATION LIFT

SITE INFORMATION

Account Name: EYHURST COURT LTD
Customer Name: BIRTLEY HOUSE
Street: BRAMLEY
Post Code: GU5 0LB

Product and job information

Material: BHM MAXI SKY 600
Serial number: LD451289471
Asset Tag:
Product Classification: Good
Date of Manufacture (if known): 2013/
Latest Location: GND FLR ANNEX
Order number: 42148172
Inspection Plan: 00028487/01/00000001
Date of Test: 20180125
Customer PO: 20170210

Examination Details


EXAMINATION DETAILS		
Safe working Load(s) (kg)	272	Accepted
	(No remarks)	
	Test Equipment	n/a
Date of Last Thorough Examination	PASS	
	12/16	
Date of Thorough Examination	PASS	
	1/18	
Date of Next Thorough Examination	PASS	
	1/19	
Is this the first examination?	No	
	(No remarks)	
If yes, assembled & installed correctly?	Yes	
	(No remarks)	
Examination interval?	Within 12 months	
	(No remarks)	
Any defects identified?	No	
	(No remarks)	
Is the defect of immediate danger?	No	
	(No remarks)	
Date when defect could become a danger?	PASS	
	(No remarks)	
Details of remedy required ?	No	
	(No remarks)	
Details of tests carried out?	Yes	
	(No remarks)	

Inspection Report:
Order number: 42148172
Job Number: 50

Sold-To Name: EYHURST COURT LTD


Is the equipment safe to operate?	Yes
	(No remarks)

Technician's signature
N Prosser



Email:

Customer's signature
See summary report



Email:
Maintenance@birtleyhouse.co.uk
tim@birtley.co.uk