

Inspection Report:

Order number: 42148172

Job Number: 200

Sold-To Name: EYHURST COURT LTD

SITE INFORMATION

Customer Name: BIRTLEY HOUSE
Street: BRAMLEY
City: GUILDFORD
Post Code: GU5 0LB

PRODUCT INFORMATION

Material: SARA 3000 EU 1M2B
Serial number: STLA0804G272
Asset Tag:
Product Classification: Fair
Latest Location: 1st Flr Magnolia room

WORK ORDER INFORMATION

Order activity type: M Maintenance
Work Order Description: Service Maintenance Visit
Customer PO: KA/16012

CONTACT INFORMATION

Site Contact: BIRTLEY HOUSE
Site Contact Email:

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments
Activity	Specification	Engineers Comments

GENERAL INFORMATION

External notes
Annual service completed. Replaced split handgrip.

PARTS USED

Product Code	Product Description	Qty	Account Indicator
HMX570	HANDGRIP BLACK LIFTING ARM	1 EA	Bronze Contract

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Account Indicator
WORK	26.01.2018 09:30 AM	26.01.2018 12:00 PM	CONTRACT

Inspection Report:

Order number: 42148172

Job Number: 200

Sold-To Name: EYHURST COURT LTD

THOROUGH EXAMINATION LIFT

SITE INFORMATION

Account Name: EYHURST COURT LTD
Customer Name: BIRTLEY HOUSE
Street: BRAMLEY
Post Code: GU5 0LB

Product and job information

Material: SARA 3000 EU 1M2B
Serial number: STLA0804G272
Asset Tag:
Product Classification: Fair
Date of Manufacture (if known): 2008/
Latest Location: 1st Flr Magnolia room
Order number: 42148172
Inspection Plan: 00028487/01/00000001
Date of Test: 20180126
Customer PO: 20170210

Examination Details

EXAMINATION DETAILS		
Safe working Load(s) (kg)	200	Accepted
	(No remarks)	
	Test Equipment	n/a
Date of Last Thorough Examination	PASS	
	12/16	
Date of Thorough Examination	PASS	
	1/18	
Date of Next Thorough Examination	PASS	
	1/19	
Is this the first examination?	No	
	(No remarks)	
If yes, assembled & installed correctly?	Yes	
	(No remarks)	
Examination interval?	Within 12 months	
	(No remarks)	
Any defects identified?	No	
	(No remarks)	
Is the defect of immediate danger?	No	
	(No remarks)	
Date when defect could become a danger?	PASS	
	(No remarks)	
Details of remedy required ?	No	
	(No remarks)	
Details of tests carried out?	Yes	
	(No remarks)	

Inspection Report:
Order number: 42148172
Job Number: 200

Sold-To Name: EYHURST COURT LTD

Is the equipment safe to operate?	Yes
	(No remarks)

Technician's signature
N Prosser



Email:

Customer's signature
See summary report



Email:
Maintenance@birtleyhouse.co.uk
tim@birtleyhouse.co.uk