

Inspection Report:

Order number: 43752563

Job Number: 20 Sold To: EYHURST COURT LTD

SITE INFORMATION

PRODUCT INFORMATION

Customer Name: BIRTLEY HOUSE Street: **BRAMLEY**

Material: MALIBU 17, LEFT, HI/LO, LIFT Serial number: 1309000986

City: **GUILDFORD** **Asset Tag:**

Post Code: GU5 0LB **Product Classification: Latest Location: GND FLR ANNEX**

WORK ORDER INFORMATION

CONTACT INFORMATION

Order activity type: R Contract Repair tim whalley 01483 892055 **Site Contact: BIRTLEY HOUSE** Site Contact Email:

Work Order Description:

Customer Order Number: TW180376

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments	
Controls not functioning			
Activity	Specification	Engineers Comments	

GENERAL INFORMATION

External notes	
Replaced faulty seat limit PCB and tested okay.	

EQUIPMENT CHECKS

Task	Specification	Engineers Comments
Product Located		
Risk Assessment Performed		
Safe to operate		

PARTS USED

Product Code	Product Description	Qty	Accounting Indicator
8562148	END LIMIT PCB BOARD ASSY	1 EA	Bronze Contract

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Accounting Indicator
WORK	07.03.2019 12:15:00	07.03.2019 13:00:00	SERVICE GOODWILL



Inspection Report: Order number: 43752563

Job Number: 20

Technician's signature N Prosser

Date: 07.03.2019

Email:

Sold To: EYHURST COURT LTD

Customer's signature S Whalley

Date: 07.03.2019

Email:

Tim@birtleyhouse.co.uk