

Inspection Report:

Order number: 43652952

Job Number: 160 Sold To: EYHURST COURT LTD

SITE INFORMATION

PRODUCT INFORMATION

Customer Name:BIRTLEY HOUSEMaterial:BEDPAN WASHERStreet:BRAMLEYSerial number:818093

City: GUILDFORD Asset Tag:

Post Code: GU5 0LB Product Classification:
Latest Location: 1ST FLOOR

WORK ORDER INFORMATION CONTACT INFORMATION

Order activity type: M Maintenance Site Contact: BIRTLEY HOUSE

Work Order Description: SERVICE VISIT Site Contact Email: Customer Order Number: TW180376

PROBLEM DESCRIPTION AND CORRECTIVE ACTION

Damage	Problem Description	Engineers Comments
Activity	Specification	Engineers Comments
Create Quotation	Qty 1 of 504029302 + 0.5 hrs labour	

GENERAL INFORMATION

Externa	al notes
Annual	service completed. Found descaler bottle empty, quotation raised to replace. Time above 80C = 1' 36", time above 90C = 25".

EQUIPMENT CHECKS

Task	Specification	Engineers Comments
Product Located		
Risk Assessment Performed		
Safe to operate		

PARTS USED

Product Code Product Description Qty Accounting Indicator		Product Description	Qty	Accounting Indicator
---	--	---------------------	-----	----------------------

LABOUR DETAILS

Activity Type	Start Date And Time	End Date And Time	Accounting Indicator
WORK	30.01.2019 13:30:00	30.01.2019 14:00:00	CONTRACT



Inspection Report: Order number: 43652952

Job Number: 160

Technician's signature N Prosser

Date: 31.01.2019

Email:

Sold To: EYHURST COURT LTD

Customer's signature See summary report

Date: 31.01.2019

Email:

Tim@birtleyhouse.co.uk