

**Inspection Report:**

Order number: 43652952

Job Number: 150

Sold To: EYHURST COURT LTD

**SITE INFORMATION**

Customer Name: BIRTLEY HOUSE  
 Street: BRAMLEY  
 City: GUILDFORD  
 Post Code: GU5 0LB

**PRODUCT INFORMATION**

Material: BEDPAN WASHER  
 Serial number: 818095  
 Asset Tag:  
 Product Classification:  
 Latest Location: GROUND FLOOR

**WORK ORDER INFORMATION**

Order activity type: M Maintenance  
 Work Order Description: SERVICE VISIT  
 Customer Order Number: TW180376

**CONTACT INFORMATION**

Site Contact: BIRTLEY HOUSE  
 Site Contact Email:

**PROBLEM DESCRIPTION AND CORRECTIVE ACTION**

Damage	Problem Description	Engineers Comments
Activity	Specification	Engineers Comments
Create Quotation	Qty 1 of 504029302 + 0.5 hrs labour	

**GENERAL INFORMATION**

External notes
Annual service completed. Found descaler fluid bottle empty, quotation raised to replace. Time above 80C = 1' 38", time above 90C = 32".

**EQUIPMENT CHECKS**

Task	Specification	Engineers Comments
Product Located		
Risk Assessment Performed		
Safe to operate		

**PARTS USED**

Product Code	Product Description	Qty	Accounting Indicator
--------------	---------------------	-----	----------------------

**LABOUR DETAILS**

Activity Type	Start Date And Time	End Date And Time	Accounting Indicator
WORK	31.01.2019 09:45:00	31.01.2019 10:15:00	Bronze Contract

## Inspection Report:


Order number: 43652952

Job Number: 150

Sold To: EYHURST COURT LTD

## Technician's signature

N Prosser




Date: 31.01.2019

Email:

## Customer's signature

See summary report



Date: 31.01.2019

Email:

Tim@birtleyhouse.co.uk