

## **LOLER: Load Test Examination Report**

Lifting Operations and Equipment Regulations 1998 Schedule 1

**Client name & Address:**

OpeMed Ltd  
Birtley House Group Ltd  
Birtley House  
Bramley  
Guildford  
GU5 0LB

**Tel:**

**Location of Examination: Room 33**

**Make: Opemed Model: OT200 Serial No: 17090719 SWL: 200Kgs**

**Local Number: Start-up Date: 2012 MA Number: Item 4 OH**

<b>Commissioning Examination:</b> No	<b>Safe to Operate:</b> Yes
--------------------------------------	-----------------------------

**Advisory:**

**Schedule: Six Months**

**Safe to Operate: Yes**

**Defective Parts (Immediate Attention)**

<u>Part No</u>	<u>Description</u>	<u>Defect</u>	<u>Action Taken</u>
----------------	--------------------	---------------	---------------------

**Defects requiring rectification at a later date:**

<u>Part No</u>	<u>Description</u>	<u>Defect</u>	<u>Action Taken</u>
----------------	--------------------	---------------	---------------------

**Conducted according to BS EN ISO 10535**

**Carried out on: Jan 2019 Next due: July 2019**

**Name of Examiner: A.Howes : Service Engineer: 08**