

LOLER: Load Test Examination Report
Lifting Operations and Equipment Regulations 1998 Schedule 1

Client name & Address:

OpeMed Ltd
Birtley House Group Ltd
Birtley House
Bramley
Guildford
GU5 0LB

Tel:

Location of Examination: Room 38

Make: Arjo Model: V4 Serial No: V4-4208-461717 SWL: 200Kgs

Local Number: Start-up Date: 2008 MA Number: Item 5 OH

Commissioning Examination:	No	Safe to Operate:	Yes
-----------------------------------	-----------	-------------------------	------------

Advisory: Replaced 12v 5ah Batteries

Schedule: Six Months

Safe to Operate: Yes

Defective Parts (Immediate Attention)

Part No	Description	Defect	Action Taken
----------------	--------------------	---------------	---------------------

Defects requiring rectification at a later date:

Part No	Description	Defect	Action Taken
----------------	--------------------	---------------	---------------------

Conducted according to BS EN ISO 10535

Carried out on: Jan 2019 Next due: July 2019

Name of Examiner: A.Howes : Service Engineer: 08