PDF page 1:

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/XXX

Appliance Serial Number: 867652075064352

Report Creation Time: P158DT6H11M48S

Upper alarm limit: Above +8.0 ℃ for 10 h

Lower alarm limit: Below -0.5 ℃ for 1 h

			Low Temperature			High Temperature								
Day	Average storage temp. [℃]	Status	Min. temp. [℃]	Total time below -0.5℃ (hh:mm)	Total low alarm time (hh:mm)	Max. temp. [℃]	Total time above +8.0℃ (hh:mm)	Total high alarm time (hh:mm)	Door open count	Door open (hh:mm)	AC Power availability (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp	Faults
today	5.96	ALARM	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	12:30	02:40	NaN	208
2	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:57	NaN	
3	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:53	NaN	
4	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:54	NaN	
5	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	4	00:01	23:44	04:51	NaN	
6	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:59	NaN	
7	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:50	NaN	
8	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:49	NaN	
9	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	05:01	NaN	
10	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:45	04:50	NaN	
11	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	05:01	NaN	
12	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:52	NaN	
13	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:50	NaN	
14	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:51	NaN	
15	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:49	NaN	
16	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:53	NaN	
17	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:49	NaN	
18	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:53	NaN	
19	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	05:00	NaN	
20	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:59	NaN	
21	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	04:52	NaN	
22	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	05:06	NaN	
23	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:56	NaN	
24	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:55	NaN	
25	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	1	00:00	23:59	05:04	NaN	
26	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:59	NaN	
27	5.94	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	05:01	NaN	
28	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	04:55	NaN	
29	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	05:05	NaN	
30	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	04:51	NaN	
31	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	05:05	NaN	
32	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:44	04:56	NaN	

PDF page 2:

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/XXX

Appliance Serial Number: 867652075064352

Report Creation Time: P158DT6H11M48S

Upper alarm limit: Above +8.0 ℃ for 10 h

Lower alarm limit: Below -0.5 ℃ for 1 h

			Low Temperature			High Temperature								
Day	Average storage temp. [℃]	Status	Min. temp. [℃]	Total time below -0.5℃ (hh:mm)	Total low alarm time (hh:mm)	Max. temp. [℃]	Total time above +8.0℃ (hh:mm)	Total high alarm time (hh:mm)	Door open count	Door open (hh:mm)	AC Power availability (hh:mm)	Compressor runtime (hh:mm)	Average ambient temp	Faults
33	5.96	ALARM	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:54	NaN	
34	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	23:59	05:04	NaN	
35	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	04:53	NaN	
36	5.96	OK	5.70	00:00	00:00	6.20	00:00	00:00	1	00:00	23:44	04:55	NaN	
37	5.95	OK	5.70	00:00	00:00	6.20	00:00	00:00	0	00:00	24:00	05:04	NaN	
38	NaN	OK	NaN	00:00	00:00	NaN	00:00	00:00	0	00:00	00:00	00:00	NaN	