60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/110e
Appliance Serial Number: 20253830459
Report Creation Time: P182DT0H1M52S
Upper alarm limit Above +10.0 C for 600 seconds
Lower alarm limit Below +2.0 C for 60 seconds

			Low Temperature			High Temperature								
Day	**Average storage temp.(C)**	**Status**	**Min. temp. (C)**	**Total time below +2.0 (hh:mm)**	**Total low alarm time (hh:mm)**	**Max. Temp. (C)**	**Total time above +10.0 (hh:mm)**	**Total high alarm time (hh:mm)**	**Door open count**	**Door open time (min)**	**Power avail. (hh:mm)**	**Compressor runtime (hh:mm)**	**Average ambient temp.(C)**	**Faults**
today														
2	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.5	
3	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:27	26.4	
4	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:28	26.2	
5	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:12	26.3	
6	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:24	26.1	
7	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	25.9	
8	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	25.9	
9	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:26	25.9	
10	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:15	26.0	
11	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.2	
12	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:25	26.2	
13	5.4	OK	5.4	0:00	0:00	5.5	0:00	0:00	0	0	24:00	2:14	26.1	
14	5.4	OK	5.3	0:00	0:00	5.5	0:00	0:00	0	0	24:00	2:13	26.0	
15	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:26	26.0	
16	5.4	OK	5.3	0:00	0:00	5.5	0:00	0:00	0	0	24:00	2:14	26.0	
17	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:14	26.1	
18	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:27	26.5	
19	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.6	
20	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:14	26.7	
21	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:27	26.8	
22	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	26.9	
23	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:26	26.9	
24	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.9	
25	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:28	27.0	
26	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:15	26.9	
27	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.9	
28	5.4	OK	5.3	0:00	0:00	5.7	0:00	0:00	0	0	24:00	2:28	26.8	
29	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:17	26.8	
30	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	27.0	
31	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:18	27.0	
32	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:25	26.8	

60 DAY PERFORMANCE REPORT

Appliance PQS Code: E003/110e
Appliance Serial Number: 20253830459
Report Creation Time: P182DT0H1M52S
Upper alarm limit Above +10.0 C for 600 seconds
Lower alarm limit Below +2.0 C for 60 seconds

			Low Temperature			High Temperature								
Day	**Average storage temp.(C)**	**Status**	**Min. temp. (C)**	**Total time below +2.0 (hh:mm)**	**Total low alarm time (hh:mm)**	**Max. Temp. (C)**	**Total time above +10.0 (hh:mm)**	**Total high alarm time (hh:mm)**	**Door open count**	**Door open time (min)**	**Power avail. (hh:mm)**	**Compressor runtime (hh:mm)**	**Average ambient temp.(C)**	**Faults**
33	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:13	26.7	
34	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:27	26.7	
35	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:19	26.6	
36	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	26.7	
37	5.4	OK	5.3	0:00	0:00	5.7	0:00	0:00	0	0	24:00	2:17	26.7	
38	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:15	26.8	
39	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:18	26.6	
40	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:17	26.5	
41	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	26.6	
42	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:18	26.5	
43	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:20	26.6	
44	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:25	26.8	
45	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:17	26.9	
46	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:17	26.9	
47	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:18	26.9	
48	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:19	26.8	
49	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	27.0	
50	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:18	27.1	
51	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:15	27.2	
52	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:29	27.2	
53	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	27.2	
54	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:14	27.4	
55	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:31	27.3	
56	5.4	OK	5.3	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:16	27.4	
57	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:15	27.6	
58	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:29	27.8	
59	5.4	OK	5.4	0:00	0:00	5.6	0:00	0:00	0	0	24:00	2:14	27.9	
60	5.4	OK	5.4	0:00	0:00	5.7	0:00	0:00	0	0	24:00	2:29	27.7	

Error codes description:

A - Vaccine temp. sensor error
B - Ambient temp. sensor error
C - Ice bank temp. sensor error
D - Surface temp. sensor error
E - Freezer temp. sensor error
F - Vaccine comp. door malfunction
G - Freezer comp. door malfunction
H - Power measurement error
I - Compressor 1 malfunction
J - Compressor 2 malfunction
K - Fan malfunction
Z - Missing/incomplete data records