


Validation: XRF-predicted vs ICP-MS measured Pb


pellet_FP

$r = 0.990$, RMSE = 1719 ppm | BA bias = 303 ppm, LOA [-3063, 3670] | $n = 34$




pellet_intensity

$r = 0.999$, RMSE = 155 ppm | BA bias = -1 ppm, LOA [-309, 308] | $n = 34$



powder_FP

$r = 0.983$, RMSE = 1799 ppm | BA bias = 293 ppm, LOA [-3235, 3821] | $n = 36$



powder_intensity

$r = 0.998$, RMSE = 271 ppm | BA bias = 36 ppm, LOA [-498, 569] | $n = 36$

