



RIE Group Pty Ltd / ABN: 30 612 894 970

2B / 90 Raglan Street, Roma QLD 4455

0427 393 217 / admin@riegroup.com.au

Calibration Test Certificate

2023-10-06 09:46:33

Detail Of Works

Job No.: **R23195**
Works Requested: **new PGM**
Works Performed: **Bump & calibrate**

Client Details

Supervisor: **Jon Hollingworth**
Device User: **Huracan**

Test Result **Pass**

Device

Serial Number: **KA423-1191216** Detector Type: **GasAlertMicroClip**
Manufacturer: **BW Technologies** Next Cal Due: **2024-04-03**

Sensors

Type:	H2S	CO	LEL	O2
Result:	Pass	Pass	Pass	Pass
Final Reading:	25.0 ppm	100.0 ppm	50.0 %LEL	20.9 %vol
Next Calibration Due:	2024-04-03	2024-04-03	2024-04-03	2024-04-03

Set Points

Type:	H2S	CO	LEL	O2
High Alarm:	15.0 ppm	200.0 ppm	10.0 %LEL	23.5 %vol
Low Alarm:	10.0 ppm	25.0 ppm	5.0 %LEL	19.5 %vol
TWA Alarm:	10.0 ppm	25.0 ppm		
STEL Alarm:	15.0 ppm	50.0 ppm		

Options

Datalog Interval: **15 seconds** Unit Programmed: **N/A**
H2S STEL Period: **15 minutes** CO STEL Period: **15 minutes**

Test Station

Dock Serial Number:	Z310-001553			Dock Location:
Used:	Inlet 1: Yes	Inlet 2: Yes	Inlet 3: No	Inlet 4: No
Concentration:	20.9 %	25.0	20.9 %	20.9 %
Type:	Purge	4 Gas Mixture 2.5% vol CH4	Purge	Purge
Notes:				