



2023-10-06 09:42:13

Job No.: **R23195**
Works Requested: **new PGM**
Works Performed: **Bump & calibrate**

Supervisor: **Jon Hollingworth**
Device User: **Huracan**

Serial Number:	KA423-1191216	Detector Type:	GasAlertMicroClip
Manufacturer:	BW Technologies	Next Cal Due:	2024-04-01

Test Result	Pass
-------------	------

Type:	H2S	CO	LEL	O2
Result:	Pass	Pass	Pass	Pass
Next Calibration Due:	2024-04-01	2024-04-01	2024-04-01	2024-04-03
Acoustic Alarm:	Pass	Visual Alarm:	Pass	

Dock Serial Number:	Z310-001553			Dock Location:
	Inlet 1:	Inlet 2:	Inlet 3:	Inlet 4:
Used:	No	Yes	No	No
Concentration:	20.9 %	25.0	20.9 %	20.9 %
Type:	Purge	4 Gas Mixture 2.5% vol CH4	Purge	Purge
Notes:				