



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2021-08-24**

Description: **SBE-32 Carousel**

### Sensor Information:

Model Number: **SBE-32**

Serial Number: **0048**

Replaced the main piston "O"-Rings.

Replaced 6-pin connector.

Replaced 2-actuator magnets.

Replaced main housing.

### Pressure Test Protocol:

Low Pressure Test: **50**      PSI      Held For: **60**      Minutes

High Pressure Test: **0**      PSI      Held For: **0**      Minutes

Passed Test: **True**

Tested By: **BLT**

High pressure is  
generally equal  
to the maximum  
depth rating of  
the instrument

