



EADAS

ABHIMANYU CHOPRA – G00885635

LALITHA BASKHARUNI – G00949261


NISARG PATEL – G00944795

SAMMY LIN – G00429253

PROBLEM STATEMENT

- In person monitoring of the elderly is not always convenient.
- Solution : EADAS – Elderly Asset Distress Alarm System
 - Monitors
 - Orientation & acceleration to detect fall
 - Heart rate to detect spike or fall
 - Location, in case the elderly doesn't respond
 - Alert
 - Sends status/location and time as sms to a set mobile number

HARDWARE BLOCK DIAGRAM




SOFTWARE FLOW DIAGRAM

INITIALIZATION SEQUENCE




SOFTWARE FLOW DIAGRAM

APPLICATION



SMS ALERT EXAMPLE




SOFTWARE TOOLS

- Code Composer Studio v6.1 – IDE for TI microcontrollers
- Git – Source code management
- Kicad – Schematic design
- SketchUp – 3D design
- Open source frameworks & components
 - Arduino C++ Libraries ported to Ti Launchpad in C (Gyro)
 - Implementation approach – pulsesensor.com (Pulse sensor)


PROJECT TIMELINE

A long time ago, in a galaxy
far, far away...

PROJECT TIMELINE



EADAS CASE DESIGN



ISSUES/SOLUTION

- SPI interface – Hard to initially understand and get configuration right on Launchpad
- Timing issues – printf to CCS console caused timing issues and UART/Timer/ADC stopped working as expected
- Pulse Sensor – A very finicky sensor. Took a lot of time to troubleshoot errors
- Integration – Required lots of testing and code changes
- Most components purchased from Sparkfun/Adafruit come with Arduino libraries. Took full advantage and ported to MSP430 C/C++
- ISRs – They had to be really fast
- Didn't include buzzard in PCB design
- 3D Printing – Takes a lot of time



DEMO TIME!

QUESTIONS?



MOTD

