Monarch Accel 3-axis 200g ▪ v1.1 *Acorn* Support

Black Oak Engineering

precision electronic instruments

Standard baud rate is 115200, 8N1.

Device Commands

dvc Basic 'AT' command. What device are you?

DVC='Monarch Accel xyz 200g', expected response.

dco Contact info

DCO='www.blackoakeng.com'

drh Hard reset

dsn Serial number?

dti='HH:MM' Set time. Use 24 hr form of hours. Seconds are reset.

dti DTI='16:11'

der Report any error messages.

DER='Sensor Overflow'

dvf What is your firmware version?

dvh What is your hardware version?

dva What is your Acorn version?

dmo What is your model number?

Copyright © 2021-26 Black Oak Engineering ®. All rights reserved.
'Monarch' and its design are protected trademarks.
All Black Oak Engineering products are designed and manufactured in the US.

Acquisition Commands

aqm What are your xyz measurements?

AQM=’1.23,4.56,7.89g’

aqr=500 Set sample rate in ms.

AQR=500ms

AQR='SR too fast'

Communications Commands

mco+|- Enable | disable serial feed.

mec+|- Enable | disable command echo.

mhd Print column header, e.g., 'time, uT'.

Copyright © 2021-26 Black Oak Engineering ®. All rights reserved.
‘Monarch’ and its design are protected trademarks.
All Black Oak Engineering products are designed and manufactured in the US.