Keysight EDU1002A Oscilloscope Quick Reference

Bandwidth Limit	Press channel # -> select BW limit
Probe	Make sure it is on 10:1 (switched down) instead of 1:1 (up) Verify: Press channel # - > Probe - > Ratio 10:1
White Background	Save/Recall -> Save -> Settings -> Invert Grat
Peak Max	Measure -> Type -> Maximum (scroll with entry button) -> Add measurement
Sample Measurement	Acquire(in Horizontal menu) -> ACQ Mode -> Scroll with entry button to averaging -> scroll # of samples
Help	Hold in any button
CH1-CH2	Math -> Operator to minus