




Republic of the Philippines  
Department of Environment and Natural Resources  
**ENVIRONMENTAL MANAGEMENT BUREAU**  
Regional Office No. VIII  
DENR 8 Compound, Brgy. 2, Jones Extension, Tacloban City  
Telefax: (053) 832-2319  
Email: [r8support@emb.gov.ph](mailto:r8support@emb.gov.ph) [emb8@emb.gov.ph](mailto:emb8@emb.gov.ph)  
Website: [r8.emb.gov.ph](http://r8.emb.gov.ph)



May 11, 2022

**JOHN JOVEN S. CHIONG**

Managing Head  
PASAR Corporation  
LIDE, Isabel, Leyte



**Attention: EDMUND P. ESGUERRA**  
Pollution Control Officer

Sir:


This refers to your submitted Relative Accuracy Test Audit (RATA) Plan received by this Office on 02 May 2022 for the Continuous Emission Monitoring System (CEMS) installed at Desulfurization Plant for source emission concentration of Sulfur Dioxide (SO<sub>2</sub>) to be tested by SGS on May 13, 2022.

Please be informed that same is approved with a directive to comply with the facility's requirements and specifications and test procedures as stated in the test plan (40CFR Part 60, Appendix B, Performance Specification 2- Specifications and Test Procedures for SO<sub>2</sub> and NO<sub>2</sub> CEMS in Stationary Sources and Method 6C: Determination of Sulfur Dioxide Emissions from Stationary Sources (Instrumental Analyzer Procedure).

This Office will send technical personnel to witness the conduct of the audit and observe the conditions of the facility.

Be guided accordingly.

Very truly yours,

  
**REYNALDO B. BARRA, PME.**  
OIC-Regional Director

