

# Pre-Deployment Testing, Augmentation and Calibration of Cross-Sensitive Sensors

Balz Maag, Olga Saukh, David Hasenfratz\*, Lothar Thiele

Computer Engineering and Networks Laboratory, ETH Zurich,  
Switzerland


\*Sensirion AG, Staefa, Switzerland



Eidgenössische Technische Hochschule Zürich  
Swiss Federal Institute of Technology Zurich



# Air Pollution



All Images **News** Videos Books More ▾ Search tools

2015

About 9.450.000 results (0,42 seconds)



[Leeds named hotspot for \*\*air pollution\*\* deaths as epidemic ...](#)

[Yorkshire Evening Post](#) - 46 minutes ago

Almost 700 people in Leeds will die this year from illnesses related to **air pollution**, with an overall cost to the local health economy of £480m, it has been ...

[Should cars be banned from Manchester city centre for one day a ...](#)

[Manchester Evening News](#) - 8 hours ago

[Odd-even will work only if implemented in entire NCR: \*\*Pollution\*\* ...](#)

Highly Cited - [Hindustan Times](#) - 13 hours ago

[Smog Blankets Rajpath Lawns On Republic Day](#)

International - [NDTV](#) - 8 hours ago

[After the Delhi experiment](#)

Opinion - [The Hindu](#) - 25 Jan 2016

[Amazing Things Happened When New Delhi Halved the Number of ...](#)

In-Depth - [VICE News](#) - 25 Jan 2016



[Hindustan Ti...](#)



[Manchester ...](#)



[Winnipeg Fre...](#)



[Houston Chr...](#)



[NDTV](#)



[VICE News](#)

# Active Research and Development

- Numerous research projects and start-ups



**Air Quality Egg**

- Similar approach
  - Small and low-cost air quality monitoring systems


# Low-cost Air Quality Sensors

- Pro's:


- Small
- Cheap (1\$ - 100\$)
- Low power consumption

- Con's:

- Low target pollutant concentrations, often at sensitivity boundaries
- Environmental conditions affect sensor output
- Low selectivity, i.e. sensors are cross-sensitive to multiple substances
- Need frequent re-calibration



SGX Sensortech  
MiCS-OZ-47 O<sub>3</sub>



AlphaSense CO-B4

# Limiting Effects

- Datasheet information
  - Sparse or not provided at all
  - Laboratory results do not cover deployment conditions

## ENVIRONMENTAL

|   |            |
|---|------------|
| Sensitivity @ -20°C (% output @ -20°C/output @ 20°C) @ 2ppm NO <sub>2</sub> | 40 to 70   |
| Sensitivity @ 50°C (% output @ 50°C/output @ 20°C) @ 2ppm NO <sub>2</sub>   | 120 to 135 |
| Zero @ -20°C nA change from 20°C  | ±10        |
| Zero @ 50°C nA change from 20°C   | 60 to 380  |

## CROSS SENSITIVITY

|                               |             |                         |                               |          |
|-------------------------------|-------------|-------------------------|-------------------------------|----------|
| H <sub>2</sub> S              | sensitivity | % measured gas @ 5ppm   | H <sub>2</sub> S              | < -130   |
| NO                            | sensitivity | % measured gas @ 5ppm   | NO                            | < 4      |
| Cl <sub>2</sub>               | sensitivity | % measured gas @ 5ppm   | Cl <sub>2</sub>               | < 100    |
| SO <sub>2</sub>               | sensitivity | % measured gas @ 5ppm   | SO <sub>2</sub>               | < -20    |
| CO                            | sensitivity | % measured gas @ 5ppm   | CO                            | < 0.1    |
| H <sub>2</sub>                | sensitivity | % measured gas @ 100ppm | H <sub>2</sub>                | < 0.1    |
| C <sub>2</sub> H <sub>4</sub> | sensitivity | % measured gas @ 100ppm | C <sub>2</sub> H <sub>4</sub> | < 0.1    |
| NH <sub>3</sub>               | sensitivity | % measured gas @ 20ppm  | NH <sub>3</sub>               | < 0.1    |
| CO <sub>2</sub>               | sensitivity | % measured gas @ 5% Vol | CO <sub>2</sub>               | < 0.1    |
| O <sub>3</sub>                | sensitivity | % measured gas @ 100ppb | O <sub>3</sub>                | 30 to 65 |
| Halothane                     | sensitivity | % measured gas @ 100ppm | Halothane                     | < 0.1    |

secondhand smoke, smoke generated from burning of wood and paper, volatiles of wine (alcohol) and cosmetics, ammonia, hydrogen sulfide, hydrogen, carbon monoxide, propane, methane, styrene, propylene glycol, phenol, acetone, thinner, insecticide, correction fluid, benzene, formaldehyde and so on.


Datasheet, TP401-A Indoor Air Quality Sensor, Shenzhen Dovelet Sensors Technology CO., LTD

Datasheet, NO2-B4 Nitrogen Dioxide Sensor, Alphasense

- Ignoring these effects limits performance
- **Goal:** Understand sensor characteristics under deployment-related conditions

# Example: Alphasense NO<sub>2</sub>-B4 Sensor

- Deployed at high-quality monitoring station
- Ordinary Least-Squares (OLS) calibration to nitrogen dioxide (NO<sub>2</sub>) reference measurements




- Root-Mean-Square-Error (RMSE) = 12.4 ppb (50%)

Sensor is highly cross-sensitive to ozone (O<sub>3</sub>),  
temperature and humidity

# Sensor Calibration


## Simple sensor calibration



Ordinary Least-Squares (OLS):

$$r = \beta_0 + \beta_1 s_1 + \varepsilon$$

## Sensor array calibration




Multiple Least-Squares (MLS):

$$r = \beta_0 + \beta_1 s_1 + \beta_2 s_2 + \beta_3 s_3 + \varepsilon$$

*Used to compensate for cross-sensitivities*

# Example: Alphasense NO<sub>2</sub>-B4 Sensor revised


- Multiple Least-Squares (MLS) sensor array calibration
  - NO<sub>2</sub>-B4, SGX O<sub>3</sub>, humidity and temperature




- RMSE = 4.6 ppb (18%)




# Challenges



# Challenges




# Sensor Testing: Signals




- In-field measurements
  - Measurements next to high-quality monitoring stations, e.g. run by governmental authorities
- Sensor-under-test  $s$
- Various reference signals  $r_i \in R$ , e.g. pollutants, temperature, humidity...
- Standardization for scale-invariant results

# Sensor Testing: Inverse Calibration




- Inverse calibration



- Multiple Least-Squares


$$s = \beta_0 + \beta_1 r_1 + \beta_2 r_2 + \beta_3 r_3 + \varepsilon$$

# Sensor Testing: Regression Error




- Regression estimation  $\hat{u}$ 
  - *Explained* part of the sensor signal with given references
- Regression error  $\varepsilon$ 
  - *Unexplained* part of the sensor signal
- Reason for substantial error can be two-fold
  - Uncaptured cross-sensitivities
  - Sensor noise

# Sensor Testing: Error Decomposition



- FFT of typical  $O_3$  concentration



- Error decomposition: Low-pass filter (cut-off:  $\frac{1}{24h}$ )
  - Low-frequent part  $\varepsilon_P$ : Uncaptured cross-sensitivities
  - High-frequent part  $\varepsilon_N$ : Sensor noise

# Experimental Evaluation


- Various low-cost sensors
- Governmental high-quality station (NABEL) in Duebendorf, Switzerland
  - 20 different reference signals
- 15 months of data



# Alphasense NO<sub>2</sub>-B4 Sensor

Root-Mean-Square  
 $RMS(\cdot) \in [0,1]$

← ■  $RMS(\epsilon_P)$  ■  $RMS(\epsilon_N)$  ■  $RMS(\hat{u})$




High error components when using only NO<sub>2</sub> reference

Decreased error and increased explained signal components when adding O<sub>3</sub>, temperature and humidity

Extending the used references




# SGX O<sub>3</sub> and Alphasense CO-B4



Similar results: Highly sensitive to target gas.  
Adding T & H reduces error components.


# Dovelet Air Quality Sensor





Not sensitive to any pollutants.  
Unqualified sensor for outdoor air quality measurements.

# Testing Conclusion

1. Need  $O_3$ , humidity and temperature measurements to compensate for cross-sensitivities of the  $NO_2$  sensor
2.  $O_3$  and CO sensor depend on humidity and temperature









# Calibration Stability: O<sub>3</sub>

- Calibration error vs. different training frequencies over 12 months
- Training time: 4 weeks



- Decreasing error with increasing calibration frequency
- OLS requires monthly recalibration to achieve same error when calibrating the array every 4 months


# Calibration Stability: CO



- Increasing error at  $f = 2$ : Unstable parameters during summer
- Sensor array calibration beneficial



# Calibration Stability: NO<sub>2</sub>



- Decreasing error
- MLS outperforms OLS

# Conclusions

- Low-cost sensors suffer from **cross-sensitivities** and **meteorological dependencies**
- In-field testing using reference measurements to **explain sensor-under-test**
- **Quantify amount** of captured and uncaptured cross-sensitivities and sensor noise
- **Improved accuracy and stability** when calibrating an augmented sensor array

# Thank You!

**ETH**

Eidgenössische Technische Hochschule Zürich  
Swiss Federal Institute of Technology Zurich

