SPATIALLY VARIABLE DECONVOLUTION FOR LIGHTSHEET MICROSCOPY

Bogdan Toader¹, TODO update Stéphane Chrétien², Andrew Thompson³

¹University of Cambridge, ²University of Lyon 2, ³National Physical Laboratory



LIGHTSHEET MICROSCOPY

TODO: a few words about why lightsheet is good

Light sheet microscopy is a type of fluorescence microscopy used in cell biology due to its fast acquisition times and low photo-damage to the sample. This is achieved by selectively illuminating a slice of the sample using a sheet of light and detecting the emitted fluorescence signal using a dedicated objective orthogonal to the plane of the sheet.

• Project aim: computationally improve the quality of light sheet microscopy images.

LIGHT-SHEET MICROSCOPE





Figure credit: Jörg Ritter, PhD thesis (2011)
A light-sheet microscope consists of two objectives: TODO etc

SPATIALLY VARYING PSF



PROBLEM

Due to the combination of light-sheet and the objective PSF, the effective PSF of the system is spatially varying, and therefore standard deconvolution approaches are not applicable.

IMAGE FORMATION MODEL

FORWARD MODEL

The sample s illuminated at $z = z_0$ by the light-sheet l and the photons are collected by an objective with PSF h:

$$f(x, y, z_0) = \iiint l_{avg_y}(u, v, w) s(u, v, w - z_0) h(x - u, y - v, w) du dv dw$$

where h is given in (1) and l is calculated similarly (TODO explain) TODO: insert diagram of the image formation to explain the forward model

PSF MODEL

TODO We calculate the objective PSF h by using a model that includes defocus etc etc. TODO: reference to paper

$$h(x,y,z) = \left| \iint g_{\sigma} * p(\kappa_x, \kappa_y) e^{2i\pi z} \sqrt{(n/\lambda)^2 - \kappa_x^2 - \kappa_y^2} e^{2i\pi(\kappa_x x + \kappa_y y)} \, d\kappa_x \, d\kappa_y \right|^2$$
 (1)


with the pupil function p defined

$$p(\kappa_x, \kappa_y) = \begin{cases} e^{2i\pi \sum_{j=1}^{15} c_i Z_j(\kappa_x, \kappa_y)} & \text{for } \rho = \sqrt{\kappa_x^2 + \kappa_y^2} \le NA/\lambda, \\ 0, & \text{otherwise.} \end{cases}$$
(2)


where the phase of the pupil function is computed by least-squares fitting of the coefficients of the first 15 Zernike polynomials using an image of a bead.

The parameters of the model are given by the experimental setup used to acquire the image:

- n refractive index
- \bullet λ wave length
- NA numerical aperture
- g_{σ} Gaussian blur to take into account other properties not accounted for in our model



Bead image (MIP)



Estimated objective PSF
(MIP)

Results — Simulated Data

RECONSTRUCTION

TODO: refine this text, also make it take into account that we don't always use L2 for fidelity or TV for regularization. Let \hat{f} be the image data and f(s) the result of applying the forward model to the sample s.

To recover s, we solve:


Find
$$\hat{s} \in \operatorname{argmin}_{s} \left\{ \|\hat{f} - f(s)\|_{L_{2}} + \lambda TV(s) \right\}$$

using a version of the Primal Dual Hybrid Gradient (PDHG) algorithm. TODO: reference.


Results – real data

BEADS

Full resolution image: 1127 x 111 x 100




Data (top), constant PSF deconvolution (middle), model deconvolution (bottom)




Data (top), constant PSF deconvolution (middle), model deconvolution (bottom)

Marchantia

TODO: include the latest results here because better Full resolution image: 1127 x 155 x 100



Data (top), constant PSF deconvolution (middle), model deconvolution (bottom)



Data (top), constant PSF deconvolution (middle), model deconvolution (bottom)

ACKNOWLEDGMENTS

This work is funded by Isaac Newton Trust/Wellcome Trust ISSF/University of Cambridge Joint Research Grants Scheme, RG89305.

REFERENCES