



CERTIFICATE OF CALIBRATION



Godrej & Boyce Mfg.Co.Ltd.,
Calibration Services,MD Babar Business Center,
Godown 3-6,2 floor,Plot BGP-109
Landewadi MIDC Bhosari Pune-411026
Dir.:+91-20-46304224-4223

FR-QT-21			
Certificate No. : P-240102-2-5		Date Of issue : 9-Jan-2024	
Date of Calibration : 4-Jan-2024		Recom. Due Date : 4-Jan-2025	
Customer Details :			
LARSEN & TOUBRO LIMITED		Calibrated At : LAB	
DEFENCE IC ,PLOT NO;A5-,7,8,10&11, NAVLAKHE UMBERE		Cond.on Receipt : SATISFACTORY	
TALEGAON PUNE MIDC TALEGAON TALEGON PUNE		Receipt No. : P-240102-2	
Maharashtra India 410507		Date Of Receipt : 04-01-2024	
Details of Under Calibration instrument		Id. No : TE/ES/0120	
Name : Digital Multimeter		Sr. No : 083059	
Range : AS PER REPORT		L.C. : AS PER REPORT	
Make : Rishabh		Model : 15S	
Location : --			
Details of Master Equipment			
Name	Sr.No.	Valid Upto	Traceability
Multi Product Calibrator	LMGP-126	7-Nov-2024	G&B, MUMBAI
Environmental Details :			
Temperature : 25.6°C		Relative Humidity :53%	
Work Instruction No. :		WI-ET-04	
REMARK :			
1)Calibration Points are taken as per Customer's Request. 2. Discipline-Electro Technical - Alternating Current(<1GHz),Direct Current			
AUTHORIZED SIGNATORY : - SANDEEP MAHADIK			
DESIGNATION : DM			
CALIBRATED BY : Lab Engineer : P.K.DASH			
ULR No : CC249524000000031F			



Godrej & Boyce Mfg. Co. Ltd.

CERTIFICATE OF CALIBRATION

Godrej & Boyce Mfg.Co.Ltd.,
Calibration Services,MD Babar Business Center,
Godown 3-6,2 floor,Plot BGP-109
Landewadi MIDC Bhosari Pune-411026
Dir.:+91-20-46304224-4223

FR-QT-21					
Certificate No. : P-240102-2-5			ULR No : CC24952400000031F		
Range		UUC Reading	Standard reading	Error	Measurement Uncertainty ($\pm$) (%)
	30 mV	3.03mV	3.000	0.03	0.200
		15.02mV	15.000	0.02	0.040
		27.02mV	27.000	0.02	0.022
		-27.01mV	-27.000	-0.01	0.022
DC Voltage:	300 mV	29.9mV	30.000	-0.1	0.190
		149.9mV	150.000	-0.1	0.043
		269.9mV	270.000	-0.1	0.043
		-269.9mV	-270.000	0.1	0.043
	3V	0.300V	0.30000	0.000	0.330
		1.500V	1.50000	0.000	0.067
		2.699V	2.7000	-0.001	0.037
		-2.699V	-2.7000	0.001	0.037
	30V	2.99V	3.0000	-0.01	0.200
		14.99V	15.000	-0.01	0.040
		26.99V	27.000	-0.01	0.093
		-26.98V	-27.000	0.02	0.093
	300V	29.9V	30.000	-0.1	0.200
		149.8V	150.000	-0.2	0.043
		269.7V	270.00	-0.3	0.043
		-269.6V	-270.00	0.4	0.043
	1000V	100V	100.000	0	0.58
		499V	500.00	-1	0.16
		900V	900.00	0	0.086
		-899V	-900.00	1	0.086
AC voltage: 3 V	50Hz	0.297V	0.30000	-0.003	0.33
	50Hz	1.495V	1.50000	-0.005	0.067
	50Hz	2.693V	2.7000	-0.007	0.19
	1kHz	2.696V	2.7000	-0.004	0.067
AC voltage: 30 V	50Hz	2.98V	3.0000	-0.02	0.27
	50Hz	14.95V	15.000	-0.05	0.067
	50Hz	26.91V	27.000	-0.09	0.067
	1kHz	26.94V	27.000	-0.06	0.067
AC voltage: 300 V	50Hz	29.8V	30.000	-0.2	0.067
	50Hz	149.7V	150.000	-0.3	0.075
	50Hz	269.5V	270.00	-0.5	0.22
	1kHz	269.4V	270.00	-0.6	0.22



Godrej & Boyce Mfg. Co. Ltd.

CERTIFICATE OF CALIBRATION

Godrej & Boyce Mfg.Co.Ltd.,
Calibration Services,MD Babar Business Center,
Godown 3-6,2 floor,Plot BGP-109
Landewadi MIDC Bhosari Pune-411026
Dir.:+91-20-46304224-4223

FR-QT-21					
Certificate No. : P-240102-2-5			ULR No : CC24952400000031F		
Range		UUC Reading	Standard reading	Error	Measurement Uncertainty ($\pm$) (%)
AC voltage: 1000 V	50Hz	99V	100.00	-1	0.58
	50Hz	498V	500.00	-2	0.16
	50Hz	898V	900.00	-2	0.11
	1kHz	902V	900.00	2	0.11
DC Current:	3A	0.301A	0.300000	0.001	0.2
		1.502A	1.50000	0.002	0.2
		2.705A	2.70000	0.005	0.2
		-2.704A	-2.70000	-0.004	0.2
	10A	1.00A	1.0000	0.00	0.2
		4.99A	5.0000	-0.01	0.2
		8.99A	9.0000	-0.01	0.2
		-8.99A	-9.0000	0.01	0.2
AC Current: 10A	50Hz	0.99A	1.0000	-0.01	0.58
	50Hz	4.97A	5.0000	-0.03	0.21
	50Hz	8.97A	9.0000	-0.03	0.21
	1kHz	8.99A	9.0000	-0.01	0.21
Resistance:	300 Ω	30.05 Ω	30.000	0.0500	0.07
		150.06 Ω	150.000	0.0600	0.07
		270.09 Ω	270.00	0.090	0.07
	3k Ω	0.301k Ω	0.30000	0.00100	0.067
		1.502k Ω	1.50000	0.00200	0.067
		2.703k Ω	2.7000	0.0030	0.33
	30k Ω	3.00k Ω	3.0000	0.0000	0.2
		14.99k Ω	15.0000	-0.0100	0.073
		26.97k Ω	27.000	-0.030	0.07
	300k Ω	29.9k Ω	30.000	-0.100	0.19
		149.8k Ω	150.000	-0.200	0.07
		269.7k Ω	270.00	-0.30	0.42
	3M Ω	0.299M Ω	0.30000	-0.00100	0.33
		1.498M Ω	1.50000	-0.00200	0.067
		2.696M Ω	2.7000	-0.0040	0.41
	30M Ω	2.98M Ω	3.0000	-0.0200	0.43
		14.91M Ω	15.0000	-0.0900	0.65
		26.88M Ω	27.0000	-0.120	0.65



Godrej & Boyce Mfg. Co. Ltd.



CC-2495

CERTIFICATE OF CALIBRATION



CALIBRATION SERVICES

Godrej & Boyce Mfg.Co.Ltd.,
Calibration Services,MD Babar Business Center,
Godown 3-6,2 floor,Plot BGP-109
Landewadi MIDC Bhosari Pune-411026
Dir.:+91-20-46304224-4223

FR-QT-21

Certificate No.: P-240102-2-5

ULR No : CC249524000000031F

Remarks:

1. Equipment used for calibration were calibrated & traceable to National & International Standards.
2. This report refers only to particular item (s) submitted for calibration.
3. Godrej Calibration Services is not liable for any change in calibration data & performance specifications on account of malfunctioning of Standards / Instruments / Equipment covered by this report due to damage caused to it after issuance of this report.
4. The calibration results reported are valid at the time of and under the stated conditions of the measurements.
5. The Certificate should not be reproduced except in full without the prior permission from Godrej Calibration Service.
6. Satisfactory calibration report in no ways implies that the equipment calibrated is approved by NABL.
7. All precautions have been taken for any error or omission while calibrating the instruments and issuing this certificate. However Godrej Calibration Services shall not be liable for any loss or liability that may be arise to any party in this regards.
8. Readings given are average of 3 Measurements.
9. The reported Expanded Uncertainty is calculated at 95.45 % Confidence level with coverage factor k=2.

----- End Of Certificate-----



Godrej & Boyce Mfg. Co. Ltd.

Regd. Office: Pirojshanagar, L.B.S. Marg, Vikhroli, Mumbai - 400 079.T.:+91-22-67961700/1800 W:www.godrej.com

(This document is digitally signed on first page and does not require signature)