

CASE
BC-SAMPLE-001

LAB
Sample Dental Lab

DELIVERED
30 Sep 2026, 08:00 CEST

Restorations

TEETH (FDI)	TYPE	MATERIAL
36	Full-contour crown	Multilayer zirconia



PARAMETERS USED

Cement gap 0.03 mm, extra gap 0.06 mm from 1.0 mm above margin, minimum occlusal thickness per zirconia IFU, light proximal contacts, occlusal contacts light-medium.

6-point quality check

PASS	Margin line	Closed, follows prep line 360°
PASS	Proximal contacts	Mesial + distal, light
PASS	Occlusion	3 centric stops, no excursive interference
PASS	Minimum thickness & connectors	No area below material minimum
PASS	Contour / emergence profile	Anatomical, cleanable embrasures
PASS	Insertion path & files	No undercuts, watertight STL

Design views



NOTES

SAMPLE REPORT — illustration only, not a real case.

QC APPROVED BY

Dr. Alireza

Dentist · BoundlessCAD

boundlesscad.design@gmail.com
boundlesscaddesign-dev.github.io

Questions or adjustments? Reply to our email — revisions within 12 h.