# DRGS VMAT Analysis

Metadata:
File: RI.Arck.OpenBeam-0.dcm, RI.Arck.T2_DR_GS...dcm
Unit: TrueBeamSTX
Author: Karina Mitarova
Analyzed on: Friday, October 12, 2018

DRGS VMAT results:
Source-to-Image Distance (mm): 1000
Tolerance (mm): 0.013 : PASS
X-offset applied (pixels): 10
Absolute mean deviation (%): 0.43
Maximum deviation (%): 0.86