

DRGS VMAT Analysis

Metadata:

File: RI.Arck.OpenBeam-0.dcm, RI.Arck.T2_DR_GS.dcm
Unit: TrueBeamSTX
Author: Stanislav
Analyzed on: Friday, October 05, 2018

DRGS VMAT results:

Source-to-Image Distance (mm): 1000
Tolerance (mm): 0.013 : PASS
X-offset applied (pixels): 10
Absolute mean deviation (%): 0.55
Maximum deviation (%): 0.85

