

DRGS VMAT Analysis

Metadata:

File: RI.Arck.OpenBeam-0.dcm, RI.Arck.T2_DR_GS.dcm

Unit: TrueBeamSTX

Author: Karina Mitarova

Analyzed on: Saturday, October 13, 2018

DRGS VMAT results:

Source-to-Image Distance (mm): 1000

Tolerance (mm): 0.013 : PASS

X-offset applied (pixels): 10

Absolute mean deviation (%): 0.46

Maximum deviation (%): -0.76

