

DRGS VMAT Analysis

Metadata:

File: RI.Arck.OpenBeam-0.dcm, RI.Arck.T2_DR_GS.dcm
Unit: TrueBeamSTX
Author: Karina Mitarova
Analyzed on: Friday, October 12, 2018

DRGS VMAT results:

Source-to-Image Distance (mm): 1000
Tolerance (mm): 0.013 : PASS
X-offset applied (pixels): 10
Absolute mean deviation (%): 0.41
Maximum deviation (%): 0.68

