

Technician: Dana Reed

Sampler contact: 720-272-8661

Date and Time: 10/03/2016

FIELD EQUIPMENT VERIFICATION (BEFORE A WEEKLY SAMPLING):

QA/QC Check	Date	Time	Calibration standard	Calibration value	Acceptable Range (Check range)	Comments/ Notes
Temperature	10/3/16	10:31 10:33	27.2 27.2	27.2 27.3	n/a (20 – 35C)	
Salinity	9/19/16	10:50 11:00	35	35.5 1 34.9 2	± 3% of calibration solution (20 – 38ppt)	33.95 – 36.05
Dissolved Oxygen	—	—	—	—	Post-check ± 5 % of pre-check (0-100%)	
pH	10/3/16	10:08 10:18	7.00	6.99 1 6.99 2	± 3 % of calibration solution (6-8)	6.79 – 7.21

General Notes:

Hui O Ka Wai Ola, 7/10/2016