Question?

Jaemin.Shin@ge.com

Introduction

• A short SE EPI acquisition that acquires both positive (fwd; PA) and negative (rev; AP) polarity phase encoding in separate temporal phases is provided. This data can be used for offline distortion correction such as UCSD's

distortion correction (used in ABCD) and FSL TOPUP.


• Widespread adoption in GE research communities


The below issues cause confusion for GE users

Issues

- The DICOM headers report the (incorrectly) same phase encoding direction for all volumes
- The reversed images are stored to disk with the row order reversed, but no difference in spatial alignment is recorded in the DICOM header. (Flipped)

CV12: pepolar (0=fwd; 1=rev; 2=r-f; 3=f-r) (default=2) Reported in the DICOM tag (0019,10B3) UserData12






Solution

- The development branch (v1.0.202100814) of dcm2niix fixed these issues. All future stable releases will include these fixes (based on PSD name + CV12). This option provides a mechanism to trouble shoot any unintended consequences of the kludge: make CFLAGS=-DmyDisableGEPEPolarFlip
- Reported PhaseEncodingDirection/PhaseEncodingPolarityGE in the BIDS sidecar correctly
- Additionally, dcm2niix splits this series into two series expected by the BIDS and FSL's TOPUP
- en-source validation data provided by U of Michigan (UHP Rx28), https://github.com/neurolabusc/dcm_qa_polar For details: https://github.com/rordenlab/dcm2niix/issues/532

epi_pepolar (fMRI Field Map) with CV12 (pepolar) = 2 (r-f), # of phases = 4, Series# = 5

Before

rsfMRI_Field_Map_5.nii



BIDS sidecar JSON


"PhaseEncodingPolarityGE": "Flipped",

"PhaseEncodingDirection": "j",

Posterior

After (dcm2niix v1.0.202100814 or later)

rsfMRI_Field_Map_5.nii




BIDS sidecar JSON

"PhaseEncodingPolarityGE": "Flipped",

"PhaseEncodingDirection": "j",







2

BIDS sidecar JSON

"PhaseEncodingPolarityGE": "Unflipped", "PhaseEncodingDirection": "j-",

Question?
Jaemin.Shin@ge.com