Model Number:	W-12	352B01			
Serial Number:		247394			
Description:	ICP® Accelerometer				
Manufacturer:		РСВ	Method:	Back-to-Back Comparis	on AT401-3
		Calibration 1	Data		
Sensitivity @ 100 Hz		0.982 mV/g		Output Bias	10.0 VDC
Sonsiervie	, 65 100 112	_	Tuon	sverse Sensitivity	2.3 %
_		(0.1001 mV/m/s²)	11403	sverse sensitivity	2.3 70
Kesonan	t Frequency	75.9 kHz			
		Sensitivity			
3.0-	nperature: 76 °F (	24 °C)	Re	lative Humidity: 34 %	
2.0-		:			
1.0-					
В 0.0-		×	· · · · · · · · · · · · · · · · · · ·		
-1.0-		* * * * * * * * * * * * * * * * * * *			
-2.0-					
-3.0-		· · · · · · · · · · · · · · · · · · ·			
10.0 Hz		100.0		1000.0	10000.0
		Data Poir			
Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency	
10	-0.1	300	-0.5	7000	-1.6
15	0.5	500	-0.8	10000	-0.5
30	0.3	1000	-1.2		
50	0.4	3000	-1.6		
REF. FREQ.	0.0	5000	-1.7		
Mounting Surface: Tungsten Adapt Acceleration Level (pk): 10.0 g (9	o 1/-25	are Orientation: Vertical  If the listed level cannot be obtained,	the calibration system uses	the following formula to set the vibration amo	itude: Acceleration Level
(g) = $0.008 \text{ x (freq)}^2$ . The gravita	nited by snaker displacement tional constant used for calculat	ions by the calibration system is; I g = 9.80665 m/s*.			
As Found: n/a		Condition of	Unii		
	Unit, In Toler	ance			
		Notes			
1. Calibration is N	NIST Traceable	thru Project 683/287323 and	PTB Traceable	thru Project 17014.	_
0 001 1 100 1	aball wat ha name	roduced, except in full, withou	t written appro	val from PCB Piezotronic	es, Inc.

Mike Ferrio MAF Date: 3/21/2018 Technician: PCB PIEZOTRONICS

VIBRATION DIVISION

3425 Walden Avenue Depew, NY 14043

TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com



CAL4-3604497419,651+0