~ Calibration Certificate ~

352B01 Model Number: Serial Number: 256970 ICP® Accelerometer Description: **PCB** AT401-3 Manufacturer: Back-to-Back Comparison Method: Calibration Data 10.2 VDC **Output Bias** Sensitivity @ 100 Hz 0.919 mV/g 2.4 % Transverse Sensitivity $(0.0937 \text{ mV/m/s}^2)$ 117.2 kHz Resonant Frequency Sensitivity Plot Temperature: 74 °F (23 °C) Relative Humidity: 57 % 3.0 2.0 1.0 dB 0.0 -1.0 -2.0 -3.0-10000.0 1000.0 100.0 10.0 Hz Data Points Frequency (Hz) Dev. (%) Frequency (Hz) Dev. (%) Dev. (%) Frequency (Hz) 7000 -1.6 300 -0.6 10 0.6 10000 -1.4 -1.1 500 15 0.4 -1.5 30 -0.21000 0.1 3000 -1.5 50 -1.6 5000 REF. FREQ. 0.0 Mounting Surface: Tungsten Adapter Fastener: Adhesive Fixture Orientation: Vertical Acceleration Level (pk)*: 10.0 g (98.1 m/s*)

The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude; Acceleration Level (g) = 0.008 x (freq)*. The gravitational constant used for calculations by the calibration system is; 1 g = 9.80665 m/s*. Condition of Unit As Found: New Unit, In Tolerance As Left: Notes 1. Calibration is NIST Traceable thru Project 683/287323 and PTB Traceable thru Project 17014. 2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc. 3. Calibration is performed in compliance with ISO 10012-1, ANSI Z540.3 and ISO 17025. 4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications. 5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician:

Richard Gardner Rly

Date: 10/4/2018



**PCB PIEZOTRONICS"
VIBRATION DIVISION

3425 Walden Avenue Depew, NY 14043

TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com

CAL63-3621502253.514+0