

~ Calibration Certificate ~

Per ISO 16063-21

247392

Model Number: 352B01

Serial Number: 247392

Description: ICP® Accelerometer


Manufacturer: PCB

Method: Back-to-Back Comparison AT401-3

Calibration Data

Sensitivity @ 100 Hz 0.902 mV/g
(0.0919 mV/m/s²) Output Bias 10.1 VDC
Resonant Frequency 119.5 kHz Transverse Sensitivity 4.2 %

Sensitivity Plot



Data Points

Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)	Frequency (Hz)	Dev. (%)
10	0.5	300	-0.5	7000	-1.4
15	0.7	500	-0.8	10000	-0.4
30	0.4	1000	-1.1		
50	0.2	3000	-1.5		
REF. FREQ.	0.0	5000	-1.6		

Mounting Surface: Tungsten Adapter Fastener: Adhesive Fixture Orientation: Vertical
Acceleration Level (pk): 10.0 g (98.1 m/s²)

*The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtained, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level (g) = 0.008 x (freq)^{1.5}. *The gravitational constant used for calculations by the calibration system is: 1 g = 9.80665 m/s².

Condition of Unit

As Found: n/a

As Left: New Unit, In Tolerance

Notes

1. Calibration is NIST Traceable thru Project 683/287323 and PTB Traceable thru Project 17014.
2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc.
3. Calibration is performed in compliance with ISO 10012-1, ANSI/NCCL Z540-1-1994 and ISO 17025.
4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications.
5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%.

Technician: Mike Ferrio MAF Date: 3/17/2018



CALIBRATION CERT #1862.01



3425 Walden Avenue Depew, NY 14043
TEL: 888-684-0013 FAX: 716-685-3886 www.pcb.com

CAL4-3604133340.791+1