

Kalibrierschein Calibration Certificate

Kalibrierscheinnummer
Number of Calibration Certificate

6808000075

Auftraggeber <i>Customer</i>	Trescal -THALES TAS- Toulouse 26 rue Champollion F-31100 Toulouse Cedex	Die Kalibrierung erfolgt durch den Vergleich mit Normalen oder Messung auf Normalmesseinrichtungen, die auf die Nationalen Normale zurückgeführt sind, mit denen die physikalischen Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI) dargestellt werden. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.
Auftragsnummer <i>Order No.</i>	ES 90779	
Gegenstand <i>Object</i>	Accelerationsensor	
Hersteller <i>Manufacturer</i>	PCB	
Typ <i>Type</i>	352B01	Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Dieser Kalibrierschein wurde elektronisch erstellt und ist ohne Unterschrift und Stempel gültig.
Fabrikat/Seriennummer <i>Serial number</i>	107263	The calibration is performed by comparison with standards or measurement on instruments that are traceable to National Standards which realize the units of measurement according to the International System of Units (SI). The user is obliged to have the object recalibrated at appropriate intervals.
Nutzer-ID <i>User-ID</i>	107263	
Inventarnummer <i>Stock number</i>		
Schlüsselnummer <i>Key number</i>	008027147900	
Standort <i>Location</i>		This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. This calibration certificates is produced with and electronic system. This calibration certificate without signature and seal are valid.
Prüfauftragsnummer <i>Test Order No.</i>	6808000075	
Datum der Kalibrierung <i>Date of calibration</i>	09.01.2018	
Seitenanzahl des Kalibrierscheins <i>Number of pages of the certificate</i>	5	

State of reception: The measured values were within the range of the specification

Statement: Equipment may be used without exception

Ausstellungsdatum <i>Print Date</i>	Sachbearbeiter <i>Person in charge</i>	Leiter des Kalibrierlabor <i>Head of the calibration laboratory</i>
09.01.2018	Dietz	Markovic

Trescal GmbH
Tel (0711) 553651-0

Limburgstraße 6
Fax (0711) 553651-51

D-73734 Esslingen

Barcode



008027147900

1. Object

2. Measurement procedure

3. Equipment

Verwendete Normale <i>Standards used</i>	Hersteller <i>Manufacturer</i>	Typ <i>Type</i>	Serien/Inv.-Nr. <i>Serial/ Inv. No.</i>	Kalibriert am <i>Calibration at</i>	Kalibrierschein-Nr. <i>Calibration Cert. No.</i>
Shock calibrator	Endevco	2925	AB92	23.02.2017	0698 D-K-15183-01-00 2017-02
Acceleration standard	Endevco	2270	10355	21.02.2017	0697 D-K-15183-01-00 2017-02
Amplifier	Spektra GmbH	SRS 35	200427	22.02.2017	WK Spektra GmbH 17-0356
Scope	National Instruments	NI 5114	-	23.02.2017	0698 D-K-15183-01-00 2017-02

4. Conditions

Umgebungsbedingungen	Temperatur	20,5 °C	Rel. Feuchte	48 %	Luftdruck	979 hPa
<i>Environmental conditions</i>	<i>Temperature</i>		<i>Rel. Humidity</i>		<i>Air Pressure</i>	

2. Mounting of calibration device:

Screw adapter: torque **Nm**
Additive glue: glue: **Loctite**
Other:

6808000075				
Page - 3 – calibration from 09.01.2018				
Sensor:	Manufacturer	PCB	Typ	352B01
	Serial-Nr.	107263		

3. Technical details of the connecting cable:

Manufacturer:

Typ:

Length: **m**

Capacity: **pF**

Connector: **Microdot**

4. Sensitivity was calculated at following values (Gravitational acceleration $g_n = 9,80665 \text{ m/s}^2$)

Acceleration peak in g: **845,8**

Pulse duration $t_{l, 10\%}$: **0,505 ms**

5. Measured voltage: **9,973 V**

6. Amplifier

6.1. Charge amplifier of the standard

Channel of standard: **1**

Amplified factor: **16**

6.2. Amplifier calibration device

Channel of calibration device: **2**

Amplified factor: **4**

Current: **4 mA**

8. Scope

Channel from standard: **1**

Channel from calibration device: **2**

Measuring range channel 1: **10 V**

Measuring range channel 2: **10 V**

Frequency of measure: **2,9 MHz**

6808000075			
Page - 4 – calibration from 09.01.2018			
Sensor:	Manufacturer	PCB	Typ 352B01
	Serial-Nr.	107263	

5. Results of measurement

The calibrated value is sensitivity.
Following results were measured:

Sensitivity

Average value (from 5 values): **0,9459 mV/g**

Standard deviation in %: **0,353**

Calibration Nr.	Shock amplitude in g	Sensitivity S in mV/g	Pulse duration in ms
1	831,2	0,9465	0,505
2	840,7	0,9463	0,505
3	842,6	0,9458	0,506
4	843,9	0,9457	0,507
5	845,8	0,9454	0,507

6808000075			
Page - 5 – calibration from 09.01.2018			
Sensor:	Manufacturer	PCB	Typ 352B01
	Serial-Nr.	107263	

6. Uncertainty of measurement

The uncertainty of measurement is: 5,0 %.

The uncertainty of the used normals, is the standard deviation with (k=2) and P=95%.

7. Statement of conformity

The statement of conformity is in following to the DIN EN ISO 14253-1 according to Trescal-KUNO variant D.

8. Remarks