

Trescal GmbH

Kalibrierlaboratorium für elektrische, mechanische und dimensionelle Größen
Calibration laboratory for electrical, mechanical and dimensional measurand

OT 2142616

Kalibrierschein Calibration Certificate

Kalibrierscheinnummer
Number of Calibration Certificate

6808000079

Auftraggeber Customer	Trescal -THALES TAS- Toulouse 26 rue Champollion F-31100 Toulouse Cedex	Die Kalibrierung erfolgt durch den Vergleich mit Normalen oder Messung auf Normalmesseinrichtungen, die auf die Nationalen Normale zurückgeführt sind, mit denen die physikalischen Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI) dargestellt werden. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.
Auftragsnummer Order No.	ES 90779	
Gegenstand Object	Accelerationsensor	
Hersteller Manufacturer	PCB	
Typ Type	352B01	Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Dieser Kalibrierschein wurde elektronisch erstellt und ist ohne Unterschrift und Stempel gültig.
Fabrikat/Seriennummer Serial number	123382	The calibration is performed by comparison with standards or measurement on instruments that are traceable to National Standards which realize the units of measurement according to the International System of Units (SI). The user is obliged to have the object recalibrated at appropriate intervals.
Nutzer-ID User-ID	123382	
Inventarnummer Stock number		
Schlüsselnummer Key number	008033917300	
Standort Location		This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. This calibration certificates is produced with and electronic system. This calibration certifacte without signature and seal are valid.
Prüfauftragsnummer Test Order No.	6808000079	
Datum der Kalibrierung Date of calibration	09.01.2018	
Seitenanzahl des Kalibrierscheins Number of pages of the certificate	5	

State of reception: The measured values were within the range of the specification

Statement: Equipment may be used without exception

Ausstellungsdatum Print Date	Sachbearbeiter Person in charge	Leiter des Kalibrierlabor Head of the calibration laboratory
09.01.2018	Dietz	Markovic
Trescal GmbH Tel (0711) 553651-0	Limburgstraße 6 Fax (0711) 553651-51	D-73734 Esslingen

Barcode



008033917300

6808000079				
Page - 3 – calibration from 09.01.2018				
Sensor:	Manufacturer	PCB	Typ	352B01
	Serial-Nr.	123382		

3. Technical details of the connecting cable:

Manufacturer:

Typ:

Length: **m**

Capacity: **pF**

Connector: **Microdot**

4. Sensitivity was calculated at following values

(Gravitational acceleration $g_n = 9,80665 \text{ m/s}^2$)

Acceleration peak in g: **843,2**

Pulse duration $t_{1, 10\%}$: **0,523 ms**

5. Measured voltage: **9,821 V**

6. Amplifier

6.1. Charge amplifier of the standard

Channel of standard: **1**

Amplified factor: **16**

6.2. Amplifier calibration device

Channel of calibration device: **2**

Amplified factor: **4**

Current: **4 mA**

8. Scope

Channel from standard: **1**

Channel from calibration device: **2**

Measuring range channel 1: **10 V**

Measuring range channel 2: **10 V**

Frequency of measure: **2,9 MHz**

6808000079			
Page - 4 – calibration from 09.01.2018			
Sensor:	Manufacturer	PCB	Typ 352B01
	Serial-Nr.	123382	

5. Results of measurement

The calibrated value is sensitivity.
Following results were measured:

Sensitivity

Average value (from 5 values): **1,029 mV/g**

Standard deviation in %: **0,342**

Calibration Nr.	Shock amplitude in g	Sensitivity S in mV/g	Pulse duration in ms
1	826,4	1,029	0,523
2	832,6	1,029	0,520
3	832,0	1,029	0,524
4	840,9	1,028	0,516
5	843,2	1,029	0,523

6808000079			
Page - 5 – calibration from 09.01.2018			
Sensor:	Manufacturer	PCB	Typ 352B01
	Serial-Nr.	123382	

6. Uncertainty of measurement

The uncertainty of measurement is: 5,0 %.

The uncertainty of the used normals, is the standard deviation with (k=2) and P=95%.

7. Statement of conformity

The statement of conformity is in following to the DIN EN ISO 14253-1 according to Trescal-KUNO variant D.

8. Remarks